

BS ISO 9022-3:2015



BSI Standards Publication

# Optics and photonics — Environmental test methods

Part 3: Mechanical stress

**bsi.**

...making excellence a habit.™

**National foreword**

This British Standard is the UK implementation of ISO 9022-3:2015. It supersedes BS ISO 9022-3:1998 which is withdrawn.

The UK participation in its preparation was entrusted to Technical Committee CPW/172, Optics and Photonics.

A list of organizations represented on this committee can be obtained on request to its secretary.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

© The British Standards Institution 2015.  
Published by BSI Standards Limited 2015

ISBN 978 0 580 87917 3  
ICS 37.020

**Compliance with a British Standard cannot confer immunity from legal obligations.**

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 31 March 2015.

**Amendments/corrigenda issued since publication**

| Date | Text affected |
|------|---------------|
|------|---------------|

---

---

---

**Optics and photonics —  
Environmental test methods —**

**Part 3:  
Mechanical stress**

*Optique et photonique — Méthodes d'essais d'environnement —  
Partie 3: Contraintes mécaniques*





**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2015

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
Case postale 56 • CH-1211 Geneva 20  
Tel. + 41 22 749 01 11  
Fax + 41 22 749 09 47  
E-mail [copyright@iso.org](mailto:copyright@iso.org)  
Web [www.iso.org](http://www.iso.org)

Published in Switzerland

# Contents

Page

|  |           |
|--|-----------|
| <b>Foreword</b> .....  | <b>iv</b> |
| <b>Introduction</b> .....  | <b>vi</b> |
| <b>1 Scope</b> .....   | <b>1</b>  |
| <b>2 Normative references</b> .....  | <b>1</b>  |
| <b>3 General information and test conditions</b> .....                             | <b>1</b>  |
| <b>4 Conditioning</b> .....  | <b>2</b>  |
| 4.1 Conditioning method 30: Shock.....   | 2         |
| 4.2 Conditioning method 31: Bump.....  | 2         |
| 4.3 Conditioning method 32: Drop and topple.....                                   | 3         |
| 4.4 Conditioning method 33: Free fall.....   | 3         |
| 4.5 Conditioning method 34: Bounce.....  | 3         |
| 4.6 Conditioning method 35: Steady-state acceleration, centrifugal.....            | 4         |
| 4.7 Conditioning method 36: Sinusoidal vibration.....                              | 4         |
| 4.7.1 General.....   | 4         |
| 4.7.2 Vibration testing using sweep frequencies.....                               | 4         |
| 4.7.3 Vibration fatigue test using characteristic frequencies.....                 | 5         |
| 4.8 Conditioning method 37: Random vibration (wide-band) digitally controlled..... | 5         |
| <b>5 Procedure</b> .....   | <b>6</b>  |
| <b>6 Environmental test code</b> .....   | <b>6</b>  |
| <b>7 Specification</b> .....   | <b>6</b>  |

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see [www.iso.org/directives](http://www.iso.org/directives)).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see [www.iso.org/patents](http://www.iso.org/patents)).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the WTO principles in the Technical Barriers to Trade (TBT) see the following URL: [Foreword - Supplementary information](#)

The committee responsible for this document is ISO/TC 172, *Optics and photonics*, Subcommittee SC 1, *Fundamental standards*.

This third edition cancels and replaces the second edition (ISO 9022-3:1998), of which it constitutes a minor revision.

ISO 9022 consists of the following parts, under the general title *Optics and photonics — Environmental test methods*:

- *Part 1: Definitions, extent of testing*
- *Part 2: Cold, heat and humidity*
- *Part 3: Mechanical stress*
- *Part 4: Salt mist*
- *Part 6: Dust*
- *Part 7: Resistance to drip or rain*
- *Part 8: High internal pressure, low internal pressure, immersion*
- *Part 9: Solar radiation and weathering*
- *Part 11: Mould growth*
- *Part 12: Contamination*
- *Part 14: Dew, hoarfrost, ice*
- *Part 17: Combined contamination, solar radiation*
- *Part 20: Humid atmosphere containing sulfur dioxide or hydrogen sulfide*

- *Part 22: Combined cold, dry heat or temperature change with bump or random vibration*
- *Part 23: Low pressure combined with cold, ambient temperature and dry and damp heat*

## Introduction

Optical instruments are affected during their use by a number of different environmental parameters which they are required to resist without significant reduction in performance and to remain within defined specifications.

The type and severity of these parameters depend on the conditions of use of the instrument (for example, in the laboratory or workshop) and on its geographical location. The environmental effects on optical instrument performance in the tropics and subtropics are totally different from those found when they are used in arctic regions. Individual parameters cause a variety of different and overlapping effects on instrument performance.

The manufacturer attempts to ensure, and the user naturally expects, that instruments will resist the likely rigours of their environment throughout their life. This expectation can be assessed by exposure of the instrument to a range of simulated environmental parameters under controlled laboratory conditions. The severity of these conditions is often increased to obtain meaningful results in a relatively short period of time.

In order to allow assessment and comparison of the response of optical instruments to appropriate environmental conditions, ISO 9022 contains details of a number of laboratory tests which reliably simulate a variety of different environments. The tests are based largely on IEC standards, modified where necessary to take into account features special to optical instruments.

As a result of continuous progress in all fields, optical instruments are no longer only precision-engineered optical products, but, depending on their range of application, also contain additional assemblies from other fields. For this reason, the principal function of the instrument is to be assessed to determine which International Standard should be used for testing. If the optical function is of primary importance, then ISO 9022 is applicable, but if other functions take precedence, then the appropriate International Standard in the field concerned should be applied. Cases can arise where application of both ISO 9022 and other appropriate International Standards will be necessary.



# Optics and photonics — Environmental test methods —

## Part 3: Mechanical stress

### 1 Scope

This part of ISO 9022 specifies the methods relating to the environmental tests of optical instruments including additional assemblies from other fields (e.g. mechanical, chemical, and electronic devices), under equivalent conditions, for their ability to resist the influence of mechanical stress.

The purpose of the testing is to investigate to what extent the optical, climatic, mechanical, chemical, and electrical (including electrostatic) performance characteristics of the specimen are affected by mechanical stress.

### 2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 9022-1, *Optics and photonics — Environmental test methods — Part 1: Definitions, extent of testing*

IEC 60068-2-6:2007, *Environmental testing — Part 2-6: Tests — Test Fc: Vibration (sinusoidal)*

IEC 60068-2-7, *Environmental testing — Part 2-7: Tests — Test Ga and guidance: Acceleration, steady state*

IEC 60068-2-27, *Environmental testing — Part 2-27: Tests — Test Ea and guidance: Shock*

IEC 60068-2-31, *Environmental testing — Part 2-31: Tests — Test Ec: Rough handling shocks, primarily for equipment type specimens*

IEC 60068-2-47, *Environmental testing — Part 2-47: Tests — Mounting of specimens for vibration, impact and similar dynamic tests*

IEC 60068-2-55, *Environmental testing — Part 2-55: Tests — Test Ee and guidance: Loose cargo testing including bounce*

IEC 60068-2-64, *Environmental testing — Part 2-64: Test methods — Test Fh: Vibration, broadband random and guidance*

### 3 General information and test conditions

The test shall be carried out at ambient atmospheric conditions and in accordance with ISO 9022-1 and with the International Standards listed in [Table 1](#). The specimens shall be mounted on the test apparatus (shock machine, acceleration facility, or electrodynamic shaker) in accordance with IEC 60068-2-47.

“ $g_n$ ” is the standard acceleration due to the earth’s gravity, which itself varies with altitude and geographical latitude.

NOTE For the purposes of this part of ISO 9022, the value of  $g_n$  is rounded up to the nearest whole number which is 10 m/s<sup>2</sup>.

**Table 1 — Conditioning methods and applicable International Standards for testing**

| Subclause           | Conditioning methods                              | International Standard |
|---------------------|---|------------------------|
| <a href="#">4.1</a> | 30: Shock   | IEC 60068-2-27         |
| <a href="#">4.2</a> | 31: Bump  | IEC 60068-2-27         |
| <a href="#">4.3</a> | 32: Drop and topple                               | IEC 60068-2-31         |
| <a href="#">4.4</a> | 33: Free fall                                     | IEC 60068-2-31         |
| <a href="#">4.5</a> | 34: Bounce  | IEC 60068-2-55         |
| <a href="#">4.6</a> | 35: Steady-state acceleration                     | IEC 60068-2-7          |
| <a href="#">4.7</a> | 36: Sinusoidal vibration                          | IEC 60068-2-6          |
| <a href="#">4.8</a> | 37: Random vibration (wide-band), digital control | IEC 60068-2-64         |

## 4 Conditioning

### 4.1 Conditioning method 30: Shock

See [Table 2](#).

When testing optical instruments, a half-sine shock pulse shall be applied. The specimen shall be subjected to three shocks in each direction along each axis.

**Table 2 — Degrees of severity for conditioning method 30: Shock**

| Degree of severity   |                   | 01          | 02  | 03  | 04  | 05  | 06  | 07    | 08 <sup>a</sup> |
|--|-------------------|-------------|-----|-----|-----|-----|-----|-------|-----------------|
| Acceleration amplitude   | m s <sup>-2</sup> | 100         | 150 | 300 | 300 | 500 | 500 | 1 000 | 5 000           |
|  | $g_n$ multiples   | 10          | 15  | 30  | 30  | 50  | 50  | 100   | 500             |
| Duration of nominal shock  | ms                | 6           | 11  | 6   | 18  | 3   | 11  | 6     | 1               |
| State of operation   |                   | 0 or 1 or 2 |     |     |     |     |     |       |                 |
| NOTE Degrees of severity 02, 03 and 05 are to be given preference.   |                   |             |     |     |     |     |     |       |                 |
| <sup>a</sup> Applicable to testing of components and assemblies. Complete optical instruments should be subjected to 500 $g_n$ acceleration and shocks of 0,5 ms duration. |                   |             |     |     |     |     |     |       |                 |

### 4.2 Conditioning method 31: Bump

See [Table 3](#).

**Table 3 — Degrees of severity for conditioning method 31: Bump**

| Degree of severity  |                   | 01          | 02    | 03    | 04    | 05    | 06    | 07    | 08    |
|---|-------------------|-------------|-------|-------|-------|-------|-------|-------|-------|
| Acceleration amplitude                                      | m s <sup>-2</sup> | 100         | 100   | 100   | 100   | 250   | 250   | 400   | 400   |
|   | $g_n$ multiples   | 10          | 10    | 10    | 10    | 25    | 25    | 40    | 40    |
| Duration of nominal shock                                   | ms                | 6           | 6     | 16    | 16    | 6     | 6     | 6     | 6     |
| Number of shocks in each direction along each axis $\pm 10$ |                   | 1 000       | 4 000 | 1 000 | 4 000 | 1 000 | 4 000 | 1 000 | 4 000 |
| State of operation  |                   | 0 or 1 or 2 |       |       |       |       |       |       |       |

### 4.3 Conditioning method 32: Drop and topple

See [Table 4](#).

**Table 4 — Degrees of severity for conditioning method 32: Drop and topple**

| Degree of severity   |                         | 01 <sup>a</sup> | 02 <sup>a</sup> | 03 <sup>a</sup> | 04 <sup>b</sup> |
|--|-------------------------|-----------------|-----------------|-----------------|-----------------|
| Height of overturn   | mm                      | 25              | 50              | 100             | Toppling over   |
|  | Acceptable deviation mm | ±5              |                 |                 | —               |
| State of operation   |                         | 0 or 1          |                 |                 |                 |
| <sup>a</sup> The specimen shall be subjected to one drop on each of four bottom corners and along each of four bottom edges. |                         |                 |                 |                 |                 |
| <sup>b</sup> The specimen shall be subjected to one topple about each of four bottom edges.                                  |                         |                 |                 |                 |                 |

### 4.4 Conditioning method 33: Free fall

See [Table 5](#).

**Table 5 — Degrees of severity for conditioning method 33: Free fall**

| Degree of severity  |                         | 01     | 02   | 03   | 04   | 05  | 06    |
|---|-------------------------|--------|------|------|------|-----|-------|
| Height of fall  | mm                      | 25     | 50   | 100  | 250  | 500 | 1 000 |
|   | Acceptable deviation mm | ±5     |      |      | ±10  |     |       |
| State of operation  |                         | 0 or 1 |      |      |      |     |       |
| Mass of specimen including packing <sup>a</sup> kg                |                         | >500   | ≤500 | ≤200 | ≤100 | ≤50 | ≤20   |
| NOTE Storage containers are not to be considered as packing.      |                         |        |      |      |      |     |       |
| <sup>a</sup> Recommendation for selection of degrees of severity. |                         |        |      |      |      |     |       |

Unpackaged optical instruments shall not be tested unless they are especially designed, constructed, and armoured (e.g. rubber armouring) for free fall. The degrees of severity are applicable to normal transport handling. Unless otherwise prescribed in the relevant specification, the specimen shall be subjected to two falls. If another number of falls is taken, the total number of falls shall be preferably taken from the following series: 10, 20, 50.

### 4.5 Conditioning method 34: Bounce

See [Table 6](#).

The test shall be carried out according to IEC 60068-2-55 on a bounce table with a double amplitude of 25,5 mm ± 0,5 mm and a frequency of 4,75 Hz ± 0,05 Hz.

**Table 6 — Degrees of severity for conditioning method 34: Bounce**

| Degree of severity   |                      | 01     | 02 | 03  |
|--|----------------------|--------|----|-----|
| Exposure time  | min                  | 15     | 60 | 180 |
|  | Acceptable deviation | ±10 %  |    |     |
| State of operation   |                      | 0 or 1 |    |     |
| NOTE Degree of severity 02 is to be given preference. The period of exposure is to be allocated in equal portions to each of the surfaces to be exposed. |                      |        |    |     |

#### 4.6 Conditioning method 35: Steady-state acceleration, centrifugal

See [Table 7](#).

**Table 7 — Degrees of severity for conditioning method 35: Steady-state acceleration, centrifugal**

| Degree of severity                                  |                 | 01               | 02  | 03  | 04  | 05    | 06    |
|---|-----------------|------------------|-----|-----|-----|-------|-------|
| Acceleration  | $m\ s^{-2}$     | 50               | 100 | 200 | 500 | 1 000 | 2 000 |
|   | $g_n$ multiples | 5                | 10  | 20  | 50  | 100   | 200   |
| Exposure time along each axis and in each direction | s               | >10 <sup>a</sup> |     |     |     |       |       |
| State of operation                                  |                 | 0 or 1 or 2      |     |     |     |       |       |

<sup>a</sup> The exposure time begins after reaching the rated number of revolutions.

#### 4.7 Conditioning method 36: Sinusoidal vibration

##### 4.7.1 General

The degrees of severity specified in [Table 8](#) are relevant to optical instruments because the low frequencies combined with large displacement amplitudes do not stress optical instruments.

In special cases, refer to IEC 60068-2-6:2007, Figure 1.

##### 4.7.2 Vibration testing using sweep frequencies

See [Table 8](#) and [Table 9](#).

**Table 8 — Degrees of severity for conditioning method 36: Sinusoidal vibration using sweep frequencies**

| Degree of severity   |                  | 01          | 02    | 03   | 04   | 05   | 06   | 07   | 08   | 09   | 10  |
|--|------------------|-------------|-------|------|------|------|------|------|------|------|-----|
| Displacement   | mm               | 0,035       | 0,075 | 0,15 | 0,15 | 0,15 | 0,15 | 0,35 | 0,35 | 0,35 | 1,0 |
| Acceleration   | $m\ s^{-2}$      | 5           | 10    | 20   | 20   | —    | 20   | 50   | 50   | 50   | —   |
|  | $g_n$ multiples  | 0,5         | 1     | 2    | 2    | —    | 2    | 5    | 5    | 5    | —   |
| Number of frequency cycles <sup>a</sup> to be used on each axis per frequency band | 10 Hz to 55 Hz   | —           | —     | —    | —    | 5    | —    | —    | —    | —    | 20  |
|  | 10 Hz to 150 Hz  | —           | —     | 20   | —    | —    | —    | 5    | —    | —    | —   |
|  | 10 Hz to 500 Hz  | 2           | —     | —    | 10   | —    | —    | —    | 10   | —    | —   |
|  | 0 Hz to 2 000 Hz | —           | 2     | —    | —    | —    | 10   | —    | —    | 10   | —   |
| State of operation   |                  | 0 or 1 or 2 |       |      |      |      |      |      |      |      |     |

<sup>a</sup> The sweep rate for the specified number of frequency cycles shall be 1 octave per minute.

**Table 9 — Typical applications**

| Frequency band<br>Hz | Examples of application  |
|----------------------|--|
| 10 to 55             | Instruments installed in ships and other naval craft or in the neighbourhood of heavy rotating machines and for general industrial requirements. |
| 10 to 150            | Instruments for general industrial requirements and for use in and transport on ground vehicles.   |
| 10 to 500            | Equipment for general airborne use and for use in ground vehicles (e.g. tracked vehicles) under special conditions.                              |

Table 9 (continued)

| Frequency band<br>Hz | Examples of application   |
|----------------------|---|
| 10 to 2 000          | Equipment for use in high-speed aircraft and missiles and in special vehicles such as hovercraft. |

#### 4.7.3 Vibration fatigue test using characteristic frequencies

See [Table 10](#).

The vibration fatigue test, using characteristic frequencies, shall not be performed unless in combination with the condition specified in [4.7.2](#).

The specimen shall be vibrated along each axis for the time specified in [Table 10](#). If the characteristic frequencies depend on the location of the specimen, they shall be specified in the relevant specification. In the event that more than one characteristic frequency is used, portions of the exposure time shall be allocated to each frequency. The portion of exposure time to be allocated to each characteristic frequency shall be specified in the relevant specification.

Table 10 — Duration of the vibration fatigue test using characteristic frequencies

| Parameter                                      |                      | Requirement                                 |    |    |
|--|----------------------|---|----|----|
| Acceleration or displacement                   |                      | To be selected from <a href="#">Table 8</a> |    |    |
| Exposure time using characteristic frequencies | min                  | 10  | 30 | 90 |
|  | Acceptable deviation | ±10 %                                       |    |    |

#### 4.8 Conditioning method 37: Random vibration (wide-band) digitally controlled

The total conditioning time which is specified in [Table 11](#), [Table 12](#), and [Table 13](#) shall be divided equally between the conditioning axes defined in the relevant specification.

Table 11 — Degrees of severity for conditioning method 37: Random vibration; frequency range from 20 Hz to 150 Hz

| Degree of severity                                       |                      | 01          | 02   | 03  | 04  |
|--|----------------------|-------------|------|-----|-----|
| Acceleration power spectral density                      | $g_n^2/\text{Hz}$    | 0,02        | 0,05 | 0,2 | 0,2 |
| Rms acceleration <sup>a</sup>                            | $g_n$ multiples      | 1,6         | 2,6  | 5,1 | 5,1 |
| Frequency range ( $f_1$ to $f_2$ )                       |                      | 20 to 150   |      |     |     |
| Conditioning time along each axis                        | min                  | 10          | 10   | 10  | 30  |
|  | Acceptable deviation | ±10 %       |      |     |     |
| State of operation                                       |                      | 0 or 1 or 2 |      |     |     |
| <sup>a</sup> The values refer to a rectangular spectrum. |                      |             |      |     |     |

Table 12 — Degrees of severity for conditioning method 37: Random vibration; frequency range from 20 Hz to 500 Hz

| Degree of severity                                       |                   | 11        | 12   | 13   | 14   | 15   |
|--|-------------------|-----------|------|------|------|------|
| Acceleration power spectral density                      | $g_n^2/\text{Hz}$ | 0,005     | 0,01 | 0,05 | 0,05 | 0,05 |
| Rms acceleration <sup>a</sup>                            | $g_n$ multiples   | 1,6       | 2,2  | 4,9  | 4,9  | 4,9  |
| Frequency range ( $f_1$ to $f_2$ )                       |                   | 20 to 500 |      |      |      |      |
| <sup>a</sup> The values refer to a rectangular spectrum. |                   |           |      |      |      |      |

Table 12 (continued)

| Degree of severity                            |                      | 11          | 12 | 13 | 14 | 15 |
|---|----------------------|-------------|----|----|----|----|
| Conditioning time along each axis             | min                  | 10          | 10 | 10 | 30 | 90 |
|   | Acceptable deviation | ±10 %       |    |    |    |    |
| State of operation                            |                      | 0 or 1 or 2 |    |    |    |    |
| a The values refer to a rectangular spectrum. |                      |             |    |    |    |    |

Table 13 — Degrees of severity for conditioning method 37: Random vibration; frequency range from 20 Hz to 2 000 Hz

| Degree of severity                            |                      | 21          | 22   | 23   | 24   | 25 <sup>a</sup> | 26 <sup>b</sup> |
|---|----------------------|-------------|------|------|------|-----------------|-----------------|
| Acceleration power spectral density           | $g_n^2/\text{Hz}$    | 0,001       | 0,01 | 0,01 | 0,05 | 0,02            | 0,05            |
| Rms acceleration <sup>b</sup>                 | $g_n$ multiples      | 1,4         | 4,5  | 4,5  | 10   | 6,3             | 10              |
| Frequency range ( $f_1$ to $f_2$ )            |                      | 20 to 2 000 |      |      |      |                 |                 |
| Conditioning time along each axis             | min                  | 10          | 10   | 30   | 30   | 90              | 90              |
|   | Acceptable deviation | ±10 %       |      |      |      |                 |                 |
| State of operation                            |                      | 0 or 1 or 2 |      |      |      |                 |                 |
| a For missiles and jet aircraft.              |                      |             |      |      |      |                 |                 |
| b The values refer to a rectangular spectrum. |                      |             |      |      |      |                 |                 |

## 5 Procedure

The tests shall be conducted in accordance with the requirements of the relevant specification and with ISO 9022-1 and the relevant parts of IEC 60068 listed in [Clause 2](#).

## 6 Environmental test code

The environmental test code shall be as defined in ISO 9022-1, giving a reference to ISO 9022 and the codes for the conditioning method chosen, the degree of severity, and the state of operation.

EXAMPLE The environmental test of optical instruments for resistance to shock, conditioning method 30, degree of severity 01, and state of operation 1 is identified as:

**Environmental test ISO 9022-30-01-1**

## 7 Specification

The relevant specification shall contain the following details:

- environmental test code;
- number of specimens;
- data as required by the International Standards listed in [Table 1](#) (the requirements of ISO 9022-3 shall prevail);
- conditioning methods 30 and 31: axes and directions of exposure;
- conditioning method 32: edge over which to tilt and number of conditionings;
- conditioning method 33: conditioning of package prior to and after conditioning, number of conditionings, and number of the edges, corners, and surfaces to be exposed;

- g) conditioning method 34: surface to be exposed;
- h) conditioning method 35: axes along which, and directions in which, specimens are to be exposed;
- i) conditioning methods 36 and 37: axes along which to expose specimens;
- j) conditioning methods 36 (4.7.3): portion of exposure time to be allocated to each characteristic frequency; characteristic frequency at the installed position of the specimen, where appropriate;
- k) preconditioning;
- l) type and scope of initial test;
- m) period of operation for state of operation 2;
- n) type and scope of intermediate test for state operation 2;
- o) recovery;
- p) type and scope of final test;
- q) criteria for evaluation;
- r) type and scope of test report.







# British Standards Institution (BSI)

BSI is the national body responsible for preparing British Standards and other standards-related publications, information and services.

BSI is incorporated by Royal Charter. British Standards and other standardization products are published by BSI Standards Limited.

## About us

We bring together business, industry, government, consumers, innovators and others to shape their combined experience and expertise into standards-based solutions.

The knowledge embodied in our standards has been carefully assembled in a dependable format and refined through our open consultation process. Organizations of all sizes and across all sectors choose standards to help them achieve their goals.

## Information on standards

We can provide you with the knowledge that your organization needs to succeed. Find out more about British Standards by visiting our website at [bsigroup.com/standards](http://bsigroup.com/standards) or contacting our Customer Services team or Knowledge Centre.

## Buying standards

You can buy and download PDF versions of BSI publications, including British and adopted European and international standards, through our website at [bsigroup.com/shop](http://bsigroup.com/shop), where hard copies can also be purchased.

If you need international and foreign standards from other Standards Development Organizations, hard copies can be ordered from our Customer Services team.

## Subscriptions

Our range of subscription services are designed to make using standards easier for you. For further information on our subscription products go to [bsigroup.com/subscriptions](http://bsigroup.com/subscriptions).

With **British Standards Online (BSOL)** you'll have instant access to over 55,000 British and adopted European and international standards from your desktop. It's available 24/7 and is refreshed daily so you'll always be up to date.

You can keep in touch with standards developments and receive substantial discounts on the purchase price of standards, both in single copy and subscription format, by becoming a **BSI Subscribing Member**.

**PLUS** is an updating service exclusive to BSI Subscribing Members. You will automatically receive the latest hard copy of your standards when they're revised or replaced.

To find out more about becoming a BSI Subscribing Member and the benefits of membership, please visit [bsigroup.com/shop](http://bsigroup.com/shop).

With a **Multi-User Network Licence (MUNL)** you are able to host standards publications on your intranet. Licences can cover as few or as many users as you wish. With updates supplied as soon as they're available, you can be sure your documentation is current. For further information, email [bsmusales@bsigroup.com](mailto:bsmusales@bsigroup.com).

## BSI Group Headquarters

389 Chiswick High Road London W4 4AL UK

## Revisions

Our British Standards and other publications are updated by amendment or revision.

We continually improve the quality of our products and services to benefit your business. If you find an inaccuracy or ambiguity within a British Standard or other BSI publication please inform the Knowledge Centre.

## Copyright

All the data, software and documentation set out in all British Standards and other BSI publications are the property of and copyrighted by BSI, or some person or entity that owns copyright in the information used (such as the international standardization bodies) and has formally licensed such information to BSI for commercial publication and use. Except as permitted under the Copyright, Designs and Patents Act 1988 no extract may be reproduced, stored in a retrieval system or transmitted in any form or by any means – electronic, photocopying, recording or otherwise – without prior written permission from BSI. Details and advice can be obtained from the Copyright & Licensing Department.

## Useful Contacts:

### Customer Services

**Tel:** +44 845 086 9001

**Email (orders):** [orders@bsigroup.com](mailto:orders@bsigroup.com)

**Email (enquiries):** [cservices@bsigroup.com](mailto:cservices@bsigroup.com)

### Subscriptions

**Tel:** +44 845 086 9001

**Email:** [subscriptions@bsigroup.com](mailto:subscriptions@bsigroup.com)

### Knowledge Centre

**Tel:** +44 20 8996 7004

**Email:** [knowledgecentre@bsigroup.com](mailto:knowledgecentre@bsigroup.com)

### Copyright & Licensing

**Tel:** +44 20 8996 7070

**Email:** [copyright@bsigroup.com](mailto:copyright@bsigroup.com)



...making excellence a habit.™