

BS ISO 10260:1992



BSI Standards Publication

Water quality — Measurement of biochemical parameters — Spectrometric determination of the chlorophyll-a concentration

bsi.

...making excellence a habit.™

National foreword

This British Standard is the UK implementation of ISO 10260:1992.

The UK participation in its preparation was entrusted to Technical Committee EH/3/2, Physical chemical and biochemical methods.

A list of organizations represented on this committee can be obtained on request to its secretary.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

© The British Standards Institution 2014.
Published by BSI Standards Limited 2014

ISBN 978 0 580 87041 5
ICS 13.060.50

Compliance with a British Standard cannot confer immunity from legal obligations.

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 30 June 2014.

Amendments/corrigenda issued since publication

Date	Text affected
------	---------------

INTERNATIONAL STANDARD

ISO
10260

First edition
1992-07-15

Water quality — Measurement of biochemical parameters — Spectrometric determination of the chlorophyll-a concentration

*Qualité de l'eau — Mesurage des paramètres biochimiques — Dosage
spectrométrique de la chlorophylle a*



Reference number
ISO 10260:1992(E)

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

International Standard ISO 10260 was prepared by Technical Committee ISO/TC 147, *Water quality*, Sub-Committee SC 2, *Physical, chemical, biochemical methods*.

Annexes A and B of this International Standard are for information only.

© ISO 1992

All rights reserved. No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

International Organization for Standardization
Case Postale 56 • CH-1211 Genève 20 • Switzerland

Introduction

Chlorophyll-a is the essential photosynthetic pigment present in all green plants. The chlorophyll content of a surface water is an indicator of its trophic state. The determination of the chlorophyll-a concentration provides information concerning the quantity and potential photosynthetic activity of algae. The most important metabolites of chlorophylls are phaeophytines and phaeophorbide. The ratio of chlorophyll to phaeopigments is indicative of the physiological state of the algae.

Water quality — Measurement of biochemical parameters — Spectrometric determination of the chlorophyll-a concentration

1 Scope

1.1 This International Standard specifies a method for the determination of the chlorophyll-a concentration. The procedure can be applied for phytoplankton in natural surface waters and for testing algal growth in bio-assays. Using appropriate sampling it can also be applied to phytobenthic communities (see annex A).

1.2 Other algal pigments such as chlorophyll-b and chlorophyll-c and some chlorophyll metabolites do not contribute to the determination. Phaeopigments may be determined semiquantitatively, to correct for interference with chlorophyll-a determination and to indicate the portion of inactive algal biomass.

1.3 Chlorophyll is sensitive to light and oxygen, especially when it is extracted. To avoid oxidative and photochemical destruction, the samples shall not be exposed to bright light or air. Homogenization of the sample may in some cases increase the extraction efficiency.

1.4 The extraction procedure with ethanol involves heating to 75 °C for 5 min to inactivate chlorophyllase and accelerate the lysis of pigments. Storage of extracts (except filters containing suspended matter) prior to photometric measurement should be kept to a minimum, but is possible up to 3 d under refrigeration at 4 °C. Storage of extracts at less than – 25 °C is possible for at least 30 d.

1.5 Even though the procedure involves filtration or centrifugation to clarify the final extract, a slight turbidity may remain. The acidification step may also cause turbidity. Therefore, the absorbance measured at 665 nm has to be corrected for turbidity by subtracting the absorbance measured at 750 nm.

1.6 The pigment of certain rarely occurring phototrophic bacteria (e.g. *Chlorobium*) interferes with the determination of chlorophyll-a concentration [1]. The contribution of chlorophyll-b and chlorophyll-c to the absorbance at 665 nm is negligible [2].

2 Normative references

The following standards contain provisions which, through reference in this text, constitute provisions of this International Standard. At the time of publication, the editions indicated were valid. All standards are subject to revision, and parties to agreements based on this International Standard are encouraged to investigate the possibility of applying the most recent editions of the standards indicated below. Members of IEC and ISO maintain registers of currently valid International Standards.

ISO 5667-1:1980, *Water quality — Sampling — Part 1: Guidance on the design of sampling programmes*.

ISO 5667-2:1991, *Water quality — Sampling — Part 2: Guidance on sampling techniques*.

3 Principle

Collection of algae and other suspended matter from a water sample by filtration. Extraction of algal pigments from the filter residue into hot ethanol. Spectrometric determination of the chlorophyll-a concentration in the extract. Evaluation of the chlorophyll-a and phaeopigment concentration from the difference in absorbance at 665 nm prior to and after acidification of the extract [3] [4].

4 Reagents

Use only reagents of recognized analytical grade and only deionized water of equivalent purity.

4.1 Hydrochloric acid, $c(\text{HCl}) = 3 \text{ mol/l}$.

4.2 Ethanol,
($\text{C}_2\text{H}_5\text{OH}$), aqueous solution 90 % (V/V).

NOTE 1 Generally, a denaturant in ethanol does not interfere. Nevertheless, a comparative determination with pure ethanol (90 %) is recommended with each unknown batch [4].

5 Apparatus

Ordinary laboratory apparatus and the following:

5.1 Spectrometer, for use in the visible range up to 750 nm, with a resolution of 1 nm, a bandwidth of 2 nm or less, sensitivity less than or equal to 0,001 absorbance units and with optical cells of path length between 1 cm and 5 cm.

5.2 Vacuum filtration device, filter holder with clamp.

5.3 Glass-fibre filters free of organic binder, for filtration of water samples, retaining more than 99 % of particles greater than 1 μm . Suitable diameters range from 25 mm to 50 mm.

5.4 Filters for filtration of extracts, as described in 5.3 but of small diameter e.g. 25 mm.

Alternative: Centrifuge, with an acceleration of 6 000 g and a rotor suitable for appropriate extraction tubes.

5.5 Water bath, adjustable to $75 \text{ }^\circ\text{C} \pm 1 \text{ }^\circ\text{C}$ with a rack for extraction vessels.

5.6 Extraction vessels, e.g. wide-necked amber glass vials with polytetrafluoroethylene (PTFE)-lined screw caps, of typical capacity 30 ml to 50 ml, suitable for centrifugation at 6 000 g .

6 Sampling and storage

Sample according to ISO 5667-1 and ISO 5667-2. Refrigerated storage of water samples in the dark for less than 8 h is acceptable but should be avoided. If possible, perform steps 7.1 to 7.3 immediately after sampling. If necessary, store the raw extracts in air-tight brown glass extraction vessels (5.6) below $-25 \text{ }^\circ\text{C}$ for up to 30 d. Do not store frozen water samples or filters plus solids.

7 Procedure

7.1 Filtration

Shake the samples in order to mix thoroughly. Filter a measured volume of sample V_s (normally in the

range 0,1 litre to 2 litres, depending on the concentration of algae) through a glass-fibre filter (5.3) clamped in a suitable holder. Dry the filter in a vacuum, as soon as it is dry remove it from the holder and place it in the extraction vessel. If it does not fit into the extraction vessel, tear it into pieces.

Avoid contact with fingers.

7.2 Extraction variant A

Heat the required volume of ethanol (4.2) to $75 \text{ }^\circ\text{C}$.

Pour a small volume (usually 30 ml to 40 ml) of the hot ethanol into the vessel containing the filter or filter pieces. After cooling for a few minutes, grind the filter to facilitate extraction, preferably with a rod-shaped homogenizer. Wash the homogenizer rod with a small volume of ethanol (4.2) to remove particles of sample adhering to it. Extract the suspension for at least 3 min.

NOTE 2 Usually the extraction is carried out at room temperature for several hours or overnight. If the extraction is prolonged, or the extract is stored for several days (e.g. over the weekend), the extraction vessels should be stored in a refrigerator.

Filter the slurry through a dense filter (5.4) into a calibrated flask (of capacity 50 ml or 100 ml) with a stopper. Wash the extraction vessel with ethanol (4.2) to remove residual extract and transfer quantitatively while rinsing the filter into the calibrated flask. Fill to the mark, stopper and mix thoroughly. This is the extract volume V_e .

Proceed with step 7.4.

7.3 Extraction variant B

Dispense an exact volume V_E (usually 20 ml or 25 ml) of the ethanol (4.2) into the extraction vessel (5.6) and allow the filter pieces to submerge. Close the screw cap tightly to avoid losses from evaporation of extractant. Shake slightly to resuspend the filter residue. Place the tube in the water bath (5.5) so that the extractant level aligns with the level of the bath. Heat for 5 min, shaking slightly if necessary. Take the extraction vessels from the water bath and allow to cool to room temperature for 15 min.

The time between extraction and measurement should be kept to a minimum.

NOTE 3 Extracts at this stage may be stored in a refrigerator overnight prior to measurement (see 7.4). Prolonged storage shall not exceed 3 d.

Filter the supernatant extract through a filter (5.4) into a clean extraction vessel (5.6) but do not rinse with fresh solvent (see note 4).

Alternatively, centrifuge the extraction vessels for a period which is sufficient to obtain a clear supernatant.

Use a clear extract or supernatant for photometry.

NOTE 4 Using this procedure, it is sufficient to recover merely a part of the extract volume, because the initial volume V_e of extractant is known exactly and is not reduced due to evaporation during extraction because the caps are tightly closed. Furthermore, the residual water content of the filters (see 7.1) is negligible, contributing to far less than 5 % of the extract volume.

7.4 Photometry

7.4.1 Transfer part of the clear extract into the spectrometer cuvette using a pipette, leaving a sufficient volume for the acidification step (see 7.4.2).

Measure the absorbance at 665 nm (A_{665}) and 750 nm (A_{750}) against a reference cell filled with ethanol (4.2).

NOTE 5 The absorbance at 665 nm should fall between 0,01 and 0,8 units. This may be achieved by suitably choosing the water volume filtered, extractant volume, dilution, optical pathlength, etc. To start with, take 0,5 litre of sample, a 50 mm diameter filter, 20 ml of ethanol and a 5 cm cuvette.

7.4.2 Acidify part of the extract (usually 5 ml to 10 ml) with 0,01 ml of hydrochloric acid (4.1) per 10 ml of extract volume, shake and measure the absorbance again at 665 nm and 750 nm, after 5 min to 30 min.

8 Calculation and expression of results

8.1 The chlorophyll-a concentration ρ_c , in micrograms per litre, is calculated according to the equation

$$\rho_c = \frac{(A - A_a)}{K_c} \times \frac{R}{R - 1} \times \frac{10^3 V_e}{V_s \cdot d} \quad \dots (1)$$

where

$A = A_{665} - A_{750}$ is the absorbance of the extract before acidification (see 7.4.1);

$A_a = A_{665} - A_{750}$ is the absorbance of the extract after acidification;

V_e is the volume, in millilitres, of extract;

V_s is the volume, in litres, of filtered sample;

$K_c = 82 \text{ l}/\mu\text{g}\cdot\text{cm}$ is the specific operational spectral absorption coefficient for chlorophyll-a (the value is taken from [2]);

$R = 1,7$ is the ratio A/A_a for a solution of pure chlorophyll-a which is transferred to phaeophytin by acidification (see 7.4.2) (the value is taken from [2]);

d is the path length, in centimetres, of the optical cell;

10^3 is the dimension factor to fit V_e .

8.2 The phaeopigment concentration ρ_p , in micrograms per litre, is calculated according to the equation

$$\rho_p = A_a \times \frac{R}{K_c} \times \frac{10^3 V_e}{V_s \cdot d} - \rho_c \quad \dots (2)$$

8.3 When the value 82 is taken for the specific spectral absorption coefficient for chlorophyll-a in 90 % ethanol, and 1,7 is taken for the maximum acid ratio (R) for pure chlorophyll-a, the chlorophyll-a concentration ρ_c in the water sample simplifies to

$$\rho_c = (A - A_a) \times 29,6 \times \frac{V_e}{V_s \cdot d} \quad \dots (3)$$

and for the concentration of phaeopigment, equation 2 simplifies to equation (4)

$$\rho_p = A_a \times 20,8 \times \frac{V_e}{V_s \cdot d} - \rho_c \quad \dots (4)$$

NOTES

6 The calculated values for phaeopigments are less reliable than for chlorophyll-a concentrations. Interlaboratory trials have shown an interlaboratory variation of 5 % to 11 % for chlorophyll-a determinations and 6 % to 46 % for phaeopigment determinations.

7 The ratio A/A_a is 1,7, if only undegraded chlorophyll-a is present in the sample. It is 1, if only degradation products of chlorophyll-a are present in the sample.

The operational absorption coefficient (82) for chlorophyll-a in ethanol (4.2) is derived from the operational absorption coefficient, recommended in [5] at 665 nm. The value (84) given there makes allowance for the presence of chlorophyll-b and chlorophyll-c. The absorbance of equal concentrations of chlorophyll-a in acetone is 2 % to 3 % higher than in ethanol at 665 nm [2] [4].

8.4 Note the results, in micrograms per litre (or milligrams per cubic metre), with at most two significant figures or one figure after the decimal point, for example:

Chlorophyll-a concentration 5,5 µg/l
Phaeopigment concentration < 0,1 µg/l

9 Precision

An interlaboratory trial, carried out in 1983, produced the results [6] given in table 1.

10 Test report

The test report shall include the following information:

- a reference to this International Standard;
- identification of the water sample;
- expression of results, in accordance with clause 8;
- sample pretreatment, if relevant;
- any deviation from this method or any other circumstances which may have influenced the results.

Table 1 — Precision data

Sample	<i>l</i>	<i>n</i>	<i>n_a</i> %	$\bar{\bar{x}}$ µg/l	σ_r µg/l	VC _r %	σ_R µg/l	VC _R %
A	18	71	0	126,5	5,46	4,3	6,36	5,0
B	18	70	2,8	20,3	3,67	18,1	2,90	11,3
C	17	67	5,6	24,6	2,21	9,0	2,8	11,4

l is the number of laboratories
n is the number of values
n_a is the percentage of outliers
 $\bar{\bar{x}}$ is the total mean

σ_r is the repeatability standard deviation
VC_r is the repeatability variation coefficient
 σ_R is the reproducibility standard deviation
VC_R is the reproducibility variation coefficient

Annex A (informative)

Determination of phytobenthic concentration

A.1 Determination of phytobenthos

The sampling of phytoplankton is carried out according to ISO 5667-1 and ISO 5667-2.

Sampling of phytobenthos (attached algae) depends on the substratum colonized. A defined surface area on stones or other submerged material may be scratched with a knife or a razor blade. Gravel and small stones are sampled directly into the solvent. The use of artificial substrates exposed to water, such as glass slides, facilitate quantitative sampling. The colonized glass slides or the removed algae are transported from the sampling site with the water. The suspended phytobenthic algae or sediments of known weight are afterwards treated similarly to phytoplankton according to the method described in this International Standard.

A.2 Extracting agent

Acetone has been widely used since it was recommended by UNESCO-SCOR in 1966 for the ex-

traction of algae. However, because it became evident that the extraction efficiency of acetone was rather poor in some cases, alcohols were used. It was shown that hot ethanol and methanol are equally efficient, and considerably more efficient than cold acetone [4]. Special problems may arise, however, from the use of methanol if the chlorophyll extracts are to be acidified in order to correct for interference from phaeopigments. A shifting of the absorption maximum to shorter wavelengths can lead to considerable variation in the maximum acid ratio depending, for example, on the acid concentration, the water content of the solvent and filter, the period between acidification and measurement. These problems can only be overcome by including additional steps in the procedure, for example, neutralization with organic bases (dimethyl aniline or diphenyl aniline). Furthermore, methanol is toxic. For these reasons, it is recommended to use ethanol as a reference extractant for the determination of chlorophyll-a in freshwater and marine samples.

Annex B
(informative)

Bibliography

- [1] TOLSTOY, A. and TOTH, I., Bacteriochlorophyll-d and its interference on determination of chlorophyll-a, *Arch. Hydrobiol.* **89** (1980), pp. 160-170.
- [2] NUSCH, E.A., Comparison of different methods for chlorophyll and phaeopigment determination, *Arch. Hydrobiol. Beih. Ergebn. Limnol.* **14** (1980), pp. 14-36.
- [3] LORENZEN, C.J., Determination of chlorophyll and phaeopigments; spectrophotometric equations, *Limnol, Oceanogr.* **12** (1967), pp. 343-346.
- [4] MARKER, A.F.H., NUSCH, E.A., RAI, H. and RIEMANN, B., The measurement of photosynthetic pigments in freshwaters and standardization of methods: Conclusions and recommendations, *Arch. Hydrobiol. Beih. Ergebn. Limnol.* **14** (1980), pp. 91-106.
- [5] VOLLENWEIDER, R.A., A manual on methods of measuring primary production in aquatic environments, *IBP Handbook No 12*, 2nd ed. (1971) Blackwell Scient. Publ. Oxford, Edinburgh.
- [6] NUSCH, E.A., Results from an interlaboratory ring test concerning the determination of chlorophyll a, *Z. Wasser Abw. Forsch.* **17** (1984), pp. 189-194.

UDC 614.777:556.15:543.42:547.979.7

Descriptors: water, quality, chemical analysis, determination of content, chlorophyll, concentration, spectrometric method.

Price based on 6 pages

British Standards Institution (BSI)

BSI is the national body responsible for preparing British Standards and other standards-related publications, information and services.

BSI is incorporated by Royal Charter. British Standards and other standardization products are published by BSI Standards Limited.

About us

We bring together business, industry, government, consumers, innovators and others to shape their combined experience and expertise into standards-based solutions.

The knowledge embodied in our standards has been carefully assembled in a dependable format and refined through our open consultation process. Organizations of all sizes and across all sectors choose standards to help them achieve their goals.

Information on standards

We can provide you with the knowledge that your organization needs to succeed. Find out more about British Standards by visiting our website at bsigroup.com/standards or contacting our Customer Services team or Knowledge Centre.

Buying standards

You can buy and download PDF versions of BSI publications, including British and adopted European and international standards, through our website at bsigroup.com/shop, where hard copies can also be purchased.

If you need international and foreign standards from other Standards Development Organizations, hard copies can be ordered from our Customer Services team.

Subscriptions

Our range of subscription services are designed to make using standards easier for you. For further information on our subscription products go to bsigroup.com/subscriptions.

With **British Standards Online (BSOL)** you'll have instant access to over 55,000 British and adopted European and international standards from your desktop. It's available 24/7 and is refreshed daily so you'll always be up to date.

You can keep in touch with standards developments and receive substantial discounts on the purchase price of standards, both in single copy and subscription format, by becoming a **BSI Subscribing Member**.

PLUS is an updating service exclusive to BSI Subscribing Members. You will automatically receive the latest hard copy of your standards when they're revised or replaced.

To find out more about becoming a BSI Subscribing Member and the benefits of membership, please visit bsigroup.com/shop.

With a **Multi-User Network Licence (MUNL)** you are able to host standards publications on your intranet. Licences can cover as few or as many users as you wish. With updates supplied as soon as they're available, you can be sure your documentation is current. For further information, email bsmusales@bsigroup.com.

BSI Group Headquarters

389 Chiswick High Road London W4 4AL UK

Revisions

Our British Standards and other publications are updated by amendment or revision.

We continually improve the quality of our products and services to benefit your business. If you find an inaccuracy or ambiguity within a British Standard or other BSI publication please inform the Knowledge Centre.

Copyright

All the data, software and documentation set out in all British Standards and other BSI publications are the property of and copyrighted by BSI, or some person or entity that owns copyright in the information used (such as the international standardization bodies) and has formally licensed such information to BSI for commercial publication and use. Except as permitted under the Copyright, Designs and Patents Act 1988 no extract may be reproduced, stored in a retrieval system or transmitted in any form or by any means – electronic, photocopying, recording or otherwise – without prior written permission from BSI. Details and advice can be obtained from the Copyright & Licensing Department.

Useful Contacts:

Customer Services

Tel: +44 845 086 9001

Email (orders): orders@bsigroup.com

Email (enquiries): cservices@bsigroup.com

Subscriptions

Tel: +44 845 086 9001

Email: subscriptions@bsigroup.com

Knowledge Centre

Tel: +44 20 8996 7004

Email: knowledgecentre@bsigroup.com

Copyright & Licensing

Tel: +44 20 8996 7070

Email: copyright@bsigroup.com



...making excellence a habit.™