

**BS ISO 16063-17:2016**



**BSI Standards Publication**

# **Methods for the calibration of vibration and shock transducers**

Part 17: Primary calibration by centrifuge

**National foreword**

This British Standard is the UK implementation of ISO 16063-17:2016.

The UK participation in its preparation was entrusted to Technical Committee GME/21/2, Mechanical vibration, shock and condition monitoring - Vibration and shock measuring instruments and testing equipment.

A list of organizations represented on this committee can be obtained on request to its secretary.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

© The British Standards Institution 2016.  
Published by BSI Standards Limited 2016

ISBN 978 0 580 90054 9

ICS 17.160

**Compliance with a British Standard cannot confer immunity from legal obligations.**

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 30 June 2016.

**Amendments issued since publication**

Date	Text affected
------	---------------

---

---

---

**Methods for the calibration of  
vibration and shock transducers —**

**Part 17:  
Primary calibration by centrifuge**

*Méthodes pour l'étalonnage des transducteurs de vibrations et de  
chocs —*

*Partie 17: Étalonnage primaire par centrifugeur*



**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2016, Published in Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
Ch. de Blandonnet 8 • CP 401  
CH-1214 Vernier, Geneva, Switzerland  
Tel. +41 22 749 01 11  
Fax +41 22 749 09 47  
copyright@iso.org  
www.iso.org

# Contents

Page

<b>Foreword</b> .....	<b>iv</b>
<b>1 Scope</b> .....	<b>1</b>
<b>2 Normative references</b> .....	<b>1</b>
<b>3 Terms and definitions</b> .....	<b>1</b>
<b>4 Requirements for apparatus and environmental conditions</b> .....	<b>1</b>
4.1 Calibration environment.....	1
4.2 Balanced table or arm (rotational table).....	1
4.3 Rotational frequency measuring instrumentation.....	2
4.4 Voltage-measuring instrumentation.....	3
<b>5 Preferred values</b> .....	<b>3</b>
<b>6 Method 1 (with measurement of the radius of rotation)</b> .....	<b>3</b>
6.1 Test procedure.....	3
6.2 Expression of results.....	3
<b>7 Method 2 (without measurement of the radius of rotation)</b> .....	<b>4</b>
7.1 Test procedure.....	4
7.2 Expression of results.....	4
<b>Annex A (informative) Expression of uncertainty of measurement in calibration</b> .....	<b>5</b>
<b>Bibliography</b> .....	<b>7</b>

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see [www.iso.org/directives](http://www.iso.org/directives)).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see [www.iso.org/patents](http://www.iso.org/patents)).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: [www.iso.org/iso/foreword.html](http://www.iso.org/iso/foreword.html).

The committee responsible for this document is ISO/TC 108, *Mechanical vibration, shock and condition monitoring*, Subcommittee SC 3, *Use and calibration of vibration and shock measuring instruments*.

This first edition of ISO 16063-17 cancels and replaces ISO 5347-7:1993, which has been technically revised.

ISO 16063 consists of the following parts, under the general title *Methods for the calibration of vibration and shock transducers*:

- *Part 1: Basic concepts*
- *Part 11: Primary vibration calibration by laser interferometry*
- *Part 12: Primary vibration calibration by the reciprocity method*
- *Part 13: Primary shock calibration using laser interferometry*
- *Part 15: Primary angular vibration calibration by laser interferometry*
- *Part 16: Calibration by Earth's gravitation*
- *Part 17: Primary calibration by centrifuge*
- *Part 21: Vibration calibration by comparison to a reference transducer*
- *Part 22: Shock calibration by comparison to a reference transducer*
- *Part 31: Testing of transverse vibration sensitivity*
- *Part 32: Resonance testing — Testing the frequency and the phase response of accelerometers by means of shock excitation*
- *Part 41: Calibration of laser vibrometers*
- *Part 42: Calibration of seismometers with high accuracy using acceleration of gravity*

— *Part 43: Calibration of accelerometers by model-based parameter identification*

The following parts are under preparation:

— *Part 33: Testing of magnetic field sensitivity*





# Methods for the calibration of vibration and shock transducers —

## Part 17: Primary calibration by centrifuge

### 1 Scope

ISO 16063 comprises a series of documents dealing with methods for the calibration of vibration and shock transducers.

This part of ISO 16063 lays down detailed specifications for the instrumentation and procedure to be used for primary calibration of accelerometers using centrifuge calibration.

This part of ISO 16063 is applicable to rectilinear accelerometers with zero-frequency response, mainly of the strain gauge or piezoresistive type, and to primary standard and working transducers.

It is applicable for a calibration range from 10 m/s<sup>2</sup> to 20 000 m/s<sup>2</sup> (higher accelerations possible) at 0 Hz.

The limits of uncertainty applicable are  $\pm 1$  % of reading.

### 2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 2041, *Mechanical vibration, shock and condition monitoring — Vocabulary*

### 3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 2041 apply.

### 4 Requirements for apparatus and environmental conditions

#### 4.1 Calibration environment

The standard reference atmospheric conditions are:  $(23 \pm 3)$  °C and 75 % relative humidity max. The temperature, humidity and atmospheric pressure shall be measured and reported.

#### 4.2 Balanced table or arm (rotational table)

The main component of the calibration apparatus consists of balanced table or arm which rotates about a vertical axis with uniform angular speed. For the calibration range from 10 m/s<sup>2</sup> to 100 m/s<sup>2</sup>, the table/arm shall be level within  $\pm 0,5^\circ$  of horizontal. For ranges higher than 100 m/s<sup>2</sup>, levelling is allowed to within  $\pm 2^\circ$ . The calibration apparatus shall be placed on a sufficiently heavy base which is sufficiently isolated from the floor vibration.


The rotational frequency shall be uniform within  $\pm 0,05$  % of the nominal value.

The transducers axis of sensitivity shall be aligned within  $\pm 0,5^\circ$ .

The radius of rotation to the centre of the transducer mass element shall be measured with an uncertainty less than  $\pm 0,15\%$ . If this is not possible, Method 2 shall be used (see [Clause 7](#)).

If the accelerometer is substituted by impedances not sensitive to acceleration, the hum and noise when the centrifuge is rotating at the calibration speeds shall be at least 0,5 % below reading.

The apparatus of the main component is illustrated by [Figure 1](#).



- Key**
- 1 signal conditioner for transducers
  - 2 balanced table
  - 3 drive motor
  - 4 bearing
  - 5 vibration isolation stage
  - 6 signal coupler
  - 7 stone plate
  - 8 transducers
  - 9 cable
  - 10 centre of mass
  - 11 rotation centre

**Figure 1 — Illustration of the apparatus of the main component**

### 4.3 Rotational frequency measuring instrumentation

The relative expanded measurement uncertainty ( $k = 2$ ) contribution of the measurement of rotational frequency shall be 0,05 % or less.

#### 4.4 Voltage-measuring instrumentation

The relative expanded measurement uncertainty ( $k = 2$ ) contribution of the voltmeter shall be 0,01 % or less.

### 5 Preferred values

Six acceleration values, in metres per second squared, equally covering the accelerometer range, shall be chosen from the following series: 10; 20; 50; 100; 200; 500 or their multiples of ten. The reference acceleration shall be 100 m/s<sup>2</sup> (second choice: 50 m/s<sup>2</sup>).

## 6 Method 1 (with measurement of the radius of rotation)

### 6.1 Test procedure

Rotate the table or arm at different frequencies determined by calculation from the standard levels using [Formula \(1\)](#):

$$a = 4\pi^2 n^2 r \quad (1)$$

where

$n$  is the rotational frequency, in hertz;

$r$  is the radius of rotation to the centre of the accelerometer mass element, in metres.

Measure the transducer output at each rotational frequency.

Determine the reference calibration factor at the reference acceleration. Then, determine the sensitivity for the other calibration amplitudes. The results shall be given as a deviation from the reference calibration factor in percent.

### 6.2 Expression of results

The calibration factor,  $S$ , in volts per (metre per second squared) [V/(m/s<sup>2</sup>)], is given by [Formula \(2\)](#):

$$S = \frac{V}{4\pi^2 n^2 r} \quad (2)$$

where

$V$  is the transducer output, in volts;

$n$  is the rotational frequency, in hertz;

$r$  is the radius of rotation to the centre of the accelerometer mass element, in metres.

When the calibration results are reported, the total uncertainty of the calibration and the corresponding confidence level, calculated in accordance with [Annex A](#), shall also be reported.

## 7 Method 2 (without measurement of the radius of rotation)

### 7.1 Test procedure

If the rotational radius cannot be measured with the specified uncertainty, the transducer can be rotated in two different positions, the radial distance between these to be measured with uncertainty maximum of  $\pm 0,5\%$ .

Measure the two rotational frequencies giving the same transducer output at the two locations.

### 7.2 Expression of results

The calibration factor,  $S$ , in volts per (metre per second squared), is given by [Formula \(3\)](#):

$$S = \frac{V}{4\pi^2 n_2^2 \left| \frac{\Delta r}{1 - (n_2 / n_1)^2} \right|} \quad (3)$$

where

$V$  is the accelerometer output, in volts;

$n_1$  is the rotational frequency at the first accelerometer position, in hertz;

$n_2$  is the rotational frequency at the second accelerometer position, in hertz;

$\Delta r$  is the distance between the two accelerometer positions, in metres.

When the calibration results are reported, the total uncertainty of the calibration and the corresponding confidence level, calculated in accordance with [Annex A](#), shall also be reported.

## Annex A (informative)

### Expression of uncertainty of measurement in calibration

#### A.1 Uncertainty analysis on Method 1

Components of uncertainty that contribute to the uncertainty of the measurement on Method 1 are listed in [Table A.1](#). The combined standard uncertainty,  $u_c$ , can be expressed by [Formula \(A.1\)](#):

$$u_c = \sqrt{\sum_{i=1}^{10} (u_i)^2} \quad (\text{A.1})$$

where  $u_i$  is uncertainty component in [Table A.1](#), respectively.

The expanded uncertainty,  $U$ , shall be determined by multiplying  $u_c$  by coverage factor,  $k$ , where a value of  $k = 2$ .

**Table A.1 — Uncertainty components on Method 1**

$i$	Standard uncertainty component	Source of uncertainty	Relative uncertainty contribution $u_i(y)$
Acceleration generation			
1	$u(\alpha)$	Alignment effect for rotational plane $\alpha$ : deviation from rotational plane	$u_1(y) = \frac{9,8(1 - \cos\alpha)}{a}$
2	$u(\beta)$	Alignment effect of transducer axe $\beta$ : alignment deviation from transducer axe	$u_2(y) = (1 - \cos\beta)$
3	$u(e_n)$	Rotational frequency effect $e_n$ : rotational frequency deviation from nominal value	$u_3(y) = 2e_n / n$
4	$u(e_{\Delta n})$	Rotational frequency effect (cyclic) $e_{\Delta n}$ : cyclic rotational frequency deviation	$u_4(y) = 2e_{\Delta n} / n$
5	$u(e_r)$	Rotational radius deviation effect $e_r$ : deviation of radius	$u_5(y) = e_r / r$
Voltage measurement			
6	$u(e_V)$	Uncertainty of voltage measurement	$u_6(y)$
7	$u(a_H)$	Hum and noise effect	$u_7(y)$
8	$u(e_P)$	Effect of power supply	$u_8(y)$
Other factors			
9	$u(S_T)$	Effect of temperature	$u_9(y)$
10	$u(S_{\text{hys}})$	Non linearity, hysteresis	$u_{10}(y)$

## A.2 Uncertainty analysis on Method 2

Components of uncertainty that contribute to the uncertainty of the measurement on Method 2 are listed in [Table A.2](#). The combined standard uncertainty,  $u_c$ , can be expressed by [Formula \(A.2\)](#):

$$u_c = \sqrt{\sum_{i=1}^{10} (u_i)^2} \quad (\text{A.2})$$

where  $u_i$  is uncertainty component in [Table A.2](#), respectively.

The expanded uncertainty,  $U$ , shall be determined by multiplying,  $u_c$ , by coverage factor,  $k$ , where a value of  $k = 2$ .

**Table A.2 — Uncertainty components on Method 2**

$i$	Standard uncertainty component	Source of uncertainty	Relative uncertainty contribution $u_i(y)$
Acceleration generation			
1	$u(\alpha)$	Alignment effect for rotational plane $\alpha$ : deviation from rotational plane	$u_1(y) = \frac{9,8(1 - \cos\alpha)}{a}$
2	$u(\beta)$	Alignment effect of transducer axe $\beta$ : alignment deviation from transducer axe	$u_2(y) = (1 - \cos\beta)$
3	$u(e_{n1})$	Rotational frequency effect at the first transducer position $e_{n1}$ : rotational frequency deviation from nominal value at rotational $n_2$ frequency $n_1$	$u_3(y) = 2e_{n1} / n_1$
4	$u(e_{n2})$	Rotational frequency effect at the second transducer position $e_{n2}$ : rotational frequency deviation from nominal value at rotational frequency	$u_4(y) = 2e_{n2} / n_2$
5	$u(e_{\Delta r})$	Distance measurement between the two transducer positions effect $e_{\Delta r}$ : deviation of measurement between the positions, $\Delta r$ : Distance between the positions	$u_5(y) = e_{\Delta r} / \Delta r$
Voltage measurement			
6	$u(e_V)$	Uncertainty of voltage measurement	$u_6(y)$
7	$u(a_H)$	Hum and noise effect	$u_7(y)$
8	$u(e_P)$	Effect of power supply	$u_8(y)$
Other factors			
9	$u(S_T)$	Effect of temperature	$u_9(y)$
10	$u(S_{hys})$	Non linearity, hysteresis	$u_{10}(y)$

## Bibliography

- [1] ISO 8042, *Shock and vibration measurements — Characteristics to be specified for seismic pick-ups*
- [2] ISO 16063-1, *Methods for the calibration of vibration and shock transducers — Part 1: Basic concepts*
- [3] ISO/IEC Guide 98-3, *Uncertainty of measurement — Part 3: Guide to the expression of uncertainty in measurement (GUM:1995)<sup>1)</sup>*

---

1) Reissue of the *Guide to the expression of uncertainty in measurement (GUM)*, 1995.







# British Standards Institution (BSI)

BSI is the national body responsible for preparing British Standards and other standards-related publications, information and services.

BSI is incorporated by Royal Charter. British Standards and other standardization products are published by BSI Standards Limited.

## About us

We bring together business, industry, government, consumers, innovators and others to shape their combined experience and expertise into standards-based solutions.

The knowledge embodied in our standards has been carefully assembled in a dependable format and refined through our open consultation process. Organizations of all sizes and across all sectors choose standards to help them achieve their goals.

## Information on standards

We can provide you with the knowledge that your organization needs to succeed. Find out more about British Standards by visiting our website at [bsigroup.com/standards](http://bsigroup.com/standards) or contacting our Customer Services team or Knowledge Centre.

## Buying standards

You can buy and download PDF versions of BSI publications, including British and adopted European and international standards, through our website at [bsigroup.com/shop](http://bsigroup.com/shop), where hard copies can also be purchased.

If you need international and foreign standards from other Standards Development Organizations, hard copies can be ordered from our Customer Services team.

## Copyright in BSI publications

All the content in BSI publications, including British Standards, is the property of and copyrighted by BSI or some person or entity that owns copyright in the information used (such as the international standardization bodies) and has formally licensed such information to BSI for commercial publication and use.

Save for the provisions below, you may not transfer, share or disseminate any portion of the standard to any other person. You may not adapt, distribute, commercially exploit, or publicly display the standard or any portion thereof in any manner whatsoever without BSI's prior written consent.

## Storing and using standards

Standards purchased in soft copy format:

- A British Standard purchased in soft copy format is licensed to a sole named user for personal or internal company use only.
- The standard may be stored on more than 1 device provided that it is accessible by the sole named user only and that only 1 copy is accessed at any one time.
- A single paper copy may be printed for personal or internal company use only.

Standards purchased in hard copy format:

- A British Standard purchased in hard copy format is for personal or internal company use only.
- It may not be further reproduced – in any format – to create an additional copy. This includes scanning of the document.

If you need more than 1 copy of the document, or if you wish to share the document on an internal network, you can save money by choosing a subscription product (see 'Subscriptions').

## Reproducing extracts

For permission to reproduce content from BSI publications contact the BSI Copyright & Licensing team.

## Subscriptions

Our range of subscription services are designed to make using standards easier for you. For further information on our subscription products go to [bsigroup.com/subscriptions](http://bsigroup.com/subscriptions).

With **British Standards Online (BSOL)** you'll have instant access to over 55,000 British and adopted European and international standards from your desktop. It's available 24/7 and is refreshed daily so you'll always be up to date.

You can keep in touch with standards developments and receive substantial discounts on the purchase price of standards, both in single copy and subscription format, by becoming a **BSI Subscribing Member**.

**PLUS** is an updating service exclusive to BSI Subscribing Members. You will automatically receive the latest hard copy of your standards when they're revised or replaced.

To find out more about becoming a BSI Subscribing Member and the benefits of membership, please visit [bsigroup.com/shop](http://bsigroup.com/shop).

With a **Multi-User Network Licence (MUNL)** you are able to host standards publications on your intranet. Licences can cover as few or as many users as you wish. With updates supplied as soon as they're available, you can be sure your documentation is current. For further information, email [subscriptions@bsigroup.com](mailto:subscriptions@bsigroup.com).

## Revisions

Our British Standards and other publications are updated by amendment or revision.

We continually improve the quality of our products and services to benefit your business. If you find an inaccuracy or ambiguity within a British Standard or other BSI publication please inform the Knowledge Centre.

## Useful Contacts

### Customer Services

**Tel:** +44 345 086 9001

**Email (orders):** [orders@bsigroup.com](mailto:orders@bsigroup.com)

**Email (enquiries):** [cservices@bsigroup.com](mailto:cservices@bsigroup.com)

### Subscriptions

**Tel:** +44 345 086 9001

**Email:** [subscriptions@bsigroup.com](mailto:subscriptions@bsigroup.com)

### Knowledge Centre

**Tel:** +44 20 8996 7004

**Email:** [knowledgecentre@bsigroup.com](mailto:knowledgecentre@bsigroup.com)

### Copyright & Licensing

**Tel:** +44 20 8996 7070

**Email:** [copyright@bsigroup.com](mailto:copyright@bsigroup.com)

### BSI Group Headquarters

389 Chiswick High Road London W4 4AL UK