

BS ISO 18158:2016



BSI Standards Publication

# Workplace air — Terminology

**bsi.**

...making excellence a habit.™

**National foreword**

This British Standard is the UK implementation of ISO 18158:2016.

The UK participation in its preparation was entrusted to Technical Committee EH/2/2, Work place atmospheres.

A list of organizations represented on this committee can be obtained on request to its secretary.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

© The British Standards Institution 2016.  
Published by BSI Standards Limited 2016

ISBN 978 0 580 80283 6

ICS 01.020; 13.040.30

**Compliance with a British Standard cannot confer immunity from legal obligations.**

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 31 March 2016.

**Amendments issued since publication**

Date	Text affected
------	---------------

---

INTERNATIONAL  
STANDARD

ISO  
18158

First edition  
2016-03-01

---

---

**Workplace air — Terminology**

*Qualité de l'air — Terminologie*



Reference number  
ISO 18158:2016(E)

© ISO 2016



## **COPYRIGHT PROTECTED DOCUMENT**

© ISO 2016, Published in Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
Ch. de Blandonnet 8 • CP 401  
CH-1214 Vernier, Geneva, Switzerland  
Tel. +41 22 749 01 11  
Fax +41 22 749 09 47  
copyright@iso.org  
www.iso.org

# Contents

Page

<b>Foreword</b> .....	<b>iv</b>
<b>Introduction</b> .....	<b>v</b>
<b>1 Scope</b> .....	<b>1</b>
<b>2 Terms and definitions</b> .....	<b>1</b>
2.1 General terms .....	1
2.1.1 Agents.....	1
2.1.2 Air pollutants.....	1
2.1.3 Health-related fractions and conventions.....	2
2.1.4 Aerosols.....	3
2.1.5 Exposure assessment.....	5
2.1.6 Other terms .....	6
2.2 Terms related to the physical and chemical processes of workplace (air) sampling.....	6
2.2.1 Workplace (air) sample.....	6
2.2.2 Workplace (air) sampler .....	7
2.2.3 Workplace (air) sampling .....	10
2.3 Terms related to the analytical method.....	12
2.4 Terms related to method performance.....	14
2.4.1 Efficiency .....	14
2.4.2 Uncertainty.....	14
2.4.3 General statistical terms .....	17
2.4.4 Other statistical terms.....	19
<b>Annex A (informative) Alphabetical index of terms defined</b> .....	<b>20</b>
<b>Annex B (informative) Additional terms</b> .....	<b>24</b>
<b>Bibliography</b> .....	<b>28</b>

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see [www.iso.org/directives](http://www.iso.org/directives)).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see [www.iso.org/patents](http://www.iso.org/patents)).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the WTO principles in the Technical Barriers to Trade (TBT) see the following URL: [Foreword - Supplementary information](#)

The committee responsible for this document is ISO/TC 146, *Air quality*, Subcommittee SC 2, *Workplace atmospheres*.

## Introduction

The health of workers in many industries is at risk through exposure by inhalation of chemical and biological agents. Industrial hygienists and other public health professionals need to determine the effectiveness of measures taken to control workers' exposure and this is generally achieved by making workplace air measurements.

Many terms that are commonly used in relation to workplace air measurements and exposure assessments are defined within individual standards or technical documents and are often defined differently from one standard or document to the next. This creates ambiguities and inconsistencies in the use of such terms. This International Standard was developed to ensure that commonly used terms have agreed-upon definitions and to eliminate ambiguities and inconsistencies in their usage. It will be of benefit to agencies concerned with health and safety at work, industrial hygienists and other public health professionals, analytical laboratories, industrial users of metals and metalloids and their workers.





# Workplace air — Terminology

## 1 Scope

This International Standard specifies terms and definitions that are related to the assessment of workplace exposure (see 2.1.5.1) to chemical and biological agents (see 2.1.1.1). These are either general terms or are specific to physical and chemical processes of air sampling, the analytical method (see 2.3.3), or method performance.

The terms included are those that have been identified as being fundamental because their definition is necessary to avoid ambiguity and ensure consistency of use.

This International Standard is applicable to all International Standards, ISO Technical Reports, ISO Technical Specifications, and ISO Guides related to workplace atmospheres.

## 2 Terms and definitions

### 2.1 General terms

#### 2.1.1 Agents

##### 2.1.1.1

##### **biological agent**

one of a number of agents such as bacteria, viruses, fungi and other micro-organisms or parts of them and their associated toxins, including those which have been genetically modified, cell cultures or endoparasites which are potentially hazardous to human health

Note 1 to entry: Dusts of organic origin, for example, cotton dust, flour dust and wood dust, are not considered to be biological agents and are therefore not covered by this definition.

[SOURCE: EN 1540:2011, 2.1.1, modified — Added “one of a number of agents such as” to definition and changed “pollen” to “cotton dust”.]

##### 2.1.1.2

##### **chemical agent**

chemical element or compound on its own or admixed as it occurs in the natural state or as produced, used, or released, including release as waste, by any work activity, whether or not produced intentionally and whether or not placed on the market

[SOURCE: Council Directive 98/24/EC, Art. 2 a]

#### 2.1.2 Air pollutants

##### 2.1.2.1

##### **air pollutant**

material emitted into the atmosphere either by human activity or natural processes and adversely affecting humans or the environment

[SOURCE: EN 1540:2011, 2.2.1, modified — Changed “man” to “humans”.]

##### 2.1.2.2

##### **airborne dust**

finely divided matter, in solid form, dispersed in air

[SOURCE: EN 1540:2011, 2.2.2]

### 2.1.2.3

#### **airborne particles**

fine matter, in solid or liquid form, dispersed in air

[SOURCE: EN 1540:2011, 2.2.3, modified — Note 1 to entry deleted.]

### 2.1.2.4

#### **total airborne particles**

*airborne particles* (2.1.2.3) present in a given volume of air

[SOURCE: EN 1540:2011, 2.2.6, modified — Added “present” after “particles”.]

### 2.1.2.5

#### **vapour**

gas phase of a substance in a state of equilibrium or disturbed equilibrium with the same substance in a liquid or solid state below its boiling or sublimation point

[SOURCE: EN 1540:2011, 2.2.7]

## 2.1.3 Health-related fractions and conventions

### 2.1.3.1

#### **health-related fractions**

<airborne particles> collective term for the fractions of *airborne particles* (2.1.2.3) penetrating to different regions of the respiratory tract, i.e. the inhalable, thoracic and respirable fractions

Note 1 to entry: The health-related fractions of airborne particles are specified in ISO 7708.

[SOURCE: EN 1540:2011, 2.3.1, modified — citation in Note 1 to entry changed from EN 481 to ISO 7708.]

#### 2.1.3.1.1

##### **inhalable fraction**

mass fraction of *total airborne particles* (2.1.2.4) which is inhaled through the nose and mouth

Note 1 to entry: The inhalable fraction depends on the speed and direction of the air movement, on the rate of breathing and other factors.

[SOURCE: ISO 7708:1995, 2.3]

#### 2.1.3.1.2

##### **extrathoracic fraction**

mass fraction of *total airborne particles* (2.1.2.4) which fail to penetrate beyond the larynx

[SOURCE: ISO 7708:1995, 2.5, modified]

#### 2.1.3.1.3

##### **thoracic fraction**

mass fraction of *total airborne particles* (2.1.2.4) which penetrate beyond the larynx

[SOURCE: ISO 7708:1995, 2.7, modified]

#### 2.1.3.1.4

##### **respirable fraction**

mass fraction of *total airborne particles* (2.1.2.4) which penetrate to the unciliated airways

[SOURCE: ISO 7708:1995, 2.11, modified]

### 2.1.3.2

#### **sampling convention**

target specification for sampling instruments for a specified airborne particulate fraction

#### 2.1.3.2.1

##### **inhalable convention**

target specification for sampling instruments when the *inhalable fraction* ([2.1.3.1.1](#)) is the fraction of interest

[SOURCE: ISO 7708:1995, 2.4]

#### 2.1.3.2.2

##### **extrathoracic convention**

target specification for sampling instruments when the *extrathoracic fraction* ([2.1.3.1.2](#)) is the fraction of interest

[SOURCE: ISO 7708:1995, 2.6, modified — “the fraction” added before “of interest” for consistency of wording.]

#### 2.1.3.2.3

##### **thoracic convention**

target specification for sampling instruments when the *thoracic fraction* ([2.1.3.1.3](#)) is the fraction of interest

[SOURCE: ISO 7708:1995, 2.8, modified — “the fraction” added before “of interest” for consistency of wording.]

#### 2.1.3.2.4

##### **respirable convention**

target specification for sampling instruments when the *respirable fraction* ([2.1.3.1.4](#)) is the fraction of interest

[SOURCE: ISO 7708:1995, 2.12, modified — “the fraction” added before “of interest” for consistency of wording.]

### 2.1.4 Aerosols

#### 2.1.4.1

##### **aerosol**

*airborne particles* ([2.1.2.3](#)) and the gas (and vapour) mixture in which they are suspended

Note 1 to entry: The airborne particles can be in or out of equilibrium with their own *vapours* ([2.1.2.5](#)).

[SOURCE: EN 1540:2011, 2.2.4]

#### 2.1.4.2

##### **bioaerosol**

*aerosol* ([2.1.4.1](#)) consisting of (a) *biological agent(s)* ([2.1.1.1](#))

Note 1 to entry: *Airborne dusts* ([2.1.2.2](#)) of organic origin, for example, cotton dust, flour dust and wood dust, are not considered to be bioaerosols and are therefore not covered by this definition.

[SOURCE: EN 1540:2011, 2.2.5, modified — “pollen” changed to “cotton dust” in Note 1 to entry.]

#### 2.1.4.3

##### **nanoaerosol**

*aerosol* ([2.1.4.1](#)) comprised of, or consisting of, *nanoparticles* ([2.1.4.4](#)) and *nanostructured particles* ([2.1.4.5](#))

[SOURCE: ISO/TR 27628:2007, 2.11]

#### 2.1.4.4

##### **nanoparticle**

material with all three dimensions in the size range from approximately 1 nm to 100 nm

[SOURCE: ISO/TS 80004-4:2011, 2.1, 2.2, and 2.4, modified]

#### 2.1.4.5

##### **nanostructured particle**

particle with structural features smaller than 100 nm, which can influence its physical, chemical and/or biological properties

Note 1 to entry: A nanostructured particle can have a maximum dimension substantially larger than 100 nm.

EXAMPLE A 500 nm diameter *agglomerate* (2.1.4.9) of *nanoparticles* (2.1.4.4) would be considered a nanostructured particle.

[SOURCE: ISO/TR 27628:2007, 2.13, modified — “may” changed to “can” in definition and Note 1 to entry.]

#### 2.1.4.6

##### **ultrafine aerosol**

*aerosol* (2.1.4.1) consisting predominantly of *ultrafine particles* (2.1.4.7)

Note 1 to entry: The term is often used in the context of particles produced as a by-product of a process (incidental particles), such as welding *fume* (B.2.2) and combustion fume

[SOURCE: ISO/TR 27628:2007, 2.20]

#### 2.1.4.7

##### **ultrafine particle**

particle with a nominal diameter (such as geometric, aerodynamic, mobility, projected-area or otherwise) of 100 nm or less, produced as a by-product of a process such as welding and combustion

[SOURCE: ISO/TR 27628:2007, 2.21, modified — Note 1 to entry moved to definition text.]

#### 2.1.4.8

##### **particle aerodynamic diameter**

##### **aerodynamic diameter**

diameter of a sphere of 1 g cm<sup>-3</sup> density with the same terminal settling velocity in calm air as the particle, under the prevailing conditions of temperature, pressure and relative humidity

Note 1 to entry: The particle aerodynamic diameter depends on the size, density and shape of the particle.

Note 2 to entry: Aerodynamic diameter is related to the inertial properties of aerosol particles.

[SOURCE: EN 1540:2011, 2.3.2, modified — Note 2 to entry derived from ISO/TR 27628:2007, 2.2.]

#### 2.1.4.9

##### **agglomerate**

<aerosols> group of particles held together by relatively weak forces, including van der Waals forces, electrostatic forces and surface tension

[SOURCE: ISO/TR 27628:2007, 2.4 modified — Note 1 to entry deleted.]

#### 2.1.4.10

##### **aggregate**

<aerosols> heterogeneous particle in which the various components are held together by relatively strong forces and thus not easily broken apart

[SOURCE: ISO/TR 27628:2007, 2.5, modified – Note 1 to entry deleted.]

#### 2.1.4.11

##### **coagulation**

formation of larger particles through the collision and subsequent adhesion of smaller particles

[SOURCE: ISO/TR 27628:2007, 2.6]

#### 2.1.4.12

##### **coalescence**

formation of homogeneous particles through the collision of smaller liquid particles and subsequent merging or mixing of constituent material

[SOURCE: ISO/TR 27628:2007, 2.7, modified — “liquid” added.]

### 2.1.5 Exposure assessment

#### 2.1.5.1

##### **exposure**

<inhalation> situation in which a *chemical agent* (2.1.1.2) or *biological agent* (2.1.1.1) is present in the air that is inhaled by a person

[SOURCE: EN 1540:2011, 2.4.1, modified — “agent” added after “chemical” to facilitate cross-referencing.]

#### 2.1.5.2

##### **dermal exposure**

contact between a *chemical agent* (2.1.1.2) or *biological agent* (2.1.1.1) and human skin

[SOURCE: EN 1540:2011, 2.4.2]

#### 2.1.5.3

##### **time-weighted average concentration**

##### **TWA concentration**

concentration of a *chemical agent* (2.1.1.2) in the air, averaged over a *reference period* (2.1.5.7)

[SOURCE: ISO 21438-1:2007, 3.1.7, modified — Replaced “atmosphere” with “air”.]

#### 2.1.5.4

##### **occupational exposure limit value**

##### **OELV**

limit of the time-weighted average of the concentration of a *chemical agent* (2.1.1.2) in the air within the *breathing zone* (2.1.5.5) of a worker in relation to a specified *reference period* (2.1.5.7)

Note 1 to entry: The term “limit value” is often used as a synonym for “occupational exposure limit value” but the term “occupational exposure limit value” is preferred because there is more than one limit value (e.g. biological limit value and occupational exposure limit value).

Note 2 to entry: Occupational exposure limit values (OELVs) are often set for reference periods of 8 h but can also be set for shorter periods or concentration excursions. OELVs for gases and *vapours* (2.1.2.5) are stated in terms independent of temperature and air pressure variables in ml/m<sup>3</sup> and in terms dependent on those variables in mg/m<sup>3</sup> for a temperature of 20 °C and a pressure of 101,3 kPa. OELVs for *airborne particles* (2.1.2.3) and mixtures of particles and vapours are given in mg/m<sup>3</sup> or multiples of that for actual environmental conditions (temperature, pressure) at the *workplace* (2.1.6.2). OELVs of fibres are given in number of fibres/m<sup>3</sup> or number of fibres/cm<sup>3</sup> for actual environmental conditions (temperature, pressure) at the workplace.

[SOURCE: Council Directive 98/24/EC, Art. 2 d, modified — “OELV” and new Note 1 to entry added; original Note 1 becomes Note 2; “limit value” changed to “occupational exposure limit value” or “OELV” in Note 2 to entry; “mostly” changed to “often” in first sentence of Note 2 to entry.]

#### 2.1.5.5

##### **breathing zone**

space around the nose and mouth from which breath is taken

Note 1 to entry: Technically, the breathing zone corresponds to a hemisphere (generally accepted to be 30 cm in radius) extending in front of the human face, centred on the midpoint of a line joining the ears. The base of the hemisphere is a plane through this line, the top of the head and the larynx. This technical description is not applicable when respiratory protective equipment is used.

[SOURCE: EN 1540:2011, 2.4.5]

#### 2.1.5.6

**measuring procedure**  
**measurement procedure**  
**measurement method**

set of operations described specifically for the sampling and *analysis* (2.3.1) of *chemical agents* (2.1.1.2) or *biological agents* (2.1.1.1) in air

Note 1 to entry: A measuring procedure usually includes preparation for sampling, conducting the sampling, transportation and storage, and *sample preparation* (2.3.11) for analysis and conducting the analysis.

[SOURCE: EN 1540:2011, 2.4.6, modified — “agents” added after “chemical” to facilitate cross-referencing and Note 1 to entry editorially modified.]

#### 2.1.5.7

**reference period**

specified period of time for which the *occupational exposure limit value* (2.1.5.4) of a *chemical agent* (2.1.1.2) or *biological agent* (2.1.1.1) applies

Note 1 to entry: The reference period is usually 8 h for long-term measurements and 15 min for short-term measurements.

[SOURCE: EN 1540:2011, 2.4.7, modified — “agent” added after “chemical” to facilitate cross-referencing.]

### 2.1.6 Other terms

#### 2.1.6.1

**dustiness**

propensity of materials to produce *airborne dust* (2.1.2.2) during handling

Note 1 to entry: Dustiness is not an intrinsic property as it depends on how it is measured.

[SOURCE: EN 1540:2011, 2.5.1]

#### 2.1.6.2

**workplace**

designated area or areas in which the work activities are carried out

[SOURCE: EN 1540:2011, 2.5.2]

## 2.2 Terms related to the physical and chemical processes of workplace (air) sampling

### 2.2.1 Workplace (air) sample

#### 2.2.1.1

**sample**  
**air sample**

<workplace air quality> product of the process of (*air*) *sampling* (2.2.3.1)

Note 1 to entry: An air sample is frequently considered to include the *collection substrate(s)* (2.2.3.7) as well as the collected *chemical agents* (2.1.1.2) and/or *biological agents* (2.1.1.1) or sometimes it is considered to be the fractional part of a larger volume of air.

[SOURCE: EN 1540:2011, 3.1.1, modified — “<workplace air quality>” added as the domain for the definition. In Note 1 to entry, “agents” added after “chemical” to facilitate cross-referencing.]

#### 2.2.1.2

##### **collected sample**

product of the process of *(air) sampling* (2.2.3.1) that consists of the collected *chemical agents* (2.1.1.2) and/or *biological agents* (2.1.1.1) only

[SOURCE: EN 1540:2011, 3.1.2, modified — “Agents” added after “chemical” to facilitate cross-referencing.]

#### 2.2.1.3

##### **personal sample**

product of the process of using a *sampler* (2.2.2.1), attached to a person, to collect gases, *vapours* (2.1.2.5), and/or *airborne particles* (2.1.2.3) in the *breathing zone* (2.1.5.5) for the purpose of measuring *exposure* (2.1.5.1) to *chemical agents* (2.1.1.2) and/or *biological agents* (2.1.1.1)

#### 2.2.1.4

##### **static sample**

##### **area sample**

product of using a *sampler* (2.2.2.1) in a stationary location that collects gases, *vapours* (2.1.2.5) and/or *airborne particles* (2.1.2.3) for the purpose of measuring *exposure* (2.1.5.1) to *chemical agents* (2.1.1.2) and/or *biological agents* (2.1.1.1)

### 2.2.2 Workplace (air) sampler

#### 2.2.2.1

##### **sampler**

##### **air sampler**

<workplace air quality> device for separating and/or collecting *chemical agents* (2.1.1.2) and/or *biological agents* (2.1.1.1) from the surrounding air

Note 1 to entry: (Air) samplers are generally designed for a particular purpose, e.g. for sampling gases and *vapours* (2.1.2.5) or for sampling airborne particles.

[SOURCE: EN 1540:2011, 3.2.1, modified — “<workplace air quality>” added as the domain for the definition. “Separating” changed to “separating and/or collecting”. “Agents” added after “chemical” to facilitate cross-referencing.]

#### 2.2.2.1.1

##### **passive sampler**

*sampler* (2.2.2.1) that collects gases, *vapours* (2.1.2.5) or *airborne particles* (2.1.2.3) on a *collection substrate* (2.2.3.7) without active air movement

Note 1 to entry: Passive samplers include *diffusive samplers* (2.2.2.1.2) for collection of gases and vapours and samplers for collection of airborne particles based on turbulent diffusion and separation by electrical or other forces.

[SOURCE: EN 1540:2011, 3.2.1.1]

#### 2.2.2.1.2

##### **diffusive sampler**

*passive sampler* (2.2.2.1.1) that collects gases or *vapours* (2.1.2.5) at a rate governed by diffusion through a static air layer and/or permeation through a membrane

[SOURCE: EN 1540:2011, 3.2.1.2, modified — Reference to airborne particles deleted and “a physical process such as” deleted.]

### 2.2.2.1.3

#### **active sampler**

*sampler* (2.2.2.1) that collects gases, *vapours* (2.1.2.5) or *airborne particles* (2.1.2.3), by means of active air movement

Note 1 to entry: Active samplers can collect *samples* (2.2.1.1) onto a *collection substrate* (2.2.3.7) such as a filter or a *sorbent tube* (2.2.2.5) or can collect samples into a canister or bag.

[SOURCE: EN 1540:2011, 3.2.1.3, modified — The phrase “on a collection substrate” was deleted from the definition and Note 1 to entry was added.]

### 2.2.2.1.4

#### **pumped sampler**

*active sampler* (2.2.2.1.3) that collects gases, *vapours* (2.1.2.5) or *airborne particles* (2.1.2.3) where the active air movement is induced by means of a pump

[SOURCE: EN 1540:2011, 3.2.1.4, modified — The phrase “on a collection substrate” was deleted from the definition and Note 1 to entry was deleted.]

### 2.2.2.1.5

#### **flow-controlled pump**

pump with nominally constant flow rate provided by an automatic flow control system

[SOURCE: ISO 13137:2013, 3.10]

### 2.2.2.1.6

#### **aerosol sampler**

#### **airborne particle sampler**

#### **airborne particulate sampler**

device that is used to collect *airborne particles* (2.1.2.3)

Note 1 to entry: The term aerosol sampler is commonly used although it is not in line with the definition of aerosol given in *aerosol* (2.1.4.1).

Note 2 to entry: The collection of airborne particles can be either active or passive.

[SOURCE: EN 1540:2011, 3.2.1.5, modified — Synonym format modified in accordance with ISO 10241-1:2011. “Sampler” changed to “device” and “transport” changed to “collect” in definition. The phrase “to a collection substrate” was deleted from definition. Cross-reference in Note 1 to entry was modified and Note 2 to entry was modified.]

#### 2.2.2.1.6.1

#### **inhalable sampler**

*aerosol sampler* (2.2.2.1.6) that is used to collect the *inhalable fraction* (2.1.3.1.1) of airborne particles from the surrounding air

[SOURCE: EN 1540:2011, 3.2.1.5.1, modified — “Of airborne particles from the surrounding air” was added.]

#### 2.2.2.1.6.2

#### **thoracic sampler**

*aerosol sampler* (2.2.2.1.6) that is used to collect the *thoracic fraction* (2.1.3.1.3) of airborne particles from the surrounding air

[SOURCE: EN 1540:2011, 3.2.1.5.2, modified — “Of airborne particles from the surrounding air” was added.]



#### 2.2.2.1.6.3

##### **respirable sampler**

*aerosol sampler* (2.2.2.1.6) that is used to collect the *respirable fraction* (2.1.3.1.4) of airborne particles from the surrounding air

[SOURCE: EN 1540:2011, 3.2.1.5.3, modified — “Of airborne particles from the surrounding air” was added.]

#### 2.2.2.1.7

##### **mixed-phase sampler**

*sampler* (2.2.2.1) or *sampling train* (2.2.2.6) that is used to collect *airborne particles* (2.1.2.3) and *vapours* (2.1.2.5) onto one or more *collection substrates* (2.2.3.7)

[SOURCE: EN 1540:2011, 3.2.1.6, modified — “Transport” changed to “collect.” “To one or more” changed to “Onto one or more”.]

#### 2.2.2.2

##### **personal sampler**

*sampler* (2.2.2.1), attached to a person, that collects gases, *vapours* (2.1.2.5) or *airborne particles* (2.1.2.3) in the *breathing zone* (2.1.5.5) for the purpose of measuring *exposure* (2.1.5.1) to *chemical agents* (2.1.1.2) and/or *biological agents* (2.1.1.1)

[SOURCE: EN 1540:2011, 3.2.2, modified — “To determine” changed to “for the purpose of measuring” and “agents” added after “chemical” to facilitate cross-referencing.]

#### 2.2.2.3

##### **static sampler**

##### **area sampler**

*stationary sampler* (2.2.2.1), not attached to a person, that collects gases, *vapours* (2.1.2.5) or *airborne particles* (2.1.2.3) at a particular location

[SOURCE: EN 1540:2011, 3.2.3, modified — “Stationary” was added to the definition.]

#### 2.2.2.4

##### **length-of-stain detector tube**

glass tube containing chemical reagents in which a colour change is produced on a graduated scale, based on concentration of a specific *chemical agent* (2.1.1.2), when a *sample* (2.2.1.1) is drawn through it

#### 2.2.2.5

##### **sorbent tube**

sampling device, usually made of metal or glass, containing a *collection substrate* (2.2.3.7) such as a *sorbent* (B.3.3) or a support impregnated with reagent, through which sampled air passes

Note 1 to entry: Some sorbent tubes are intended for use as *active samplers* (2.2.2.1.3) and some as *passive samplers* (2.2.2.1.1).

[SOURCE: EN 1540:2011, 3.2.5, modified — “Sampling” and “through which sampled air passes” were added to the definition.]

#### 2.2.2.6

##### **sampling train**

apparatus consisting of one or more (*air*) *samplers* (2.2.2.1) connected in series, along with associated sampling equipment and connecting tubing, used to collect one or more *chemical agents* (2.1.1.2)

## 2.2.3 Workplace (air) sampling

### 2.2.3.1 sampling air sampling

<workplace air quality> process consisting of the collection of *chemical agents* (2.1.1.2) and/or *biological agents* (2.1.1.1) from air or the withdrawal or isolation of a fractional part of a larger volume of air

[SOURCE: EN 1540:2011, 3.3.1, modified — “<Workplace air quality>” was added as the domain for the definition. “Separation” was changed to “collection.” “Onto a collection substrate” was deleted. “Agents” was added after “chemical” to facilitate cross-referencing.]

### 2.2.3.2 sampling method air sampling method

<workplace air quality> all steps of the *measuring procedure* (2.1.5.6) that describe the physical process of (*air sampling* (2.2.3.1)

[SOURCE: EN 1540:2011, 3.3.2, modified — “<Workplace air quality>” added as the domain for the definition.]

### 2.2.3.3 personal sampling

process of using a *sampler* (2.2.2.1), attached to a person, to collect gases, *vapours* (2.1.2.5) or *airborne particles* (2.1.2.3) in the *breathing zone* (2.1.5.5) for the purpose of measuring *exposure* (2.1.5.1) to *chemical agents* (2.1.1.2) and/or *biological agents* (2.1.1.1)

### 2.2.3.4 static sampling area sampling

process of using a *sampler* (2.2.2.1) in a stationary location that collects gases, *vapours* (2.1.2.5) or *airborne particles* (2.1.2.3) for the purpose of measuring *exposure* (2.1.5.1) to *chemical agents* (2.1.1.2) and/or *biological agents* (2.1.1.1)

### 2.2.3.5 loading

<sample> amount of *sample* (2.2.1.1) collected

[SOURCE: EN 1540:2011, 3.3.5, modified — definition pertaining to <analyte> is in *loading* (2.2.3.6) and “on the collection substrate” changed to “collected”]

### 2.2.3.6 loading

<analyte> amount of *analyte* (2.3.2) collected

[SOURCE: EN 1540:2011, 3.3.5, modified — definition pertaining to (sample) is in *loading* (2.2.3.5) and “on the collection substrate” changed to “collected”]

### 2.2.3.7 collection substrate sampling substrate collection medium sampling medium

medium on which airborne *chemical agents* (2.1.1.2) and/or *biological agents* (2.1.1.1) are collected for subsequent *analysis* (2.3.1)

Note 1 to entry: Filters, polyurethane foams, impinger solutions and sampling cassettes are examples of collection substrates for *airborne particles* (2.1.2.3).

Note 2 to entry: Activated carbon, silica gel and reagent impregnated filters are examples of *collection substrates* (2.2.3.7) for gases and *vapours* (2.1.2.5).

Note 3 to entry: Agar and water media are examples of collection substrates for *bioaerosols* (2.1.4.2).

[SOURCE: EN 1540:2011, 3.3.6, modified — In Note 1 to entry, “impinger solutions” added; in Note 3 to entry, “and water” added; “agents” added after “chemical” to facilitate cross-referencing.]

#### 2.2.3.8

##### **blank (sample)**

unused *collection substrate* (2.2.3.7), taken from the same batch used for *sampling* (2.2.3.1), processed so as to measure artifacts in the measurement (sampling and analysis) process

[SOURCE: EN 14902:2005, 3.1.9]

#### 2.2.3.9

##### **field blank**

*blank (sample)* (2.2.3.8) that is transported to the sampling site, but not used for sample collection

Note 1 to entry: A field blank is loaded in the *sampler* (2.2.2.1), where applicable, and returned to the laboratory in the same way as a *sample* (2.2.1.1).

Note 2 to entry: The results from the *analysis* (2.3.1) of field blanks are used to identify contamination of the sample arising from handling in the field and during transport.

[SOURCE: EN 1540:2011, 3.3.8]

#### 2.2.3.10

##### **laboratory blank**

##### **method blank**

*blank (sample)* (2.2.3.8) that is not transported to the field

Note 1 to entry: The laboratory blank undergoes the same handling as the sample substrate in the laboratory, including conditioning and placing into the *samplers* (2.2.2.1) or transport containers when this is done in the laboratory.

Note 2 to entry: The results from the *analysis* (2.3.1) of laboratory blanks are used to correct sample results for contamination with *analyte* (2.3.2) and/or *interferents* (2.3.6).

#### 2.2.3.11

##### **breakthrough volume**

<gas/vapour> volume of air that can be passed through a *sampler* (2.2.2.1) before the gas or *vapour* (2.1.2.5) exceeds the capacity of the sampler

Note 1 to entry: For practical application, see EN 1076:2009, Annex A.

[SOURCE: EN 1540:2011, 3.3.9, modified — “elutes from” changed to “exceeds the capacity of”.]

#### 2.2.3.12

##### **sampling efficiency**

##### **sampler efficiency**

<aerosol sampler> for each particle *aerodynamic diameter* (2.1.4.8), relative fraction of the concentration of *airborne particles* (2.1.2.3) collected from the undisturbed air onto the *collection substrate* (2.2.3.7) for *analysis* (2.3.1)

Note 1 to entry: The sampling efficiency is independent of whether the particle concentration is determined by number, surface area or mass.

Note 2 to entry: As used in this definition, the word “undisturbed” applies to ideal laboratory conditions where the presence of the *sampler* (2.2.2.1) and the body onto which it is mounted do not disturb the determination of the reference concentration. The word “undisturbed” does not refer to movement of the air itself.

Note 3 to entry: For an *aerosol sampler* (2.2.2.1.6) with internal separation, e.g. size-selective sampling, the sampling efficiency is the product of the inlet efficiency and the internal penetration.

[SOURCE: EN 1540:2011, 3.3.10, modified — Synonym “efficiency curve” deleted; “aerodynamic diameter of a particle” changed to “particle aerodynamic diameter”; “transported” changed to “collected”; “to the collection substrate” changed to “onto the collection substrate”; new Note 2 to entry added; in Note 3 to entry, “a sampler” changed to “an aerosol sampler”, “separation efficiency” changed to “internal penetration”; “e.g. size selective sampling” added.]

#### 2.2.3.13

##### **pressure drop**

<sampling train> difference between ambient pressure and the pressure at the inlet of the pump, for a constant volume flow rate setting

Note 1 to entry: The pressure drop, sometimes referred to as back pressure, is measured across the *sampler* (2.2.2.1), the *collection substrate* (2.2.3.7) and the tubing.

[SOURCE: ISO 13137:2013, 3.9]

#### 2.2.3.14

##### **operating time**

period during which the sampling pump can be operated at specified flow rate and *pressure drop* (2.2.3.13) without recharging or replacing the battery

[SOURCE: ISO 13137:2013, 3.12, modified — “sampling” added.]

### 2.3 Terms related to the analytical method

#### 2.3.1

##### **analysis**

all operations carried out after *sample preparation* (2.3.11) to determine the amount or concentration of the *analyte(s)* (2.3.2) of interest present in the *sample* (2.2.1.1)

Note 1 to entry: Adapted from EN 14902:2005, 3.1.1.

[SOURCE: EN 1540:2011, 4.1]

#### 2.3.2

##### **analyte**

substance or chemical constituent that is determined in an *analytical method* (2.3.3)

[SOURCE: EN 1540:2011, 4.2]

#### 2.3.3

##### **analytical method**

all steps of the *measuring procedure* (2.1.5.6) that describe the overall process of *sample preparation* (2.3.11) and *analysis* (2.3.1)

Note 1 to entry: In the context of this International Standard, determination of mass by weighing is considered to be an analytical method.

[SOURCE: EN 1540:2011, 4.3]

#### 2.3.4

##### **homologous series**

series of compounds possessing similar physicochemical properties, each member of which differs from the preceding member by addition of a repeating unit

Note 1 to entry: A common example of the repeating unit is the –CH<sub>2</sub>– methylene group.

[SOURCE: EN 1540:2011, 4.4, modified — In Note 1, “The most” changed to “A”.]

### 2.3.5

#### **instrumental detection limit**

##### **IDL**

lowest concentration at which an instrument can distinguish the presence of *analyte* (2.3.2) from the background generated by a matrix, such as a *reagent blank* (2.3.9), having a minimal content of that analyte

[SOURCE: ISO 30011:2010, 3.2.7, modified — “analyte content” changed to “the presence of analyte”, “minimal matrix” changed to “matrix, such as a reagent blank, having a minimal content of that analyte” for clarity, and Note 1 to entry not included.]

### 2.3.6

#### **interferent**

constituent of the (*air*) *sample* (2.2.1.1) or other aspect of the sampling or analytical procedure having an adverse effect on the accuracy of the measurement

Note 1 to entry: Interferents can include components of sampling or analysis equipment, reagents, etc.

[SOURCE: EN 1540:2011, 4.5, modified — Added “or other aspect of the sampling or analytical procedure” and added Note 1 to entry.]

### 2.3.7

#### **dynamic range**

range of concentrations over which the measured *analyte* (2.3.2) response is suitable for the *measuring procedure* (2.1.5.6)

Note 1 to entry: For example, in methods such as atomic spectrometry, the linear dynamic range extends from the *limit of quantification* (2.4.3.5) to the onset of calibration curvature. See ISO 30011 for more information.

### 2.3.8

#### **measurand**

particular quantity subject to measurement

[SOURCE: EN 1540:2011, 4.6]

### 2.3.9

#### **reagent blank**

all reagents used in *sample preparation* (2.3.11), in the same quantities used to prepare blank and sample solutions

Note 1 to entry: The reagent blank is used to assess contamination from the laboratory environment and to characterize background from the reagents used in sample preparation.

[SOURCE: ISO 30011:2010, 3.2.12, modified — Editorially improved.]

### 2.3.10

#### **reference sample**

*sample* (2.2.1.1) having a known or measured content and/or *loading* (2.2.3.6) of the *analyte* (2.3.2) of interest

Note 1 to entry: A reference sample can be analysed to determine the *analytical bias* (2.4.3.1.1) or the *analytical precision* (2.4.2.6) of a *measuring procedure* (2.1.5.6).

[SOURCE: EN 1540:2011, 4.7, modified — “or loading” changed to “and/or loading”.]

### 2.3.11

#### **sample preparation**

all operations carried out on a *sample* (2.2.1.1), usually after transportation and storage, to prepare it for *analysis* (2.3.1), including transformation of the sample into a measurable state, where necessary

Note 1 to entry: Adapted from EN 14902:2005, 3.1.24.

[SOURCE: EN 1540:2011, 4.8]

## 2.4 Terms related to method performance

### 2.4.1 Efficiency

#### 2.4.1.1

##### **analytical recovery**

all operations carried out after *sample preparation* (2.3.11) to determine the amount or concentration of the *analyte(s)* (2.3.2) of interest present in the collected *sample* (2.2.1.2)

Note 1 to entry: Adapted from EN 14902:2005, 3.1.1.

[SOURCE: EN 1540:2011, 5.1.1]

#### 2.4.1.2

##### **method recovery**

ratio of the measured concentration of *chemical agent* (2.1.1.2) in air to its actual concentration

Note 1 to entry: The method recovery is usually given as a percentage.

Note 2 to entry: The method recovery incorporates both *sampling efficiency* (2.2.3.12) and *analytical recovery* (2.4.1.1).

[SOURCE: EN 1540:2011, 5.1.2, modified — Changed “determined” to “measured”.]

#### 2.4.1.3

##### **selectivity**

<measuring procedure> extent of independence of a *measuring procedure* (2.1.5.6) from interferences

[SOURCE: EN 1540:2011, 5.1.3]

### 2.4.2 Uncertainty

#### 2.4.2.1

##### **uncertainty of measurement**

##### **measurement error**

*measured quantity value* (2.4.2.1.1) minus a *reference quantity value* (2.4.2.1.2)

Note 1 to entry: Application of the concept of measurement uncertainty is described in JCGM 200:2012.

[SOURCE: JCGM 200:2012, 2.16, modified — Only the JCGM definition is used; Note 1 to entry is added.]

##### 2.4.2.1.1

##### **measured quantity value**

##### **value of a measured quantity**

##### **measured value**

quantity value representing a measurement result

Note 1 to entry: Additional information can be found in JCGM 200:2012.

[SOURCE: JCGM 200:2012, 2.10, modified – Only the JCGM definition is used; Note 1 to entry is added.]

##### 2.4.2.1.2

##### **reference quantity value**

##### **reference value**

quantity value used as a basis for comparison with values of quantities of the same kind

Note 1 to entry: Additional information can be found in JCGM 200:2012.

[SOURCE: JCGM 200:2012, 5.18, modified — Only the JCGM definition is used; Note 1 to entry is added.]

**2.4.2.1.3**  
**non-random uncertainty**  
**systematic measurement error**  
**systematic error of measurement**  
**systematic error**

component of *uncertainty of measurement* (2.4.2.1) that, in replicate measurements, remains constant or varies in a predictable manner

Note 1 to entry: Additional information can be found in JCGM 200:2012.

[SOURCE: JCGM 200:2012, 2.17, modified — Only the JCGM definition is used; Note 1 to entry is added.]

**2.4.2.1.4**  
**random uncertainty**  
**random measurement error**  
**random error of measurement**  
**random error**

uncertainty associated with random errors

Note 1 to entry: Additional information can be found in JCGM 200:2012.

[SOURCE: EN 1540:2011, 5.2.1.2, modified — Note 1 to entry added.]

**2.4.2.2**  
**analytical uncertainty**  
**analytical error**

$u_a$

combined uncertainty of the *analytical method* (2.3.3) including contributions from the *analytical recovery* (2.4.1.1) and the analytical variability

[SOURCE: EN 1540:2011, 5.2.2, modified — Synonym “analytical error” has been added and Note 1 to entry has been moved into the definition.]

**2.4.2.2.1**  
**non-random analytical uncertainty**  
**systematic analytical error**

$u_{a,nr}$

uncertainty associated with non-random analytical error

[SOURCE: EN 1540:2011, 5.2.2.1]

**2.4.2.2.2**  
**random analytical uncertainty**  
**random analytical error**

$u_{a,r}$

uncertainty associated with random analytical error

Note 1 to entry: The random analytical error (in some cases called “analytical variability”) is equivalent to the *analytical precision* (2.4.2.6) determined under reproducibility conditions or the analytical precision determined under repeatability conditions together with other random uncertainty components associated with interferences, calibration, instrument response drift, blank correction, etc.

[SOURCE: EN 1540:2011, 5.2.2.2]

### 2.4.2.3 sampling uncertainty sampling error

$u_s$   
combined uncertainty of the *sampling method* (2.2.3.2) including, where relevant, contributions from sampled air volume, *sampling efficiency* (2.2.3.12), sample storage and, if applicable, transport

[SOURCE: EN 1540:2011, 5.2.3, modified — Synonym “sampling error” has been added, “sampling procedure” changed to “sampling method”, and note moved into definition.]

#### 2.4.2.3.1 non-random sampling uncertainty systematic sampling uncertainty systematic sampling error

$u_{s,nr}$   
uncertainty associated with non-random sampling error

[SOURCE: EN 1540:2011, 5.2.3.1]

#### 2.4.2.3.2 random sampling uncertainty random sampling error

$u_{s,r}$   
uncertainty associated with random sampling error

[SOURCE: EN 1540:2011, 5.2.3.2]

### 2.4.2.4 combined standard uncertainty

$u_c$   
uncertainty of the result of a measurement, expressed as a standard deviation, when that result is obtained from the values of a number of other quantities, equal to the positive square root of a sum of terms, the terms being the variances or covariances of these other quantities weighted according to how the measurement result varies with changes in these quantities

[SOURCE: EN 1540:2011, 5.2.5, modified — Definition of “standard uncertainty” from EN 1540 incorporated into this definition.]

#### 2.4.2.4.1 combined non-random standard uncertainty

$u_{c,nr}$   
combined uncertainty associated with *non-random sampling uncertainty* (2.4.2.3.1) and *non-random analytical uncertainty* (2.4.2.2.1)

[SOURCE: EN 1540:2011, 5.2.5.2, modified — “error” changed to “uncertainty”.]

#### 2.4.2.4.2 combined random standard uncertainty

$u_{c,r}$   
combined uncertainty associated with *random sampling uncertainty* (2.4.2.3.2) and *random analytical uncertainty* (2.4.2.2.2)

Note 1 to entry: In some cases, the combined random sampling uncertainty and random analytical uncertainty is called method variability.

[SOURCE: EN 1540:2011, 5.2.5.1, modified — “error” changed to “uncertainty”.]



#### 2.4.2.5 expanded uncertainty

*U*

quantity defining an interval about a result of a measurement, expected to encompass a large fraction of the distribution of values that could reasonably be attributed to the *measurand* (2.3.8)

[SOURCE: EN 1540:2011, 5.2.6]

#### 2.4.2.6 analytical precision

closeness of agreement between indications or *measured quantity values* (2.4.2.1.1) obtained by replicate analytical measurements on the same or similar objects under specified conditions

Note 1 to entry: Analytical precision is usually expressed numerically, e.g. by standard deviation, variance, or coefficient of variation under specified conditions of measurement.

Note 2 to entry: The analytical precision can be determined under *repeatability conditions of measurement* (2.4.3.7), *intermediate precision conditions of measurement* (B.5.5), or *reproducibility conditions of measurement* (2.4.3.9). See ISO 5725-1.

[SOURCE: JCGM 200:2012 (VIM), 2.15]

#### 2.4.2.7 method precision

variability arising from the *measuring procedure* (2.1.5.6) as a whole, including sampling and *analysis* (2.3.1)

[SOURCE: EN 1540:2011, 5.2.8, modified — “imprecision” changed to “variability”.]

### 2.4.3 General statistical terms

#### 2.4.3.1 measurement bias

**bias**

estimate of a *systematic measurement error* (2.4.2.1.3)

Note 1 to entry: An additional definition can be found in ISO 3534-2:2006, 3.3.2.

[SOURCE: JCGM 200:2012 (VIM), 2.18, modified — Note 1 to entry added.]

##### 2.4.3.1.1 analytical bias

estimate of systematic analytical error

##### 2.4.3.1.2 method bias

*systematic measurement error* (2.4.2.1.3) of the *measuring procedure* (2.1.5.6) as a whole, incorporating both *sampling efficiency* (2.2.3.12) and *analytical bias* (2.4.3.1.1)

[SOURCE: EN 1540:2011, 5.3.1.2, modified — In definition text, “bias” changed to “systematic measurement error” and Note 1 moved into definition.]

#### 2.4.3.2 coverage factor

*k*

numerical factor used as a multiplier of the *combined standard uncertainty* (2.4.2.4) in order to obtain an *expanded uncertainty* (2.4.2.5)

Note 1 to entry: A coverage factor, *k*, is typically in the range from 2 to 3.

[SOURCE: EN 1540:2011, 5.3.2, modified – Note 2 to entry deleted.]

#### 2.4.3.3

**precision**  
**measurement precision**

closeness of agreement between independent test/measurement results obtained under stipulated conditions

Note 1 to entry: Precision can be expressed as a standard deviation or relative standard deviation (coefficient of variation).

[SOURCE: ISO 3534-2:2006, 3.3.4, modified — Note 1 to entry added.]

#### 2.4.3.4

**limit of detection**  
**detection limit**

**LOD**

lowest amount of an *analyte* (2.3.2) that is detectable with a given confidence level

Note 1 to entry: The limit of detection can be calculated as three times the standard deviation of blank measurements. This represents a probability of 50 % that the analyte will not be detected when it is present at the concentration of the LOD.

Note 2 to entry: The LOD can be used as a threshold value to assert the presence of a substance with a known confidence.

[SOURCE: EN 1540:2011, 5.3.4]

#### 2.4.3.5

**limit of quantification**  
**quantification limit**

**LOQ**

lowest amount of an *analyte* (2.3.2) that is quantifiable with a given confidence level

Note 1 to entry: The limit of quantification can be calculated as ten times the standard deviation of blank measurements.

Note 2 to entry: The value LOQ can be used as a threshold value to assure quantitative measurement of an analyte accurately.

[SOURCE: EN 1540:2011, 5.3.5, modified — Note 2 to entry modified and Note 3 to entry deleted.]

#### 2.4.3.6

**measurement repeatability**  
**repeatability**

*measurement precision* (2.4.3.3) under a set of repeatability conditions of measurement

[SOURCE: JCGM 200:2012 (VIM), 2.21]

#### 2.4.3.7

**repeatability condition of measurement**  
**repeatability condition**

condition of measurement, out of a set of conditions that includes the same *measuring procedure* (2.1.5.6), same operators, same measuring system, same operating conditions and same location, and replicate measurements on the same or similar objects over a short period of time

Note 1 to entry: A condition of measurement is a repeatability condition only with respect to a specified set of repeatability conditions.

[SOURCE: JCGM 200:2012 (VIM), 2.20, modified — “measurement procedure” changed to “measuring procedure”.]

**2.4.3.8**  
**measurement reproducibility**  
**reproducibility**

*measurement precision* (2.4.3.3) under *reproducibility conditions of measurement* (2.4.3.9)

[SOURCE: JCGM 200:2012 (VIM), 2.25]

**2.4.3.9**  
**reproducibility condition of measurement**  
**reproducibility condition**

condition of measurement, out of a set of conditions that includes different locations, operators, measuring systems, and replicate measurements on the same or similar objects

Note 1 to entry: The different measuring systems may use different *measuring procedures* (2.1.5.6).

Note 2 to entry: A specification should give the conditions changed and unchanged, to the extent practical.

[SOURCE: JCGM 200:2012 (VIM), 2.24, modified — In Note 1 to entry, “measurement” changed to “measuring”.]

**2.4.3.10**  
**true value**

value which characterizes a quantity or quantitative characteristic perfectly defined in the conditions which exist when that quantity or quantitative characteristic is considered

Note 1 to entry: The true value of a quantity or quantitative characteristic is a theoretical concept and, in general, cannot be known exactly. In practice, a reference value is commonly accepted as the true value.

[SOURCE: ISO 3534-2:2006, 3.2.5, modified — Note 2 to entry deleted.]

**2.4.4 Other statistical terms**

**2.4.4.1**  
**validation**

process of evaluating the performance of a *measuring procedure* (2.1.5.6) and checking that the performance meets certain pre-set criteria

Note 1 to entry: Performance characteristics to be considered include confirmation of identity, *selectivity* (2.4.1.3)/specificity, *limit of detection* (2.4.3.4), *limit of quantification* (2.4.3.5), *analytical recovery* (2.4.1.1), working and linear *dynamic ranges* (2.3.7), accuracy, *measurement repeatability* (2.4.3.6), *measurement reproducibility* (2.4.3.8), *ruggedness* (B.4.5), and *robustness* (B.4.4).

[SOURCE: EN 1540:2011, 5.4.1]

**2.4.4.2**  
**measuring range**

concentration range for which the *expanded uncertainty* (2.4.2.5) of a validated *measuring procedure* (2.1.5.6) is a specified value

Note 1 to entry: EN 482:2012 specifies maximum values for expanded uncertainty.

[SOURCE: EN 1540:2011, 5.4.2]

## Annex A (informative)

### Alphabetical index of terms defined

[Table A.1](#) provides an alphabetical index of terms defined in this International Standard. Preferred terms (presented in bold in the main body) are listed in non-italicized font. Synonyms (admitted terms) are listed in italics.

**Table A.1 — Index of terms defined**

<b>Term</b>	<b>Number</b>
active sampler	<a href="#">2.2.2.1.3</a>
<i>aerodynamic diameter</i>	<a href="#">2.1.4.8</a>
aerosol	<a href="#">2.1.4.1</a>
aerosol sampler	<a href="#">2.2.2.1.6</a>
agglomerate	<a href="#">2.1.4.9</a>
aggregate	<a href="#">2.1.4.10</a>
air pollutant	<a href="#">2.1.2.1</a>
<i>air sample</i>	<a href="#">2.2.1.1</a>
<i>air sampler</i>	<a href="#">2.2.2.1</a>
<i>air sampling</i>	<a href="#">2.2.3.1</a>
<i>air sampling method</i>	<a href="#">2.2.3.2</a>
airborne dust	<a href="#">2.1.2.2</a>
<i>airborne particle sampler</i>	<a href="#">2.2.2.1.6</a>
airborne particles	<a href="#">2.1.2.3</a>
<i>airborne particulate sampler</i>	<a href="#">2.2.2.1.6</a>
analysis	<a href="#">2.3.1</a>
analyte	<a href="#">2.3.2</a>
analytical bias	<a href="#">2.4.3.1.1</a>
<i>analytical error</i>	<a href="#">2.4.2.2</a>
analytical method	<a href="#">2.3.3</a>
analytical precision	<a href="#">2.4.2.6</a>
analytical recovery	<a href="#">2.4.1.1</a>
analytical uncertainty	<a href="#">2.4.2.2</a>
<i>area sample</i>	<a href="#">2.2.1.4</a>
<i>area sampler</i>	<a href="#">2.2.2.3</a>
<i>area sampling</i>	<a href="#">2.2.3.4</a>
<i>bias</i>	<a href="#">2.4.3.1</a>
bioaerosol	<a href="#">2.1.4.2</a>
biological agent	<a href="#">2.1.1.1</a>
blank (sample)	<a href="#">2.2.3.8</a>
breakthrough volume	<a href="#">2.2.3.11</a>
breathing zone	<a href="#">2.1.5.5</a>
chemical agent	<a href="#">2.1.1.2</a>

Table A.1 (continued)

Term	Number
coagulation	<a href="#">2.1.4.11</a>
coalescence	<a href="#">2.1.4.12</a>
collected sample	<a href="#">2.2.1.2</a>
<i>collection medium</i>	<a href="#">2.2.3.7</a>
collection substrate	<a href="#">2.2.3.7</a>
combined non-random standard uncertainty	<a href="#">2.4.2.4.1</a>
combined random standard uncertainty	<a href="#">2.4.2.4.2</a>
combined standard uncertainty	<a href="#">2.4.2.4</a>
coverage factor	<a href="#">2.4.3.2</a>
dermal exposure	<a href="#">2.1.5.2</a>
<i>detection limit</i>	<a href="#">2.4.3.4</a>
diffusive sampler	<a href="#">2.2.2.1.2</a>
dustiness	<a href="#">2.1.6.1</a>
dynamic range	<a href="#">2.3.7</a>
expanded uncertainty	<a href="#">2.4.2.5</a>
exposure	<a href="#">2.1.5.1</a>
extrathoracic convention	<a href="#">2.1.3.2.2</a>
extrathoracic fraction	<a href="#">2.1.3.1.2</a>
field blank	<a href="#">2.2.3.9</a>
flow-controlled pump	<a href="#">2.2.2.1.5</a>
health-related fractions	<a href="#">2.1.3.1</a>
homologous series	<a href="#">2.3.4</a>
<i>IDL</i>	<a href="#">2.3.5</a>
inhalable convention	<a href="#">2.1.3.2.1</a>
inhalable fraction	<a href="#">2.1.3.1.1</a>
inhalable sampler	<a href="#">2.2.2.1.6.1</a>
instrumental detection limit	<a href="#">2.3.5</a>
interferent	<a href="#">2.3.6</a>
laboratory blank	<a href="#">2.2.3.10</a>
length-of-stain detector tube	<a href="#">2.2.2.4</a>
limit of detection	<a href="#">2.4.3.4</a>
limit of quantification	<a href="#">2.4.3.5</a>
loading (analyte)	<a href="#">2.2.3.6</a>
loading (sample)	<a href="#">2.2.3.5</a>
<i>LOD</i>	<a href="#">2.4.3.4</a>
<i>LOQ</i>	<a href="#">2.4.3.5</a>
measurand	<a href="#">2.3.8</a>
measured quantity value	<a href="#">2.4.2.1.1</a>
<i>measured value</i>	<a href="#">2.4.2.1.1</a>
measurement bias	<a href="#">2.4.3.1</a>
<i>measurement error</i>	<a href="#">2.4.2.1</a>
<i>measurement method</i>	<a href="#">2.1.5.6</a>
<i>measurement precision</i>	<a href="#">2.4.3.3</a>
<i>measurement procedure</i>	<a href="#">2.1.5.6</a>

**Table A.1** (continued)

<b>Term</b>	<b>Number</b>
measurement repeatability	<a href="#">2.4.3.6</a>
measurement reproducibility	<a href="#">2.4.3.8</a>
measuring procedure	<a href="#">2.1.5.6</a>
measuring range	<a href="#">2.4.4.2</a>
method bias	<a href="#">2.4.3.1.2</a>
method precision	<a href="#">2.4.2.7</a>
method recovery	<a href="#">2.4.1.2</a>
mixed-phase sampler	<a href="#">2.2.2.1.7</a>
nanoaerosol	<a href="#">2.1.4.3</a>
nanoparticle	<a href="#">2.1.4.4</a>
nanostructured particle	<a href="#">2.1.4.5</a>
non-random analytical uncertainty	<a href="#">2.4.2.2.1</a>
non-random sampling uncertainty	<a href="#">2.4.2.3.1</a>
non-random uncertainty	<a href="#">2.4.2.1.3</a>
occupational exposure limit value	<a href="#">2.1.5.4</a>
<i>OELV</i>	<a href="#">2.1.5.4</a>
operating time	<a href="#">2.2.3.14</a>
particle aerodynamic diameter	<a href="#">2.1.4.8</a>
passive sampler	<a href="#">2.2.2.1.1</a>
personal sample	<a href="#">2.2.1.3</a>
personal sampler	<a href="#">2.2.2.2</a>
personal sampling	<a href="#">2.2.3.3</a>
precision	<a href="#">2.4.3.3</a>
pressure drop	<a href="#">2.2.3.13</a>
pumped sampler	<a href="#">2.2.2.1.4</a>
<i>quantification limit</i>	<a href="#">2.4.3.5</a>
<i>random analytical error</i>	<a href="#">2.4.2.2.2</a>
random analytical uncertainty	<a href="#">2.4.2.2.2</a>
<i>random error</i>	<a href="#">2.4.2.1.4</a>
<i>random error of measurement</i>	<a href="#">2.4.2.1.4</a>
<i>random measurement error</i>	<a href="#">2.4.2.1.4</a>
<i>random sampling error</i>	<a href="#">2.4.2.3.2</a>
random sampling uncertainty	<a href="#">2.4.2.3.2</a>
random uncertainty	<a href="#">2.4.2.1.4</a>
reagent blank	<a href="#">2.3.9</a>
reference period	<a href="#">2.1.5.7</a>
reference quantity value	<a href="#">2.4.2.1.2</a>
reference sample	<a href="#">2.3.10</a>
<i>reference value</i>	<a href="#">2.4.2.1.2</a>
<i>repeatability</i>	<a href="#">2.4.3.6</a>
<i>repeatability condition</i>	<a href="#">2.4.3.7</a>
repeatability condition of measurement	<a href="#">2.4.3.7</a>
<i>reproducibility</i>	<a href="#">2.4.3.8</a>
<i>reproducibility condition</i>	<a href="#">2.4.3.9</a>

Table A.1 (continued)

Term	Number
reproducibility condition of measurement	<a href="#">2.4.3.9</a>
respirable convention	<a href="#">2.1.3.2.4</a>
respirable fraction	<a href="#">2.1.3.1.4</a>
respirable sampler	<a href="#">2.2.2.1.6.3</a>
sample	<a href="#">2.2.1.1</a>
sample preparation	<a href="#">2.3.11</a>
sampler	<a href="#">2.2.2.1</a>
<i>sampler efficiency</i>	<a href="#">2.2.3.12</a>
sampling	<a href="#">2.2.3.1</a>
sampling convention	<a href="#">2.1.3.2</a>
sampling efficiency	<a href="#">2.2.3.12</a>
<i>sampling error</i>	<a href="#">2.4.2.3</a>
<i>sampling medium</i>	<a href="#">2.2.3.7</a>
sampling method	<a href="#">2.2.3.2</a>
<i>sampling substrate</i>	<a href="#">2.2.3.7</a>
sampling train	<a href="#">2.2.2.6</a>
sampling uncertainty	<a href="#">2.4.2.3</a>
selectivity	<a href="#">2.4.1.3</a>
sorbent tube	<a href="#">2.2.2.5</a>
static sample	<a href="#">2.2.1.4</a>
static sampler	<a href="#">2.2.2.3</a>
static sampling	<a href="#">2.2.3.4</a>
<i>systematic analytical error</i>	<a href="#">2.4.2.2.1</a>
<i>systematic error</i>	<a href="#">2.4.2.1.3</a>
<i>systematic error of measurement</i>	<a href="#">2.4.2.1.3</a>
<i>systematic measurement error</i>	<a href="#">2.4.2.1.3</a>
<i>systematic sampling error</i>	<a href="#">2.4.2.3.1</a>
<i>systematic sampling uncertainty</i>	<a href="#">2.4.2.3.1</a>
thoracic convention	<a href="#">2.1.3.2.3</a>
thoracic fraction	<a href="#">2.1.3.1.3</a>
thoracic sampler	<a href="#">2.2.2.1.6.2</a>
time-weighted average concentration	<a href="#">2.1.5.3</a>
total airborne particles	<a href="#">2.1.2.4</a>
true value	<a href="#">2.4.3.10</a>
<i>TWA concentration</i>	<a href="#">2.1.5.3</a>
ultrafine aerosol	<a href="#">2.1.4.6</a>
ultrafine particle	<a href="#">2.1.4.7</a>
uncertainty of measurement	<a href="#">2.4.2.1</a>
validation	<a href="#">2.4.4.1</a>
<i>value of a measured quantity</i>	<a href="#">2.4.2.1.1</a>
vapour	<a href="#">2.1.2.5</a>
workplace	<a href="#">2.1.6.2</a>

## Annex B (informative)

### Additional terms

#### B.1 General

The normative text of this International Standard contains terms and definitions that are used in workplace air quality International Standards developed by ISO/TC 146/SC 2. This Annex provides additional terms of importance in workplace air quality but not used in International Standards developed by ISO/TC 146/SC 2. This informative Annex is provided for the convenience of users of this International Standard.

These terms are not included in the alphabetical index in Annex A.

#### B.2 Terms used in general

##### B.2.1

##### **contaminant**

material added by human or natural activities which may, in sufficient concentrations, render the atmosphere unacceptable

Note 1 to entry: Contaminants refer to gases, vapours, mists, *aerosols* (2.1.4.1), *fumes* (B.2.2), particles, or dusts, and so forth, that are airborne. The term does not apply to elements that make up the components of the earth's atmosphere, such as nitrogen, oxygen, argon, and so forth.

[SOURCE: ASTM D1356-14b]

##### B.2.2

##### **fume**

*aerosol* (2.1.4.1) of solid particles, usually from metallurgical processes, generated by condensation from the gaseous state, generally after volatilization from melted substances and often accompanied by chemical reactions such as oxidation

[SOURCE: ISO 4225:1994, 3.38]

#### B.3 Terms related to the physical and chemical processes of workplace (air) sampling

##### B.3.1

##### **sampling rate**

rate at which data collection from a *sampler* (2.2.2.1) occurs, usually presented as sample volume per unit time

[SOURCE: ASTM D1356-14b.]

##### B.3.2

##### **representative sample**

*sample* (2.2.1.1) with the same quality and characteristics for the material of interest as that of its source at the time of *sampling* (2.2.3.1)

[SOURCE: ISO 2889:2010, 3.67]



### B.3.3

#### **sorbent**

insoluble material or mixture of materials used to recover fluids through the mechanisms of absorption or adsorption or both

[SOURCE: ISO 16165:2013, 2.6.1.13, modified]

## B.4 Terms related to the analytical method

### B.4.1

#### **calibration curve**

graphical representation of measuring signal as a function of quantity of *analyte* (2.3.2)

[SOURCE: EURACHEM Guide on the Fitness for Purpose of Analytical Methods, A3]

### B.4.2

#### **reference material**

material or substance one or more of whose property values are sufficiently homogenous and well established to be used for the calibration of an apparatus, the assessment of a measurement method, or for assigning values to materials.

Note 1 to entry: A reference material may be in the form of a pure or mixed gas, liquid, or solid. Examples are water for the calibration of viscometers, sapphire as a heat-capacity calibrant in calorimetry, and solutions used for calibration in chemical *analysis* (2.3.1).

[SOURCE: ISO Guide 30:2015, 2.1.1, modified]

### B.4.3

#### **certified reference material**

*reference material* (B.4.2), accompanied by a certificate, one or more of whose property values are certified by a procedure which establishes its traceability to an accurate realization of the unit in which the property values are expressed, and for which each certified value is accompanied by an uncertainty at a stated level of confidence.

[SOURCE: ISO Guide 30:2015, 2.1.2, modified]

### B.4.4

#### **robustness**

capacity of an *analytical method* (2.3.3) to remain unaffected by small but deliberate variations in method parameters

Note 1 to entry: The robustness of an analytical method provides an indication of its reliability during normal usage.

[SOURCE: EURACHEM Guide on the Fitness for Purpose of Analytical Methods, A25.2, modified — Part of the EURACHEM Guide definition was moved to Note 1 to entry.]

### B.4.5

#### **ruggedness**

insensitivity of an *analytical method* (2.3.3) to departures from specified test or environmental conditions

Note 1 to entry: An evaluation of the ruggedness of a test method or an empirical model derived from an experiment is useful in determining whether the results or decisions will be relatively invariant over some range of environmental variability under which the test method or the model is likely to be applied.

[SOURCE: ASTM E1169-13a, modified — “a test method” changed to “an analytical method.”]

## B.5 Terms related to method performance and statistics

### B.5.1

#### **measurement accuracy** **accuracy of measurement** **accuracy**

closeness of agreement between a *measured quantity value* (2.4.2.1.1) and a true quantity value of a *measurand* (2.3.8)

Note 1 to entry: The concept “measurement accuracy” is not a quantity and is not given a numerical quantity value. A measurement is said to be more accurate when it offers a smaller measurement error.

Note 2 to entry: The term “measurement accuracy” should not be used for measurement trueness and the term “measurement precision” should not be used for “measurement accuracy”, which, however, is related to both these concepts.

Note 3 to entry: “Measurement accuracy” is sometimes understood as closeness of agreement between measured quantity values that are being attributed to the measurand.

[SOURCE: JCGM 200:2012 (VIM), 2.13]

### B.5.2

#### **sensitivity of a measuring system** **sensitivity**

quotient of the change in an indication of a measuring system and the corresponding change in a value of a quantity being measured

[SOURCE: JCGM 200:2012 (VIM), 4.12, modified — Notes 1 and 2 to entry deleted.]

### B.5.3

#### **type A evaluation of measurement uncertainty** **type A evaluation**

evaluation of a component of measurement uncertainty by a statistical analysis of *measured quantity values* (2.4.2.1.1) obtained under defined measurement conditions

Note 1 to entry: For various types of measurement conditions, see *repeatability condition of measurement* (2.4.3.7), *intermediate precision condition of measurement* (B.5.5), and *reproducibility condition of measurement* (2.4.3.9).

[SOURCE: JCGM 200:2012 (VIM), 2.28, modified — Notes 2 and 3 to entry were omitted.]

### B.5.4

#### **type B evaluation of measurement uncertainty** **type B evaluation**

evaluation of a component of measurement uncertainty determined by means other than a *type A evaluation of measurement uncertainty* (B.5.3)

EXAMPLE Evaluation based on information

- associated with authoritative published quantity values,
- associated with the quantity value of a *certified reference material* (B.4.3),
- obtained from a calibration certificate,
- about drift,
- obtained from the accuracy class of a verified measuring instrument,
- obtained from limits deduced through personal experience.

[SOURCE: JCGM 200:2012 (VIM), 2.29, modified — Note 1 to entry was omitted.]

**B.5.5**  
**intermediate precision condition of measurement**  
**intermediate precision condition**

condition of measurement, out of a set of conditions that includes the same *measuring procedure* (2.1.5.6), same location, and replicate measurements on the same or similar objects over an extended period of time but may include other conditions involving changes

Note 1 to entry: The changes can include new calibrations, calibrators, operators, and measuring systems.

[SOURCE: JCGM 200:2012 (VIM), 2.22, modified — “Measurement procedure” changed to “measuring procedure.” Notes 2 and 3 to entry were omitted.]

**B.5.6**  
**proficiency testing**

evaluation of participant performance against pre-established criteria by means of interlaboratory comparisons

[SOURCE: ISO/IEC 17043:2010, 3.7, modified]

## Bibliography

- [1] ISO Guide 30:2015, *Reference materials — Selected terms and definitions*
- [2] ISO 2889:2010, *Sampling airborne radioactive materials from the stacks and ducts of nuclear facilities*
- [3] ISO 3534-2:2006, *Statistics — Vocabulary and symbols — Part 2: Applied statistics*
- [4] ISO 4225:1994, *Air quality — General aspects — Vocabulary*
- [5] ISO 5725-1, *Accuracy (trueness and precision) of measurement methods and results — Part 1: General principles and definitions*
- [6] ISO 7708:1995, *Air quality — Particle size fraction definitions for health-related sampling*
- [7] ISO 10241-1:2011, *Terminological entries in standards — Part 1: General requirements and examples of presentation*
- [8] ISO 13137:2013, *Workplace atmospheres — Pumps for personal sampling of chemical and biological agents — Requirements and test methods*
- [9] ISO/TR 14294, *Workplace atmospheres — Measurement of dermal exposure — Principles and methods*
- [10] ISO 16165:2013, *Ships and marine technology — Marine environment protection — Terminology relating to oil spill response*
- [11] ISO/IEC 17043:2010, *Conformity assessment — General requirements for proficiency testing*
- [12] ISO 21438-1:2007, *Workplace atmospheres — Determination of inorganic acids by ion chromatography — Part 1: Non-volatile acids (sulfuric acid and phosphoric acid)*
- [13] ISO/TR 27628:2007, *Workplace atmospheres — Ultrafine, nanoparticle and nano-structured aerosols — Inhalation exposure characterization and assessment*
- [14] ISO 30011:2010, *Workplace air — Determination of metals and metalloids in airborne particulate matter by inductively coupled plasma mass spectrometry*
- [15] ISO/TS 80004-4:2011, *Nanotechnologies — Vocabulary — Part 4: Nanostructured materials*
- [16] JCGM 200:2012, *International vocabulary of metrology — Basic and general concepts and associated terms (VIM)*
- [17] ASTM D1356-14b, *Standard Terminology Relating to Sampling and Analysis of Atmospheres*
- [18] ASTM E1169-13a, *Standard Practice for Conducting Ruggedness Tests*
- [19] EN 481:1993, *Workplace Atmospheres — Size Fraction Definitions for Measurement of Airborne Particles*
- [20] EN 482:2012, *Workplace exposure — General requirements for the performance of procedures for the measurement of chemical agents*
- [21] EN 1076:2009, *Workplace exposure — Procedures for measuring gases and vapours using pumped samplers — Requirements and test methods*
- [22] EN 1540:2011, *Workplace exposure — Terminology*
- [23] EN 14902:2005, *Ambient air quality — Standard method for the measurement of Pb, Cd, As and Ni in the PM<sub>10</sub> fraction of suspended particulate matter*

- [24] Council Directive 98/24/EC of 7 April 1988 on the protection of the health and safety of workers from the risks related to chemical agents at work (fourteenth individual Directive within the meaning of Article 16(1) of Directive 89/391/EEC)
- [25] EURACHEM GUIDE. The Fitness for Purpose of Analytical Methods: A Laboratory Guide to Method Validation and Related Topics. First Internet Version, 1998
- [26] EURACHEM/CITAC GUIDE. *Measurement Uncertainty arising from sampling – A guide to methods and approaches*, First Edition, 2007
- [27] BRITISH COLUMBIA FIELD SAMPLING MANUAL. 2003 Edition, accessible at [www.env.gov.bc.ca/wsd/data\\_searches/field\\_sampling\\_manual/field\\_man\\_pdfs/fld\\_man\\_03.pdf](http://www.env.gov.bc.ca/wsd/data_searches/field_sampling_manual/field_man_pdfs/fld_man_03.pdf)





# British Standards Institution (BSI)

BSI is the national body responsible for preparing British Standards and other standards-related publications, information and services.

BSI is incorporated by Royal Charter. British Standards and other standardization products are published by BSI Standards Limited.

## About us

We bring together business, industry, government, consumers, innovators and others to shape their combined experience and expertise into standards-based solutions.

The knowledge embodied in our standards has been carefully assembled in a dependable format and refined through our open consultation process. Organizations of all sizes and across all sectors choose standards to help them achieve their goals.

## Information on standards

We can provide you with the knowledge that your organization needs to succeed. Find out more about British Standards by visiting our website at [bsigroup.com/standards](http://bsigroup.com/standards) or contacting our Customer Services team or Knowledge Centre.

## Buying standards

You can buy and download PDF versions of BSI publications, including British and adopted European and international standards, through our website at [bsigroup.com/shop](http://bsigroup.com/shop), where hard copies can also be purchased.

If you need international and foreign standards from other Standards Development Organizations, hard copies can be ordered from our Customer Services team.

## Subscriptions

Our range of subscription services are designed to make using standards easier for you. For further information on our subscription products go to [bsigroup.com/subscriptions](http://bsigroup.com/subscriptions).

With **British Standards Online (BSOL)** you'll have instant access to over 55,000 British and adopted European and international standards from your desktop. It's available 24/7 and is refreshed daily so you'll always be up to date.

You can keep in touch with standards developments and receive substantial discounts on the purchase price of standards, both in single copy and subscription format, by becoming a **BSI Subscribing Member**.

**PLUS** is an updating service exclusive to BSI Subscribing Members. You will automatically receive the latest hard copy of your standards when they're revised or replaced.

To find out more about becoming a BSI Subscribing Member and the benefits of membership, please visit [bsigroup.com/shop](http://bsigroup.com/shop).

With a **Multi-User Network Licence (MUNL)** you are able to host standards publications on your intranet. Licences can cover as few or as many users as you wish. With updates supplied as soon as they're available, you can be sure your documentation is current. For further information, email [bsmusales@bsigroup.com](mailto:bsmusales@bsigroup.com).

## Revisions

Our British Standards and other publications are updated by amendment or revision.

We continually improve the quality of our products and services to benefit your business. If you find an inaccuracy or ambiguity within a British Standard or other BSI publication please inform the Knowledge Centre.

## Copyright

All the data, software and documentation set out in all British Standards and other BSI publications are the property of and copyrighted by BSI, or some person or entity that owns copyright in the information used (such as the international standardization bodies) and has formally licensed such information to BSI for commercial publication and use. Except as permitted under the Copyright, Designs and Patents Act 1988 no extract may be reproduced, stored in a retrieval system or transmitted in any form or by any means – electronic, photocopying, recording or otherwise – without prior written permission from BSI. Details and advice can be obtained from the Copyright & Licensing Department.

## Useful Contacts:

### Customer Services

**Tel:** +44 845 086 9001

**Email (orders):** [orders@bsigroup.com](mailto:orders@bsigroup.com)

**Email (enquiries):** [cservices@bsigroup.com](mailto:cservices@bsigroup.com)

### Subscriptions

**Tel:** +44 845 086 9001

**Email:** [subscriptions@bsigroup.com](mailto:subscriptions@bsigroup.com)

### Knowledge Centre

**Tel:** +44 20 8996 7004

**Email:** [knowledgecentre@bsigroup.com](mailto:knowledgecentre@bsigroup.com)

### Copyright & Licensing

**Tel:** +44 20 8996 7070

**Email:** [copyright@bsigroup.com](mailto:copyright@bsigroup.com)

## BSI Group Headquarters

389 Chiswick High Road London W4 4AL UK

