

BS ISO 18241:2016



BSI Standards Publication

Cardiovascular implants and extracorporeal systems — Cardiopulmonary bypass systems — Venous bubble traps

National foreword

This British Standard is the UK implementation of ISO 18241:2016.

The UK participation in its preparation was entrusted to Technical Committee CH/150/2, Cardiovascular implants.

A list of organizations represented on this committee can be obtained on request to its secretary.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

© The British Standards Institution 2016.
Published by BSI Standards Limited 2016

ISBN 978 0 580 81390 0

ICS 11.040.40

Compliance with a British Standard cannot confer immunity from legal obligations.

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 31 August 2016.

Amendments/corrigenda issued since publication

Date	Text affected
------	---------------

INTERNATIONAL
STANDARD

ISO
18241

First edition
2016-08-15

**Cardiovascular implants and
extracorporeal systems —
Cardiopulmonary bypass systems —
Venous bubble traps**

Implants cardiovasculaires et systèmes extracorporels — Systèmes de pontage cardiopulmonaire — Pièges à bulles veineuses



Reference number
ISO 18241:2016(E)

© ISO 2016



COPYRIGHT PROTECTED DOCUMENT

© ISO 2016, Published in Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Ch. de Blandonnet 8 • CP 401
CH-1214 Vernier, Geneva, Switzerland
Tel. +41 22 749 01 11
Fax +41 22 749 09 47
copyright@iso.org
www.iso.org

Contents

Page

Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Requirements	3
4.1 Biological characteristics.....	3
4.1.1 Sterility and non-pyrogenicity.....	3
4.1.2 Biocompatibility.....	3
4.2 Physical characteristics.....	3
4.2.1 Blood pathway integrity.....	3
4.2.2 Prime volume.....	3
4.2.3 Connectors.....	3
4.3 Performance characteristics.....	3
4.3.1 Blood cell damage.....	3
4.3.2 Air removal efficiency.....	3
4.3.3 Flow rate capacity.....	3
4.3.4 Shelf life.....	4
5 Tests and measurements to determine compliance with this document	4
5.1 General.....	4
5.2 Biological characteristics.....	4
5.2.1 Sterility and non-pyrogenicity.....	4
5.2.2 Biocompatibility.....	4
5.3 Physical characteristics.....	4
5.3.1 Blood pathway integrity.....	4
5.3.2 Prime volume.....	4
5.3.3 Connectors.....	4
5.4 Performance characteristics.....	5
5.4.1 Blood cell damage.....	5
5.4.2 Air removal efficiency.....	6
5.4.3 Flow rate and pressure drop.....	6
5.4.4 Shelf life.....	7
6 Information supplied by the manufacturer	7
6.1 Information on the venous bubble trap.....	7
6.2 Information on the packaging.....	7
6.2.1 Information on the unit container.....	7
6.2.2 Information on the shipping container.....	7
6.3 Information in the accompanying documents.....	8
6.4 Information in the accompanying documents in a prominent form.....	8
7 Packaging	8
Bibliography	9

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: www.iso.org/iso/foreword.html.

The committee responsible for this document is ISO/TC 150, *Implants for surgery*, Subcommittee SC 2, *Cardiovascular implants and extracorporeal systems*.

Introduction

This document is intended to ensure that devices designed to remove air entering the venous line during surgical procedures requiring extracorporeal circulatory support have been adequately tested for both their safety and function, and that extracorporeal device characteristics are appropriately disclosed when labeling the device.

This document therefore contains procedures to be used for evaluation of extracorporeal venous bubble traps. Test procedures for determination of the air removal efficiency, blood cell damage and other performance characteristics are described, although limits for these characteristics are not specified. Ready identification of the performance characteristics should, however, assist the user in the selection of a venous bubble trap that will suit the needs of the patient.

This document also includes minimum reporting requirements, which will allow the user to compare performance characteristics of venous bubble traps of different designs in a standard way.

This document makes reference to other International Standards in which methods for determination of characteristics common to medical devices can be found.

Requirements for animal and clinical studies have not been included in this document.

Such studies may be part of a manufacturer's quality system.

This document contains only those requirements that are specific to venous bubble traps. Nonspecific requirements are covered by references to other International Standards listed in the normative references section.

Cardiovascular implants and extracorporeal systems — Cardiopulmonary bypass systems — Venous bubble traps

1 Scope

This document specifies requirements for sterile, single-use, venous bubble traps intended to remove air entering the venous line during surgical procedures requiring extracorporeal circulatory support, which may include cardiopulmonary bypass (CPB), extracorporeal membrane oxygenation (ECMO), or venovenous bypass for liver transplantation.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 594-2, *Conical fittings with 6 % (Luer) taper for syringes, needles and certain other medical equipment — Part 2: Lock fittings*

ISO 10993-1, *Biological evaluation of medical devices — Part 1: Evaluation and testing within a risk management process*

ISO 10993-4, *Biological evaluation of medical devices — Part 4: Selection of tests for interaction with blood*

ISO 10993-7, *Biological evaluation of medical devices — Part 7: Ethylene oxide sterilization residuals*

ISO 10993-11, *Biological evaluation of medical devices — Part 11: Tests for systemic toxicity*

ISO 11135-1, *Sterilization of health care products — Ethylene oxide — Part 1: Requirements for development, validation and routine control of a sterilization process for medical devices*

ISO 11137-1, *Sterilization of health care products — Radiation — Part 1: Requirements for development, validation and routine control of a sterilization process for medical devices*

ISO 11607-1, *Packaging for terminally sterilized medical devices — Part 1: Requirements for materials, sterile barrier systems and packaging systems*

ISO 11607-2, *Packaging for terminally sterilized medical devices — Part 2: Validation requirements for forming, sealing and assembly processes*

ISO 14937, *Sterilization of health care products — General requirements for characterization of a sterilizing agent and the development, validation and routine control of a sterilization process for medical devices*

ISO 17665-1, *Sterilization of health care products — Moist heat — Part 1: Requirements for the development, validation and routine control of a sterilization process for medical devices*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses.

- ISO Online browsing platform: available at <https://www.iso.org/obp/>
- IEC Electropedia: available at <http://www.electropedia.org/>

3.1

venous bubble trap

device for removing air from the venous line of an extracorporeal circuit

3.2

blood pathway

blood-contacting surfaces of the venous bubble trap during its intended clinical use

3.3

blood cell damage

loss or destruction of cellular components of the blood

3.4

platelet reduction

percentage reduction of platelets contained in a circuit incorporating a venous bubble trap, as a function of time

3.5

plasma-free hemoglobin level

difference between the concentration of plasma-free hemoglobin in a circuit incorporating a venous bubble trap, as a function of time

3.5.1

normalized index of hemolysis

NIH

grams of plasma-free hemoglobin released after pumping 100 l of blood

$$NIH \left\{ g / 100 L \right\} = \Delta fHb \times V \times \frac{100 - Hct}{100} \times \frac{100}{Q \times T}$$

where

ΔfHb is the increase of plasma free hemoglobin concentration (g/L) over the sampling time interval;

V is the circuit volume (L);

Q is the flow rate (L/min);

Hct is the hematocrit (%);

T is the sampling time interval (min)

3.6

white blood cell reduction

percentage reduction of white blood cells contained in a circuit incorporating a venous bubble trap, as a function of time

3.7

air removal efficiency

ability of the venous bubble trap to remove air from the blood, expressed as a percentage

3.8

blood analogue

test solution which simulates blood viscosity between $2,0 \times 10^{-3}$ Pa·s (2,0 cP), to $3,5 \times 10^{-3}$ Pa·s (3,5 cP)

3.9

predicate venous bubble trap

similar venous bubble trap to the test venous bubble trap that has previously been approved and used for the same intended clinical use

4 Requirements

4.1 Biological characteristics

4.1.1 Sterility and non-pyrogenicity

The blood pathway shall be sterile and non-pyrogenic. Compliance shall be verified in accordance with [5.2.1](#).

4.1.2 Biocompatibility

The parts of the blood pathway shall be biocompatible with respect to their intended use. Compliance shall be verified in accordance with [5.2.2](#).

4.2 Physical characteristics

4.2.1 Blood pathway integrity

When tested in accordance with [5.3.1](#), the blood pathway shall not leak.

4.2.2 Prime volume

The volume of the blood pathway shall be within the tolerances specified by the manufacturer (see [6.3](#)).

4.2.3 Connectors

Connectors for connection to the blood pathway shall, when tested in accordance with [5.3.3](#), allow a secure connection.

NOTE 1 Connectors of a type that allows connection of tubes with an inside diameter of 4,8 mm, 6,3 mm, 9,5 mm or 12,7 mm, or a type that complies with ISO 8637:2010, Figure 1, or a type that complies with ISO 594-2, have been found satisfactory.

Connection for accessory ports shall meet the requirements of ISO 594-2.

NOTE 2 Connectors corresponding to ISO 8637:2010, Figure 3, are considered as one way to comply with this requirement.

4.3 Performance characteristics

4.3.1 Blood cell damage

When determined in accordance with [5.4.1](#), the percentage change (positive or negative) of plasma-free hemoglobin, platelets, and white blood cells, shall be within the range of values specified by the manufacturer.

The hemolysis results shall be reported as mg/dL and NIH.

4.3.2 Air removal efficiency

When tested in accordance with [5.4.2](#), the air removal efficiency shall be as expressed as a percentage. The manufacturer should specify the air challenge conditions. The test methodology should account for and measure gaseous microemboli for size and number and a second measurement of gross air volume.

4.3.3 Flow rate capacity

When tested in accordance with [5.4.3](#), the test results shall demonstrate the flow rate and pressure limitation(s), as specified by the manufacturer.

4.3.4 Shelf life

When tested in accordance with [5.4.4](#), the test results shall demonstrate the rated shelf life, as specified by the manufacturer.

5 Tests and measurements to determine compliance with this document

5.1 General

5.1.1 Tests and measurements shall be performed with the device in its terminally sterilized form and prepared according to the manufacturer's instructions for intended clinical use.

5.1.2 Operating variables shall be those specified by the manufacturer for intended clinical use, unless otherwise specified.

5.1.3 Unless otherwise stated, the temperature of test liquids shall be $37\text{ °C} \pm 1\text{ °C}$.

5.1.4 If the relationship between variables is non-linear, sufficient determinations shall be made to permit valid interpolation between data points.

5.1.5 The test or measurement procedures shall be regarded as reference procedures. Other procedures can be accepted, provided that the alternative procedure has been shown to be of comparable precision.

5.2 Biological characteristics

5.2.1 Sterility and non-pyrogenicity

Compliance shall be verified by inspection of the manufacturer's documentation on sterilization and pyrogen testing, in accordance with ISO 10993-11, ISO 11135-1, ISO 11137-1, ISO 14937 or ISO 17665-1, as applicable.

5.2.2 Biocompatibility

Compliance shall be verified by test or by inspection of the manufacturer's documentation on biocompatibility for the finished device, in accordance with ISO 10993-1 and ISO 10993-7, as applicable.

5.3 Physical characteristics

5.3.1 Blood pathway integrity

Fill the blood pathway of the device with water and subject it to a negative pressure of $1,5 \times$ the manufacturer's rated negative pressure and maintain the pressure for 6 h or for the intended time of use specified by the manufacturer. Visually inspect the device for evidence of air entrainment.

5.3.2 Prime volume

The test liquid shall be anticoagulated whole blood or water. The volume of the blood pathway shall be determined as specified by the manufacturer.

5.3.3 Connectors

The connection shall be made in accordance with the manufacturer's instructions for use. The connection shall withstand a pull force of 15 N for 15 s without separating.

5.4 Performance characteristics

5.4.1 Blood cell damage

5.4.1.1 Test liquid

The test liquid for the blood pathway shall be anticoagulated whole blood.

5.4.1.2 Procedure

Two sets of appropriate, identical circuit components, including a pump, connecting tubing, a reservoir (as specified by the manufacturer and of suitable size relative to the device under test), and a means of controlling temperature, shall be assembled. The device under test shall be placed in one of the circuits between the inflow of a blood pump and the outlet of a patient simulating reservoir per the manufacturer's instructions for use. A predicate device shall be placed in the second test circuit. Priming and debubbling of the circuits by recirculating with an appropriate solution is recommended before blood is added. The blood pathway test liquid volumes shall, at the initiation of the test, be within 1 % of each other. Perform the test *in vitro* using the conditions given in [Table 1](#). If the instructions for use call for connection to ancillary devices, these should be connected and operated at the worst case conditions allowed per the manufacturer's instructions for use. A sufficient number of paired tests should be performed to support a statistical analysis. The predicate venous bubble trap should be tested under the same conditions. Compliance shall be verified by test or by inspection of the manufacturer's documentation on blood cell damage for the finished device, in accordance with ISO 10993-4, as applicable.

Table 1 — Conditions for in vitro testing of blood cell damage

Item	Level	Maximum variation
Blood flow rate	The maximum specified by the manufacturer for intended clinical use (see 6.3)	±5 %
Blood glucose	10 mmol/L	±5 mmol/L
Hemoglobin	12 g/dl	±1 g/dl

The sampling schedule shall be in accordance with [Table 2](#). More frequent sampling times are optional.

Table 2 — Sampling schedule

Parameter	Time, after initiation of test			
	Prior to test	30 min	180 min	360 min
Plasma-free hemoglobin	X	X	X	X
White blood cell	X	X	X	X
Platelets	X	X	X	X
Hemoglobin	X	X	X	X
Glucose	X			
Activated clotting time	X	X	X	X
Temperature	X	X	X	X
Flow rates	X	X	X	X

5.4.2 Air removal efficiency

5.4.2.1 Test liquid

The test liquid shall be anticoagulated whole blood with a hemoglobin content of (12 ± 1) g/dl.

5.4.2.2 Procedure

Prepare a test set up which places the venous bubble trap, per the manufacturer's instructions for use, between the inflow of a blood pump and the outlet of a patient simulating reservoir. (If the instructions for use call for connection to ancillary devices such as active air removal systems, these should be connected per the instructions for use.)

The inflow tubing between the venous bubble trap and the reservoir should include an access connector located upstream of the venous bubble trap for injecting the air challenge and an inline micro bubble detector upstream from this connector. Another inline micro bubble detector should be located after the venous bubble trap and the pump.

The test liquid should have a baseline micro bubble value of <1% of the air challenge bolus volume.

After the baseline is stabilized, introduce air at the injected site. The volume of air injected and the method of injection must be defined in the test procedure. In addition, a bubble volume collection chamber placed before the micro bubble detector may be used as an adjunctive method to determine air removal efficiency. The test should be performed at the minimum and maximum flow rates recommended by the manufacturer as well at one flow rate midway between the minimum and maximum.

NOTE This procedure will produce bubbles of different sizes.

The total volume of air downstream of the venous bubble trap should be measured by the micro bubble detector placed between the venous bubble trap and the pump. If a bubble volume collection chamber is used, the volume of air collected should be measured and added to the volume of air measured by the micro bubble detector to document total volume of air downstream of the venous bubble trap. The volume of air collected should be measured and the bigger between the volume provided by the micro bubble detector and the air collected will be considered the total volume of air downstream of the venous bubble trap

5.4.2.3 Results

The results shall be reported as the percentage efficiency of gross air removal.

$$\text{EFF} = (\text{BOLUS VOLUME} - \text{DOWNSTREAM VOLUME}) / \text{BOLUS VOLUME}$$

NOTE The downstream volume is the sum of the volume measured by the microbubble detector plus the volume collected by the collection chamber.

5.4.3 Flow rate and pressure drop

5.4.3.1 Test liquid

The test liquid shall be anticoagulated whole blood with a hemoglobin content of (12 ± 1) g/dl or blood analogue.

NOTE This test may be performed concurrently with the blood trauma test (see [5.4.1](#)).

5.4.3.2 Procedure

Place the device under test in an appropriate test circuit. Set the flow rate at the maximum rated flow and monitor the inlet and outlet pressures across the venous bubble trap for 6 h or greater according

to the manufacturer's instructions for use. Measure the flow rate using a calibrated flow meter. Note any pressure changes during the test. If anticoagulated whole blood is used this test shall not take into account the effects of formed elements or proteinaceous aggregates.

5.4.4 Shelf life

Using a validated method, ageing should be performed on final, finished, sterilized, devices in primary packaging in order to determine nominal shelf life.

6 Information supplied by the manufacturer

6.1 Information on the venous bubble trap

The following information shall be given on the venous bubble trap:

- a) the manufacturer's identity;
- b) model designation;
- c) the direction of blood flow.

6.2 Information on the packaging

6.2.1 Information on the unit container

The following shall be visible through or given on the unit container:

- a) the manufacturer's name and address;
- b) description of contents;
- c) model designation;
- d) statement on sterility and method of sterilization and non-pyrogenicity;
- e) expiry date;
- f) batch, lot or serial number designation;
- g) the words, "Read instructions before use" or equivalent symbol;
- h) any special handling or storage conditions;
- i) statement on single-use.

6.2.2 Information on the shipping container

The following information shall appear on the shipping container:

- a) the manufacturer's name and address;
- b) description of contents, including number of units;
- c) model designation;
- d) expiry date;
- e) any special handling, storage or unpacking instructions.

6.3 Information in the accompanying documents

Each shelf box shall contain an "Instructions for Use" leaflet with the following information:

- a) the manufacturer's address and telephone or fax numbers;
- b) model designation;
- c) required ancillary equipment;
- d) instructions on necessary, special or unique procedures, as applicable;
- e) directions for placing the bubble trap in a support or operational fixture;
- f) placement, type and securing of tubing connections;
- g) location and purpose of additional entry or exit ports;
- h) priming procedure;
- i) direction of blood flow;
- j) general operating procedures for normal use;
- k) air removal efficiency;
- l) maximum and minimum recommended blood flow rates;
- m) blood pathway pressure drop over the range of blood flow rates specified by the manufacturer; for intended clinical use;
- n) priming volume;
- o) a statement that the following are available upon request:
 - 1) a list of materials of the blood pathway;
 - 2) data related to blood cell damage;
 - 3) relevant tolerances for data presented;
- p) statement on sterility, method of sterilization, and non-pyrogenicity.

6.4 Information in the accompanying documents in a prominent form

The following information shall be given in accompanying documents in a prominent form:

- a) flow rate limitations;
- b) other device limitations, for example, material incompatibility with known volatile anesthetic agents, solvents or disinfectants.

7 Packaging

Packaging shall comply with the appropriate requirements of ISO 11607-1 and ISO 11607-2.

Bibliography

- [1] ISO 8637:2010, *Cardiovascular implants and extracorporeal systems — Haemodialysers, haemodiafilters, haemofilters and haemoconcentrators*
- [2] HUYBREGTS R.M.A.J.M., VEERMAN D.P., VONK A.B.A., NESSELAAR A.F., PAULUS R.C.E., THONE-PASSCHIER D.H. First clinical experience with the air purge control and electrical remote-controlled tubing clamp in mini bypass. *Artif. Organs*. 2006, **30** pp. 721–724
- [3] ROSENHOFF T.P.A., STEHOUWER M.C., de VROEGE R., BUTTER R.P., VAN BOVEN W.-J., BRUINS P. Air removal efficiency of a venous bubble trap in a minimal extracorporeal circuit during coronary artery bypass grafting. *Artif. Organs*. 2010, **34** pp. 1092–1098
- [4] SIMONS A.P., & WEERWIND P.W. Letter: Microbubble formation during minimized cardiopulmonary bypass. *Artif. Organs*. 2011, **35** p. 554
- [5] STEHOUWER M. Letter: Air removal efficiency of a venous bubble trap in a minimal extracorporeal circuit during coronary artery bypass. *Artif. Organs*. 2011, **35** p. 555

British Standards Institution (BSI)

BSI is the national body responsible for preparing British Standards and other standards-related publications, information and services.

BSI is incorporated by Royal Charter. British Standards and other standardization products are published by BSI Standards Limited.

About us

We bring together business, industry, government, consumers, innovators and others to shape their combined experience and expertise into standards-based solutions.

The knowledge embodied in our standards has been carefully assembled in a dependable format and refined through our open consultation process. Organizations of all sizes and across all sectors choose standards to help them achieve their goals.

Information on standards

We can provide you with the knowledge that your organization needs to succeed. Find out more about British Standards by visiting our website at bsigroup.com/standards or contacting our Customer Services team or Knowledge Centre.

Buying standards

You can buy and download PDF versions of BSI publications, including British and adopted European and international standards, through our website at bsigroup.com/shop, where hard copies can also be purchased.

If you need international and foreign standards from other Standards Development Organizations, hard copies can be ordered from our Customer Services team.

Copyright in BSI publications

All the content in BSI publications, including British Standards, is the property of and copyrighted by BSI or some person or entity that owns copyright in the information used (such as the international standardization bodies) and has formally licensed such information to BSI for commercial publication and use.

Save for the provisions below, you may not transfer, share or disseminate any portion of the standard to any other person. You may not adapt, distribute, commercially exploit, or publicly display the standard or any portion thereof in any manner whatsoever without BSI's prior written consent.

Storing and using standards

Standards purchased in soft copy format:

- A British Standard purchased in soft copy format is licensed to a sole named user for personal or internal company use only.
- The standard may be stored on more than 1 device provided that it is accessible by the sole named user only and that only 1 copy is accessed at any one time.
- A single paper copy may be printed for personal or internal company use only.

Standards purchased in hard copy format:

- A British Standard purchased in hard copy format is for personal or internal company use only.
- It may not be further reproduced – in any format – to create an additional copy. This includes scanning of the document.

If you need more than 1 copy of the document, or if you wish to share the document on an internal network, you can save money by choosing a subscription product (see 'Subscriptions').

Reproducing extracts

For permission to reproduce content from BSI publications contact the BSI Copyright & Licensing team.

Subscriptions

Our range of subscription services are designed to make using standards easier for you. For further information on our subscription products go to bsigroup.com/subscriptions.

With **British Standards Online (BSOL)** you'll have instant access to over 55,000 British and adopted European and international standards from your desktop. It's available 24/7 and is refreshed daily so you'll always be up to date.

You can keep in touch with standards developments and receive substantial discounts on the purchase price of standards, both in single copy and subscription format, by becoming a **BSI Subscribing Member**.

PLUS is an updating service exclusive to BSI Subscribing Members. You will automatically receive the latest hard copy of your standards when they're revised or replaced.

To find out more about becoming a BSI Subscribing Member and the benefits of membership, please visit bsigroup.com/shop.

With a **Multi-User Network Licence (MUNL)** you are able to host standards publications on your intranet. Licences can cover as few or as many users as you wish. With updates supplied as soon as they're available, you can be sure your documentation is current. For further information, email subscriptions@bsigroup.com.

Revisions

Our British Standards and other publications are updated by amendment or revision.

We continually improve the quality of our products and services to benefit your business. If you find an inaccuracy or ambiguity within a British Standard or other BSI publication please inform the Knowledge Centre.

Useful Contacts

Customer Services

Tel: +44 345 086 9001

Email (orders): orders@bsigroup.com

Email (enquiries): cservices@bsigroup.com

Subscriptions

Tel: +44 345 086 9001

Email: subscriptions@bsigroup.com

Knowledge Centre

Tel: +44 20 8996 7004

Email: knowledgecentre@bsigroup.com

Copyright & Licensing

Tel: +44 20 8996 7070

Email: copyright@bsigroup.com

BSI Group Headquarters

389 Chiswick High Road London W4 4AL UK