

# Essential oils — Analysis by gas chromatography on chiral capillary columns — General method

ICS 71.100.60

## National foreword

This British Standard reproduces verbatim ISO 22972:2004 and implements it as the UK national standard.

The UK participation in its preparation was entrusted to Technical Committee AW/54, Essential oils, which has the responsibility to:

- aid enquirers to understand the text;
- present to the responsible international/European committee any enquiries on the interpretation, or proposals for change, and keep the UK interests informed;
- monitor related international and European developments and promulgate them in the UK.

A list of organizations represented on this committee can be obtained on request to its secretary.

### Cross-references

The British Standards which implement international publications referred to in this document may be found in the *BSI Catalogue* under the section entitled “International Standards Correspondence Index”, or by using the “Search” facility of the *BSI Electronic Catalogue* or of British Standards Online.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

**Compliance with a British Standard does not of itself confer immunity from legal obligations.**

### Summary of pages

This document comprises a front cover, an inside front cover, the ISO title page, pages ii to iv, pages 1 to 4, an inside back cover and a back cover.

The BSI copyright notice displayed in this document indicates when the document was last issued.

### Amendments issued since publication

Amd. No.	Date	Comments

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 4 June 2004

© BSI 4 June 2004

ISBN 0 580 43839 2

INTERNATIONAL  
STANDARD

**ISO**  
**22972**

First edition  
2004-05-15

---

---

**Essential oils — Analysis by gas  
chromatography on chiral capillary  
columns — General method**

*Huiles essentielles — Analyse par chromatographie en phase gazeuse  
sur colonne capillaire chirale — Méthode générale*



Reference number  
ISO 22972:2004(E)



## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 22972 was prepared by Technical Committee ISO/TC 54, *Essential oils*.

## Introduction

Since the description of gas chromatographic analysis methods is long, it has been considered worthwhile to draw up general standards giving detailed information on all the recurrent parameters (apparatus, products, methods, calculation formulae, etc.), and then short-version standards (abridged analysis sheets) relating to the determination of the enantiomeric excess or distribution of chiral compounds present in essential oils.

This is the case with this International Standard, which makes reference to ISO 7609 for the general clauses.

# Essential oils — Analysis by gas chromatography on chiral capillary columns — General method

## 1 Scope

This International Standard specifies a general method for the analysis of essential oils by gas chromatography on capillary chiral columns, for the purpose of determining the specific enantiomeric excess or distribution of the chiral compounds contained in the essential oils.

## 2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 356, *Essential oils — Preparation of test samples*

ISO 7609:1985, *Essential oils — Analysis by gas chromatography on capillary columns — General method*

## 3 Principle

The essential oil is analysed by gas chromatography, under specified analytical conditions, on a column of appropriate diameter and length, the inside wall of this column having been previously impregnated, either directly with a bonded chiral stationary phase or with a support of this stationary phase.

The different chiral constituents of interest in the essential oil are identified with respect to standard substances and, if necessary, by measurement of their retention index and determination of the relative proportions of the enantiomers of a constituent.

## 4 Reagents and products

During the course of the analysis, unless otherwise specified, use only reagents of recognized analytical grade.

### 4.1 Gas.

**4.1.1 Carrier gas:** hydrogen, helium or nitrogen, or any other gas, depending on the type of detector being used.

If a detector requiring a carrier gas other than those mentioned is used, this should be stated in the test report.

**4.1.2 Detector supply gas:** all suitable gases, depending on the detector being used:

— for a flame ionization detector, air and hydrogen of high purity;

— for a mass detector, helium of high purity.

**4.2 Reference substance,** corresponding to the pure and mixed enantiomers to be detected.

## 5 Apparatus

**5.1 Chromatograph**, equipped with an injection system, an appropriate detector and a temperature programmer.

The injection and detection systems shall be fitted with devices for independent control of their respective temperatures.

**5.2 Capillary column**, made of an inert material, having an inner diameter  $\leq 0,53$  mm and a length between 10 m and 60 m.

NOTE Common chiral stationary phases are based on substituted  $\beta$ -cyclodextrins (e.g. 1,3,6-permethylated  $\beta$ -cyclodextrin at 10 % in OV 101®).

**5.3 Recorder and electronic integrator.**

## 6 Preparation of the test sample

See ISO 356.

The test sample may have to undergo the following special preparation:

- selective transfer of the relevant constituent(s) (e.g. in-line two-dimensional chromatography or trapping);
- derivatization prior to injection.

If it is the case, this will be mentioned in the standard on the relevant essential oil.

## 7 Operating conditions

### 7.1 Oven temperature

The oven temperature shall be regulated in such a way as to obtain the required efficiency (see 8.1).

### 7.2 Carrier gas flow rate

The flow rate shall be regulated in such a way as to obtain the required efficiency (see 8.1).

### 7.3 Flow rate of detector supply gases

Refer to the appliance manufacturer's instructions in order to obtain the optimum response from the detector.

## 8 Column performance

### 8.1 Column efficiency

Determine the column efficiency on the less-retained enantiomer of the selected reference substance (see 4.2), according to the method described in ISO 7609.

The efficiency shall be at least 25 000 theoretical plates.

NOTE When a two-dimensional chromatographic system is used, the temperature programming conditions can be optimized for each component of interest (or chemical family), so as to obtain a satisfactory resolution with a reduced analysis time. In this case, the efficacy may be lower than 25 000 theoretical plates.



## 8.2 Resolution and separation

Determine the resolution by injecting a quantity of the selected reference substance (see 4.2) under the test conditions. Carry out the calculations in accordance with ISO 7609:1985, 8.3.1 and 8.3.2.

The injection of the reference substance under analysis conditions shall not lead to any deterioration of the enantiomers or to any variation in their ratio.

## 9 Determination of the retention indexes

See ISO 7609.

NOTE When a two-dimensional chromatographic system is used, the determination of the retention indexes is optional if the conditions of elution on both chromatographs and the resolution of the enantiomer are similar to those of the reference substances, within the random variations.

## 10 Determination of the enantiomeric excess or distribution

Use the same temperature and flow rate conditions as those adopted for the column efficiency test (see 8.1). After stabilization of the column temperature, inject an appropriate quantity of the test portion.

On the obtained essential oil chromatogram, verify, using the indications provided by the integrator, compliance with the limits set out in the standards specific to the relevant essential oil.

Particularly in the case of use of a mono-dimensional chromatographic system, it is recommended to verify by all appropriate means (e.g. mass detector) the absence of co-elution for the substances being sought.

## 11 Expression of results

### 11.1 General

Express the results either in enantiomeric excess or in enantiomeric distribution in accordance with the requirements of the standard specific to the relevant essential oil.

### 11.2 Enantiomeric excess: $E$

Determine the enantiomeric excess,  $E$ , from the area of each of the enantiomer peaks:

$$E_x = \frac{(A_{\max} - A_{\min})}{(A_{\max} + A_{\min})} \times 100$$

where

$x = (R)$  if  $A_{\max}$  is for enantiomer (R);

$x = (S)$  if  $A_{\max}$  is for enantiomer (S);

$A_{\max}$  is the peak area presenting the largest surface;

$A_{\min}$  is the peak area presenting the smallest surface.

### 11.3 Enantiomeric distribution

$$(R) \% = \frac{A(R)}{A(R) + A(S)} \times 100$$

$$(S) \% = \frac{A(S)}{A(R) + A(S)} \times 100$$

where

$A(R)$  is the enantiomer (R) peak area;

$A(S)$  is the enantiomer (S) peak area.

In the event of the enantiomeric excess or distribution being respectively greater than 95 % or 97,5 %, it is recommended to pay particular attention to the choice of the integration parameters in order to minimize the error on the quantification of the least abundant enantiomer. The latter will be quantified, as far as possible, under acceptable chromatographic resolution conditions.

## 12 Test report

See ISO 7609.



---

---

## BSI — British Standards Institution

BSI is the independent national body responsible for preparing British Standards. It presents the UK view on standards in Europe and at the international level. It is incorporated by Royal Charter.

### Revisions

British Standards are updated by amendment or revision. Users of British Standards should make sure that they possess the latest amendments or editions.

It is the constant aim of BSI to improve the quality of our products and services. We would be grateful if anyone finding an inaccuracy or ambiguity while using this British Standard would inform the Secretary of the technical committee responsible, the identity of which can be found on the inside front cover.  
Tel: +44 (0)20 8996 9000. Fax: +44 (0)20 8996 7400.

BSI offers members an individual updating service called PLUS which ensures that subscribers automatically receive the latest editions of standards.

### Buying standards

Orders for all BSI, international and foreign standards publications should be addressed to Customer Services. Tel: +44 (0)20 8996 9001.  
Fax: +44 (0)20 8996 7001. Email: [orders@bsi-global.com](mailto:orders@bsi-global.com). Standards are also available from the BSI website at <http://www.bsi-global.com>.

In response to orders for international standards, it is BSI policy to supply the BSI implementation of those that have been published as British Standards, unless otherwise requested.

### Information on standards

BSI provides a wide range of information on national, European and international standards through its Library and its Technical Help to Exporters Service. Various BSI electronic information services are also available which give details on all its products and services. Contact the Information Centre.  
Tel: +44 (0)20 8996 7111. Fax: +44 (0)20 8996 7048. Email: [info@bsi-global.com](mailto:info@bsi-global.com).

Subscribing members of BSI are kept up to date with standards developments and receive substantial discounts on the purchase price of standards. For details of these and other benefits contact Membership Administration.  
Tel: +44 (0)20 8996 7002. Fax: +44 (0)20 8996 7001.  
Email: [membership@bsi-global.com](mailto:membership@bsi-global.com).

Information regarding online access to British Standards via British Standards Online can be found at <http://www.bsi-global.com/bsonline>.

Further information about BSI is available on the BSI website at <http://www.bsi-global.com>.

### Copyright

Copyright subsists in all BSI publications. BSI also holds the copyright, in the UK, of the publications of the international standardization bodies. Except as permitted under the Copyright, Designs and Patents Act 1988 no extract may be reproduced, stored in a retrieval system or transmitted in any form or by any means – electronic, photocopying, recording or otherwise – without prior written permission from BSI.

This does not preclude the free use, in the course of implementing the standard, of necessary details such as symbols, and size, type or grade designations. If these details are to be used for any other purpose than implementation then the prior written permission of BSI must be obtained.

Details and advice can be obtained from the Copyright & Licensing Manager.  
Tel: +44 (0)20 8996 7070. Fax: +44 (0)20 8996 7553.  
Email: [copyright@bsi-global.com](mailto:copyright@bsi-global.com).