

BRITISH STANDARD

**BS 2000 :
Part 130: 1998
ISO 3839 :1996**

Methods of test for

Petroleum and its products

**Part 130. Petroleum products - Determination
of bromine number of distillates and aliphatic
olefins - Electrometric method**

(Identical with IP 130/98)

National foreword

This British Standard was published under the authority of the Materials and Chemicals Sector Board and comes into effect on 1st March 1998. It is the English language version of ISO 3839 : 1996 *Petroleum products - Determination of bromine number of distillates and aliphatic olefins - Electrometric method*, published by the International Organization for Standardization (ISO).

This British Standard supersedes BS 5089 : 1980 (1988), which will be withdrawn.

BS 2000 comprises a series of test methods for petroleum and its products that are published by the Institute of Petroleum (IP) and have been accorded the status of a British Standard. Each method should be read in conjunction with the preliminary pages of 'IP Standard methods for analysis and testing of petroleum and related products and British Standard 2000 Parts' which gives details of the BSI/IP agreement for publication of the series, provides general information on safety precautions, sampling and other matters, and lists the methods published as Parts of BS 2000.

Under the terms of the agreement between BSI and the Institute of Petroleum, BS 2000 : Part 130 / ISO 3839 will be published by the IP (in 'Standard methods for analysis and testing of petroleum and related products and British Standard 2000 Parts' and as a separate publication). The numbering of the Parts of BS 2000 follows that of the corresponding IP methods. BS 2000 : Part 130 : 1998 is thus identical with IP 130/98.

Compliance with a British Standard does not of itself confer immunity from legal obligations.

© The Institute of Petroleum
& BSI 1998

The following BSI references
relate to the work of this
standard:
Committee reference PTI/13
Draft for public comment no:
94/501004 DC

Petroleum products - Determination of bromine number of distillates and aliphatic olefins - Electrometric method

WARNING - The use of this international standard may involve hazardous materials, operations and equipment. This standard does not purport to address all of the safety problems associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.

1 Scope

This International Standard specifies a method for the determination of the bromine number of the following materials:

a) petroleum distillates that are substantially free of material lighter than 2-methylpropane, and that have 90 % (V/V) (i.e. volume fraction 90 %) distillation recovery temperatures under 327 °C. The method is generally applicable to gasolines (including leaded, unleaded and oxygenated fuels), kerosines and distillates in the gas oil range that fall within the following limits:

90 % (V/V) recovery distillation temperature (ISO 3405)	Bromine number, max. (see note 1)
Under 250 °C	175
205 °C to 327 °C	10

b) commercial olefins that are essentially mixtures of aliphatic monoolefins and that fall within the range of 95 to 165 bromine number (see note 1).

The method has been found suitable for such materials as commercial propene trimer and tetramer, butene dimer, and mixed nonenes, octenes and heptenes. The method is not suitable for normal alpha-olefins.

NOTES

1 These limits are imposed since the precision of the method has been determined only up to or within the range of these bromine numbers.

2 The value of the bromine number is an indication of the quantity of bromine-reactive constituents, not an identification of constituents. Annex A and table A.1 give information related to the use of this International Standard as a measure of olefinic unsaturation.

2 Normative references

The following standards contain provisions which, through reference in this text, constitute provisions of this International Standard. At the time of publication, the editions indicated were valid. All standards are subject to revision, and parties to agreements based on this International Standard are encouraged to investigate the possibility of applying the most recent editions of the standards as indicated below. Members of IEC and ISO maintain registers of currently valid International Standards.

ISO 3405: 1988, *Petroleum products - Determination of distillation characteristics*.

ISO 3696: 1987, *Water for analytical laboratory use - Specification and test methods*.

3 Definition

For the purposes of this international Standard, the following definition applies:

3.1 bromine number: Mass, in grams, of bromine which will combine with 100 g of the sample under standardized conditions.

4 Principle

A known mass of the test portion dissolved in a specific solvent maintained at 0 °C to 5 °C is titrated with standard volumetric bromide/bromate solution. The end-point is indicated by a sudden change in potential on an electrometric end-point titration apparatus due to the presence of free bromine.

5 Reagents and materials

During the analysis, use only reagents of recognised analytical grade, and water equivalent to grade 3 of ISO 3696.

5.1 1,1,1-Trichloroethane (CH₃CCl₃).

CAUTION - 1,1,1-trichloroethane is hazardous to the environment. A substitute is under active investigation.

5.2 Methanol (CH₃OH).

5.3 Potassium iodide solution, 150 g/l.

Dissolve 150 g of potassium iodide (KI) in water and dilute to 1 l.

5.4 Sulfuric acid, dilute solution (1:5).

Carefully mix 1 volume of concentrated sulfuric acid [H₂SO₄, 98 % (m/m) (i.e. mass fraction 98 %) min.] with 5 volumes of water.

5.5 Titration solvent.

Prepare 1 l of titration solvent by mixing the following volumes of materials: 714 ml of acetic acid (5.9), 134 ml of 1,1,1-trichloroethane (5.1), 134 ml of methanol (5.2) and 18 ml of sulfuric acid solution (5.4).

5.6 Bromide/bromate solution, [c(Br₂) = 0,250 mol/l].

Dissolve 51,0 g ± 0,1 g of potassium bromide (KBr) and 13,92 g ± 0,01 g of potassium bromate (KBrO₃), both dried at 105 °C for 30 min, in water and dilute to 1 l.

NOTE - If the bromine numbers of the reference olefins specified in clause 7 and determined using this solution do not conform to the limits specified, or if there is some uncertainty as to the quality of primary reagents, it is recommended that the concentration (mol/l) be determined (and used in subsequent calculations) by standardizing the solution. The standardization procedure shall be carried out as follows:

Place 50 ml of acetic acid (5.9) and 1 ml of concentrated hydrochloric acid (5.10) in a 500 ml iodine-number flask. Chill the solution in an ice bath for approximately 10 min, and with constant swirling of the contents of the flask, add from a 10 ml calibrated burette, 5,00 ml ± 0,01 ml of bromide/bromate solution being standardized at the rate of 1 drop/s or 2 drop/s. Stopper the flask immediately, shake the contents, place it again in the ice bath, and add 5 ml of potassium iodide solution (5.3) in the lip of the stoppered flask. After 5 min, remove the

flask from the ice bath and allow the potassium iodide solution to flow into the flask by slowly removing the stopper. Shake vigorously, add 100 ml of water in such a manner as to rinse the stopper, lip and walls of the flask, and titrate promptly with sodium thiosulfate solution (5.7). Near the end of the titration, add 1 ml of starch solution (5.8) and titrate slowly to disappearance of the blue colour. Calculate the concentration c_1 (Br₂), expressed in moles per litre, of the bromide/bromate solution as follows:

$$c_1 = \frac{V_0 c_0}{2V_1}$$

where

V_0 is the volume of sodium thiosulfate solution required for titration of the bromide/bromate solution, in millimetres;

V_1 is the volume of bromide/bromate solution, in millimetres (nominally 5,00);

c_0 is the concentration of the sodium thiosulfate solution, in moles per litre;

2 is the number of electrons transferred during redox titration of bromide/bromate.

Repeat the standardization until two successive determinations do not differ from their mean value by more than 0,002 mol/l.

5.7 Sodium thiosulfate solution, 0,1 mol/l.

Dissolve 25,0 g ± 0,1 g of sodium thiosulfate pentahydrate (Na₂S₂O₃·5H₂O) in water and add 0,01 g of sodium carbonate (Na₂CO₃) to stabilize the solution. Dilute to 1 l and mix thoroughly by shaking. Standardize by any accepted procedure that determines the concentration with an error not greater than ± 0,000 2 mol/l. Restandardize at intervals frequent enough to detect changes in concentration of ± 0,000 5 mol/l.

5.8 Starch solution

Grind and mix thoroughly 5 g of starch and 5 mg to 10 mg of mercury (II) iodide (HgI₂) with 3 ml to 5 ml of water. Add the suspension to 2 l boiling water and boil for 5 min to 10 min. Allow to cool and decant the clear, supernatant liquid into bottles having ground-glass stoppers.

CAUTION - Mercury (II) iodide is toxic. A substitute is under active investigation.

5.9 Acetic acid, glacial, 99,0 % (m/m) (i.e. mass fraction 99,0 %) minimum purity.

BROMINE NUMBER - ELECTROMETRIC METHOD, IP 130

5.10 Hydrochloric acid, concentrated, 35,4 % (*m/m*) (i.e. mass fraction 35,4 %) HCl.

5.11 Nitric acid, concentrated, 69,0 % (*m/m*) to 70,5 % (*m/m*) (i.e. mass fraction 69,0 % to 70,5 %).

6 Apparatus

6.1 Electrometric end-point titration apparatus.

Use any apparatus designed to perform titrations to pre-set end points in conjunction with a high-resistance polarizing current supply capable of maintaining approximately 0,8 V across two platinum electrodes, and with a sensitivity such that a voltage change of approximately 50 mV at these electrodes is sufficient to indicate the end-point.

NOTE - Other types of commercially available electronic titrimeters, including certain pH-meters, have also been found suitable.

6.2 Titration vessel.

A jacketed glass vessel approximately 120 mm high and 45 mm in internal diameter and of a form that can be conveniently maintained at 0 °C to 5 °C.

6.3 Stirrer.

Any magnetic stirrer system.

6.4 Electrodes.

A platinum wire electrode pair with each wire approximately 12 mm long and 1 mm in diameter. The wires shall be located 5 mm apart and approximately 55 mm below the level of the titration solvent. Clean the electrode pair at regular intervals with nitric acid (5.11) and rinse with water before use.

6.5 Burette.

Any delivery system capable of measuring titrant in 0,05 ml or smaller graduations.

7 Check test

If there are reservations in applying the procedure to actual test portions, check the reagents and techniques by means of determinations on freshly purified cyclohexene or diisobutene. Proceed in accordance with clause 8, using a test portion of 0,6 g to 1,0 g of either cyclohexene or diisobutene (see table 1), or 6 g to 10 g of a 10 % (*m/m*) (i.e. mass fraction 10 %) solution of these materials in

1,1,1-trichloroethane (5.1). If the reagents and techniques are correct, values within the following ranges will be obtained:

Standard	Bromine number
Cyclohexene, purified (see notes 1, 2 and 3)	187 to 199 (see note 4)
Cyclohexene, 10 % solution	18 to 20
Diisobutene, purified (see notes 2 and 3)	136 to 144 (see note 4)
Diisobutene 10 % solution	13 to 15

NOTES

1 Purified test samples of cyclohexene and diisobutene may be prepared from cyclohexene concentrates with a boiling range of 81 °C to 83 °C and from diisobutene (1-pentene, 2,2,4-trimethyl isomer only) concentrates with a boiling range of 100 °C to 102 °C, by the following procedure:

Add 65 g of activated silica (75 µm to 150 µm particle size, manufactured to ensure minimum olefin polymerization) to a column of approximately 16 mm inside diameter and 760 mm length, that has a stopcock at the lower end and that contains a small plug of glass wool immediately above the latter. A 100 ml burette, or any column providing a height-to-diameter ratio of the silica gel of at least 30:1, is suitable. Tap the column during addition of the gel to ensure uniform packing.

To the column add 30 ml of the olefin to be purified. When the olefin disappears into the gel, fill the column with methanol (5.2). Discard the first 10 ml of percolate and collect the next 10 ml, which is the purified olefin ready for use in the procedure for determining bromine number. Determine and record the density and refractive index of the purified test samples at 20 °C. Discard the remaining percolate.

2 If distillation of these olefins is required as a prepurification step, a few pellets of potassium hydroxide (KOH) should be placed in the distillation flask and distillation should not be continued beyond 90 °C (*V/V*) (i.e. volume fraction 90 %) recovery to minimise the hazards from decomposition of any peroxides that may be present.

3 The reference olefins yielding the above results are characterized by the properties given in table 1.

4 The theoretical bromine numbers of cyclohexene and diisobutene are 194,5 and 142,4 respectively.

Table 1 - Physical properties of purified olefins

Compound	Boiling point °C	Density at 20 °C kg/m ³	Refractive index n_D^{20}
Cyclohexene	82,5 to 83,5	810,0	1,446 5
Diisobutene	101,0 to 102,5	717,5 ± 1,5	1,411 2

8 Procedure

8.1 Place 10 ml of 1,1,1-trichloroethane (5.1) in a 50 ml volumetric flask and, by means of a pipette, introduce a quantity of sample as indicated in table 2. Obtain the mass of sample introduced either by taking the difference between the mass (to the nearest 1 mg) of the flask before and after addition of sample or, if the density is known accurately, by calculating the mass from the measured volume. Fill the flask to the mark with 1,1,1-trichloroethane and mix well.

Table 2 - Recommended test portion mass

Bromine number	Test portion mass g
0 to 10	20 to 16
Over 10 to 20	10 to 8
Over 20 to 50	5 to 4
Over 50 to 100	2 to 1,5
Over 100 to 150	1,0 to 0,8
Over 150 to 200	0,8 to 0,6

NOTES

1 If the order of magnitude of the bromine number of a test portion is unknown, a trial test is recommended using a 2 g test portion in order to obtain the approximate magnitude of the bromide number, followed by another determination using the appropriate test portion mass as indicated in table 2. The mass of the test portion should also be such that the volume of bromide/bromate titrant used does not exceed 10 ml, and that no separation of the reaction mixture into two phases occurs during the titration.

2 Difficulty may be experienced in dissolving test portions of the high-boiling range products in the titration solvent; this difficulty can be prevented by the addition of a small quantity of toluene.

8.2 Cool the titration vessel (6.2) to between 0 °C and 5 °C and maintain at this temperature throughout the titration. Switch on the trimeter (6.1) and allow the electrical circuit to stabilize.

8.3 Introduce 110 ml of titration solvent (5.5) into the vessel and pipette in a 5 ml aliquot of the sample solution (8.1) from the 50 ml volumetric flask. Switch on the stirrer (6.3) and adjust to a rapid stirring rate, but avoid any tendency for air bubbles to be drawn down into the solution.

8.4 Set the end-point potential. With each instrument, follow the manufacturer's instructions for end-point setting and to achieve the sensitivity in the platinum electrode circuit specified in 6.1.

8.5 Depending on the titrator apparatus, add the bromide/bromate solution (5.6) manually or by microprocessor control in small increments from the burette (6.5).

With commercial titrimeters, a sudden change in potential is indicated on the meter or recorder of the instrument as the end-point is approached. The end-point of the titration has been reached when the change in potential persists for 30 s.

8.6 Carry out a blank titration of each batch of titration solvent and reagents by repeating the entire procedure using 5 ml of 1,1,1-trichloroethane in place of the sample aliquot. If more than 0,1 ml of bromide/bromate solution is required to reach the end-point, disregard the analysis, prepare fresh titration solvent and fresh reagents and repeat the analysis.

9 Calculation

Calculate the bromine number, Br No., as follows:

$$\text{Br No.} = \frac{(V_1 - V_2) c_1 \times 15,98}{m}$$

where

V_1 is the volume of bromide/bromate solution required for titration of the test solution aliquot, in millimetres;

V_2 is the volume of bromide/bromate solution required for titration of the blank, in millimetres;

- c_1 is the concentration of the bromide/bromate solution, expressed as moles bromine per litre of solution;
- 15,98 is the factor for converting grams of bromine per 100 g of sample and incorporating molecular mass of bromine (as Br₂) and converting millimetres to litres;
- m is the mass of sample in the aliquot used, in grams.

10 Expression of results

Report the result, rounded to the nearest 0,1 for bromine numbers below 10,0, and to the nearest whole number for those above.

11 Precision

The precision of the method, as obtained by statistical examination of interlaboratory test results, is as follows.

11.1 Repeatability, r

The difference between two test results, obtained by the same operator with the same apparatus under constant operating conditions on identical test material would, in the normal and correct operation of the test method, exceed the following values only in one case in 20.

Petroleum distillates:

- a) 90 % (V/V) distillation recovery under 205 °C
 $r = 0,11 (X^{0,70})$
- b) 90 % (V/V) distillation recovery 205 °C to 327 °C
 $r = 0,11(X^{0,67})$

where X is the average value of the samples being tested.

Commercial olefins:

$$r = 3$$

11.2 Reproducibility, R

The difference between two single and independent results obtained by different operators working in different laboratories on nominally identical test material would, in the normal and correct operation of the test method, exceed the following value only in one case in 20.

Petroleum distillates:

- a) 90 % (V/V) distillation recovery under 205 °C
 $R = 0,72 (X^{0,70})$
- b) 90 % (V/V) distillation recovery 205 °C to 327 °C
 $R = 0,78 (X^{0,67})$

where X is the average value of the samples being tested.

Commercial olefins:

$$R = 12^*)$$

12 Test report

The test report shall contain at least the following information:

- a reference to this International Standard;
- the type and identification of the product tested;
- the result of the test (see clause 10);
- any deviation, by agreement or otherwise, from the procedure specified;
- the date of the test.

*) Provisional value obtained from a limited amount of data.

Annex A (informative)

Interpretation of bromine number and reported data

Technically, the bromine number is the number of grams of bromine that will react with 100 g of the test portion under specified conditions. By this definition, bromine consumed by addition, substitution, oxidation and reactions with sulfur-, nitrogen-, and oxygen-containing compounds is included in the bromine number of the material. The use of bromine number in the estimation of olefinic unsaturation rests on the fact that the addition reaction proceeds rapidly and completely under most conditions. The addition of bromine proceeds readily at temperatures down to or below 0 °C. Decreasing reaction temperature, time of contact, and concentration of free bromine tend to retard both substitution and oxidation reactions. Other factors, such as solvent medium, extent of agitation and exposure to actinic light, also influence the rate of the various reactions.

Experience has shown that no single set of test conditions will direct the reaction of bromine in one manner to the exclusion of the others. For this reason, the conditions for bromine number determinations are usually established on an empirical basis to give reasonable values with representative materials.

The possibility of multiple reactions occurring concurrently and the variable behaviour of certain materials in the presence of bromine impose an element of uncertainty in the interpretation of results. A knowledge of the material being handled and of its response to bromine greatly reduces the risk of misinterpretation.

Bromine number data have been obtained on a large number of petroleum hydrocarbons and certain nonhydrocarbons associated with petroleum using the electrometric procedure. These data, which were submitted by cooperators, are presented in table A.1.

This information is given in order to serve as a general guide in the interpretation of bromine numbers on petroleum products. It is recognised that the bromine number data recorded in this table are of limited value owing to incompleteness; however, it is considered that their usefulness will be amplified as more bromine number data obtained by the method specified in this International Standard are contributed by cooperators. Such additional data should be submitted to the ISO/TC 28 Secretariat, from whom further information regarding experimental conditions can also be obtained.

