

BRITISH STANDARD

BS 5752 :
Part 15 : 1997
ISO 11292 : 1995
(corrected and reprinted
1997)

Methods of test for

Coffee and coffee products

**Part 15. Instant coffee: determination of
free and total carbohydrate contents by
high performance anion-exchange
chromatography**

ICS 67.140.20

S

National foreword

This British Standard reproduces verbatim ISO 11292 : 1995 (corrected and reprinted in 1997) and implements it as the UK national standard. It supersedes BS 5752 : Part 15 : 1995 which is withdrawn. This new edition includes new informative references in annex B. The international standard has been amended to reflect the technical content of the 1995 British Standard.

This standard embodies an agreement, to which the UK was a party, reached in Subcommittee 15, Coffee, of Technical Committee 34, Agricultural food products, of ISO.

The UK participation in its preparation was entrusted to Technical Committee AW/15, Coffee, which has the responsibility to:

- aid enquirers to understand the text;
- present to the responsible international/European committee any enquiries on the interpretation, or proposals for change, and keep the UK interests informed;
- monitor related international and European developments and promulgate them in the UK.

A list of organizations represented on this committee can be obtained on request to its secretary.

Cross-references

The British Standards which implement international or European publications referred to in this document may be found in the BSI Standards Catalogue under the section entitled 'International Standards Correspondence Index', or using the 'Find' facility of the BSI Standards Electronic Catalogue.

Compliance with a British Standard does not of itself confer immunity from legal obligations.

Summary of pages

This document comprises a front cover, an inside front cover, the ISO title page, page ii, pages 1 to 16, an inside back cover and a back cover.

This British Standard, having been prepared under the direction of the Consumer Products and Services Sector Board, was published under the authority of the Standards Board and comes into effect on 15 May 1997

© BSI 1997

Amendments issued since publication

Amd. No.	Date	Text affected

INTERNATIONAL
STANDARD

ISO
11292

First edition
1995-06-15

Corrected and reprinted
1997-02-01

**Instant coffee — Determination of free
and total carbohydrate contents —
Method using high-performance
anion-exchange chromatography**

*Café soluble — Détermination des teneurs en hydrates de carbone libres
et totaux — Méthode par chromatographie d'échange d'anions à haute
performance*



Reference number
ISO 11292:1995(E)

*
UJ
*

BS 5752 : Part 15 : 1997**Foreword**

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

International Standard ISO 11292 was prepared by Technical Committee ISO/TC 34, *Agricultural food products*, Subcommittee SC 15, *Coffee*.

Annexes A and B of this International Standard are for information only.

Descriptors: agricultural products, plant products, coffee, chemical analysis, determination of content, carbohydrates, high performance liquid chromatography.

Instant coffee — Determination of free and total carbohydrate contents — Method using high-performance anion-exchange chromatography

1 Scope

This International Standard specifies a method for the determination of free and total carbohydrate contents in instant coffee using high-performance anion-exchange chromatography. In particular, it determines the content of individual monosaccharides, sucrose and mannitol.

2 Normative references

The following standards contain provisions which, through reference in this text, constitute provisions of this International Standard. At the time of publication, the editions indicated were valid. All standards are subject to revision, and parties to agreements based on this International Standard are encouraged to investigate the possibility of applying the most recent editions of the standards indicated below. Members of IEC and ISO maintain registers of currently valid International Standards.

ISO 1042:1983, *Laboratory glassware — One-mark volumetric flasks*.

ISO 3509:1989, *Coffee and its products — Vocabulary*.

ISO 3726:1983, *Instant coffee — Determination of loss in mass at 70 °C under reduced pressure*.

3 Definitions

For the purposes of this International Standard, the definitions given in ISO 3509 and the following definitions apply.

3.1 free carbohydrate content: Content of each individual monosaccharide (arabinose, fructose, galactose, glucose, mannose), and the sucrose and mannitol contents, determined under the conditions described (method A). Content is expressed as a percentage by mass on a dry basis.

3.2 total carbohydrate content: Content of each individual monosaccharide (arabinose, galactose, glucose, mannose, xylose) and the mannitol content, determined under the conditions described, which includes a strong hydrolysis step (method B). Content is expressed as a percentage by mass on a dry basis.

4 Principle

4.1 Method A

Dissolution of a test portion in water. Separation of the carbohydrates present in the filtered extract by ion chromatography on a high-performance anion-exchange column (HPAEC) using pure water as eluent. Electrochemical detection of the eluted compounds by means of a pulsed amperometric detector (PAD) and quantification by comparison with peak areas given by standard solutions.

BS 5752 : Part 15 : 1997**4.2 Method B**

Hydrolysis of a test portion with aqueous hydrochloric acid. Analysis of the carbohydrates present in the filtered hydrolysed solution as described in method A.

5 Reagents

Use only reagents of recognized analytical grade, unless otherwise specified, and distilled or demineralized water or water of equivalent purity.

5.1 Sodium hydroxide (NaOH), 50 % (m/m) aqueous solution.

The reagent should contain the minimum amount of sodium carbonate and mercury. Do not shake or stir the solution before use.

5.2 Hydrochloric acid (HCl), 1,00 mol/l standard volumetric solution.**5.3 Eluent 1 (S1), demineralized water (18 M Ω ·cm).**

Filter the demineralized water through 0,2 μ m membrane filters. Degas by sparging with helium for between 20 min and 30 min.

5.4 Eluent 2 (S2), sodium hydroxide (NaOH), 300 mmol/l solution.

To 985 ml of degassed water (5.3), pipette 15,6 ml of the sodium hydroxide solution (5.1).

CAUTION — It is extremely important to remove dissolved carbon dioxide from the eluents prior to use. Carbonate will act as a strong "pusher" on the column, resulting in a drastic reduction in resolution and efficiency. Prepare the solution the day before the analysis.

5.5 Carbohydrate standard solutions.

Prepare fresh solutions of arabinose, fructose, galactose, glucose, mannose, sucrose and mannitol.

Weigh, to the nearest 0,1 mg, approximately 100 mg of each carbohydrate into separate 100 ml volumetric flasks (6.2) and dilute to the mark with water (stock standard solutions of 1 000 mg/l).

Mixed standard solutions can also be prepared from separate stock solutions once the retention time of each carbohydrate is known under the prevailing chromatographic conditions.

Further dilute the standard solutions to reach carbohydrate concentrations similar to those found in the non-hydrolysed or hydrolysed instant coffee sample solutions.

The resolution of rhamnose from arabinose is difficult to achieve. If these two monosaccharides coelute, do not add rhamnose in a mixed standard solution.

6 Apparatus

Usual laboratory apparatus and, in particular, the following.

6.1 Analytical balance, capable of weighing to an accuracy of $\pm 0,1$ mg.**6.2 One-mark volumetric flasks**, of capacity 100 ml (in accordance with class A of ISO 1042).**6.3 Graduated cylinders**, of capacities 1 000 ml and 50 ml, tall form.**6.4 Vacuum filtration system.****6.5 Folded filter papers**, medium fast, qualitative.**6.6 Disposable C18 filter cartridges**¹⁾, to be used according to the manufacturer's recommendations.**6.7 Disposable membrane filters**, 0,2 μ m pore size.**6.8 Water bath**, capable of being maintained at 100 °C \pm 5 °C.

1) Sep-Pack C18 (Waters) and Supelclean LC-18 (Supelco) are examples of suitable products available commercially. This information is given for the convenience of users of this International Standard and does not constitute an endorsement by ISO of these products.

6.9 Metal-free liquid chromatograph²⁾, with a high-performance anion-exchange column³⁾ filled with pellicular polystyrene-divinylbenzene resin and pre-column (guard column)⁴⁾ and postcolumn delivery system.

6.10 Pulsed amperometric detector (PAD) with gold electrode⁵⁾.

6.11 Integrator chromatography data station⁶⁾.

6.12 Disposable cartridges⁷⁾, to be used according to the manufacturer's recommendations.

7 Sampling

It is important that the laboratory receive a sample which is truly representative and has not been damaged or changed during transport or storage.

Sampling is not part of the method specified in this International Standard. A recommended sampling method is given in ISO 6670:1983, *Instant coffee in cases with liners — Sampling*.

When the instant coffee is not in cases with liners, take a well-mixed representative sample from single packed units.

8 Procedure

8.1 Determination of dry matter

Calculate the dry matter determined on a portion of the laboratory sample in accordance with ISO 3726.

8.2 Preparation of sample for analysis — Method A

Weigh, to the nearest 0,1 mg, approximately 300 mg of the laboratory sample directly into a 100 ml volumetric flask (6.2). Add, using a graduated cylinder (6.3), 70 ml of water and shake until dissolution is

complete. Dilute to the mark with water. Filter 5 ml to 10 ml of this solution through a cartridge (6.6). Discard the first few millilitres.

8.3 Preparation of sample for analysis — Method B

Weigh, to the nearest 0,1 mg, approximately 300 mg of the laboratory sample directly into a 100 ml volumetric flask (6.2). Add 50 ml of the hydrochloric acid (5.2) and swirl. Place the flask in a boiling water bath (6.8) for 150 min.

Keep the level of the sample solution always below that of the water in the bath. Swirl the solution by hand every 30 min. Cool to room temperature by passing the flask under tap water. Dilute to the mark with water and filter the solution through a folded filter paper (6.5). Pass 3 ml of the filtrate through a disposable cartridge (6.12). Discard the first millilitre.

8.4 Chromatographic analysis

Set up the chromatograph (6.9), detector (6.10) and integrator (6.11).

Allow the chromatograph to equilibrate.

Filter the standard solutions (5.5) and the test solutions (8.2 or 8.3) through 0,2 µm membrane filters (6.7).

Inject the same volume of filtered standard and test solutions into the chromatograph and separate carbohydrates under the conditions given in tables 1 and 2.

Identify and quantify carbohydrates in the sample solution by comparison with retention times and areas of corresponding peaks obtained using the standard solution.

Inject a standard solution every four injections, in order to account for any changes in retention times or peak integrations.

2) The BioLC system (Dionex) consisting of a model GPM-II quaternary gradient pump (Dionex) with a model SP8875 autosampler (Spectra - Physics) filled with a 20 µl loop, model EDM-II eluent degas module (Dionex) and reagent reservoir for NaOH postcolumn addition (Dionex) are examples of suitable equipment available commercially.

3) CarboPac PA1 (10 µm, 250 mm × 4 mm) (Dionex) is an example of a suitable analytical column available commercially.

4) CarboPac PA (Dionex) is an example of a suitable precolumn available commercially.

5) Model PAD-II (Dionex) is an example of suitable equipment available commercially.

6) Model Autolon AI-450 is an example of suitable equipment available commercially.

7) On Guard-AG (Dionex) is an example of a suitable cartridge commercially available.

This information is given for the convenience of users of this International Standard and does not constitute an endorsement by ISO of these products.

Table 1 — Preparation of column

Eluent	Time min	Eluent S1 ml	Eluent S2 ml	Procedure
Isocratic	0	100	0	Start data acquisition
	50,0	100	0	Stop data acquisition
	50,1	0	100	Start clean-up
	65,0	0	100	Stop clean-up
	65,1	100	0	Start re-equilibrium
	80,0	100	0	Stop re-equilibrium

NOTES

1) Retention times tend to vary from one column to another. Start clean-up only when the last monosaccharide (ribose) has been eluted.

2) It may be necessary to perform two or three injections of standard solution or to increase the re-equilibrium time in order to achieve a good separation of sucrose and xylose.

Table 2 — Conditions for analysis

Injection	20 µl
Flowrate	1,0 ml/min
Postcolumn addition	Eluent S2 (5.4) at a flowrate of 0,8 ml/min
Temperature	Ambient
Detector	Fill up the reference cell with Eluent S2 (5.4). Use the optimum conditions given by the manufacturer.

9 Calculation

The carbohydrate content, ω , expressed as a percentage by mass, is equal to

$$\omega = \frac{A \cdot m_0 \cdot V}{A_0 \cdot m \cdot V_0} \times 100$$

where

- A is the peak area of the individual carbohydrate in the test solution (8.4);
- A₀ is the peak area of the individual carbohydrate in the standard solution (8.4);

m is the mass, in grams, of the test portion in the test solution (8.2 or 8.3), expressed on a dry basis;

m_0 is the mass, in grams, of the carbohydrate in the standard solution (5.5);

V is the volume, in millilitres, of the test solution (8.2 or 8.3);

V_0 is the volume, in millilitres, of the standard solution (5.5).

Take as the result the arithmetic mean of the two determinations. Express the result either as free (method A) or total (method B) carbohydrate content to the nearest 0.01 % (m/m) for each carbohydrate of interest, or the total of all carbohydrates detected.

10 Precision

Details of an interlaboratory test on the precision of the method are summarized in annex A. The values derived from this interlaboratory test may not be applicable to concentration ranges and matrices other than those given.

The values of repeatability and reproducibility for each carbohydrate show a marked relationship with the wide percentage concentration of each carbohydrate, which in many cases in the samples examined was very low. This is reflected especially in some poor reproducibility figures, as might be expected. However, with contents at higher concentrations, both precision values are acceptable, as is also indicated in annex A.

10.1 Repeatability

Repeatability is defined as the absolute difference between two independent single test results obtained on identical test material in the same laboratory by the same operator using the same equipment within a short interval of time.

It is not possible therefore to give an exact repeatability figure for each and every carbohydrate, both free and total, for a full range of possible percentage contents.

Provided, however, that the content of a given carbohydrate is greater than 0,3 % (*m/m*), the data show that the relative standard deviation averages 4,5 %.

10.2 Reproducibility

Reproducibility is defined as the absolute difference between two single test results, obtained using the same method on identical test material in different

laboratories with different operators using different equipment.

Again it is not possible to give an exact reproducibility figure for each and every carbohydrate, both free and total, for a full range of possible carbohydrate contents.

Provided, however, that the content of a given carbohydrate is greater than 0,3 % (*m/m*), the data show that the relative standard deviation averages 14,3 % (except for fructose data). At a content of less than 0,3 % (*m/m*), the coefficient of variation increases very sharply.

For each carbohydrate, the reproducibility and repeatability ranges as found in the interlaboratory test and the applied mean content range are given in table 3.

11 Test report

The test report shall specify

- the method in accordance with which sampling was carried out, if known,
- the method used,
- the result obtained for each carbohydrate of interest,
- if the repeatability has been checked, the final quoted result obtained.

It shall also mention all operating details not specified in this International Standard, or regarded as optional, together with details of any incidents which may have influenced the test result.

The test report shall include all information necessary for the complete identification of the sample.



*
05
*

Table 3 — Coefficient of variation of reproducibility and repeatability

Carbohydrate	Mean content range applied in test % (m/m) d.b.	Range of coefficient of variation of reproducibility, %	Range of coefficient of variation of repeatability, %
Free mannitol	0,02 to 1,6	59,5 to 9,9	0,9 to 8,2
Free arabinose	0,46 to 1,3	13,8 to 5,1	1,6 to 7,3
Total arabinose	3,5 to 4,8	21,1 to 4,9	6,6 to 3,0
Free galactose	0,19 to 0,56	13,0 to 4,1	9,8 to 3,0
Total galactose	8,1 to 18,5	7,5 to 12,9	8,1 to 1,7
Free glucose	0,04 to 2,0	23,8 to 6,1	10,2 to 2,5
Total glucose	0,68 to 16,6	12,5 to 24,3	8,7 to 3,8
Free mannose	0,16 to 1,0	40,0 to 16,9	8,2 to 3,8
Total mannose	2,6 to 19,1	10,6 to 21,7	2,0 to 5,8
Total xylose	0,1 to 1,9	37,7 to 20,2	22,9 to 3,7
Free fructose	0,05 to 3,6	45,2 to 15,5	21,0 to 0,2
Sucrose	0,15 to 1,3	41,6 to 10,0	15,1 to 1,8

Annex A (informative)

Results of an interlaboratory test

An interlaboratory test, carried out at the international level in 1991 under the auspices of ISO/TC 34, *Agricultural food products*, SC 15, *Coffee*, and executed by the technical committee of AFCASOLE, in which 11 laboratories participated, each of which carried out 2 determinations on each sample of 6 different commercial coffees, gave the statistical results (evaluated in accordance with ISO 5725⁸⁾) shown in tables A.1 to A.8 .



8) ISO 5725:1986, *Precision of test methods — Determination of repeatability and reproducibility for a standard test method by interlaboratory tests.*

Table A.1 — Determination of mannitol content for instant coffees

Sample	1		2		3		4		5		6	
	7	11	10	11	10	10	11	11	9	11	8	10
Number of laboratories retained after eliminating outliers	7	11	10	11	10	10	11	11	9	11	8	10
Mannitol content	Free	Total	Free	Total	Free	Total	Free	Total	Free	Total	Free	Total
Mean content, % (m/m)	0,024	0,179	0,060	0,151	1,582	1,854	0,619	0,782	0,192	0,300	0,059	0,179
Standard deviation of repeatability, s_r	0,000	0,013	0,001	0,011	0,044	0,04	0,022	0,036	0,016	0,032	0,004	0,018
Repeatability, 2,83 s_r	0,001	0,035	0,002	0,032	0,124	0,115	0,063	0,102	0,045	0,091	0,012	0,050
Coefficient of variation of repeatability, %	1,547	6,986	0,914	7,609	2,777	2,198	3,594	4,594	8,245	10,776	7,071	9,801
Standard deviation of reproducibility, s_R	0,015	0,075	0,034	0,069	0,157	0,331	0,150	0,161	0,065	0,112	0,029	0,090
Reproducibility, 2,83 s_R	0,041	0,213	0,098	0,196	0,444	0,938	0,424	0,454	0,183	0,316	0,083	0,254
Coefficient of variation of reproducibility, %	59,454	41,996	57,581	45,961	9,925	17,866	24,162	20,535	33,789	37,260	49,132	50,084

Table A.2 — Determination of arabinose content for instant coffees

Sample	1		2		3		4		5		6	
	Free	Total	Free	Total	Free	Total	Free	Total	Free	Total	Free	Total
Number of laboratories retained after eliminating outliers	11	11	9	11	11	11	11	11	10	9	11	11
Arabinose content	Free	Total	Free	Total	Free	Total	Free	Total	Free	Total	Free	Total
Mean content, % (m/m)	0,891	3,544	1,325	4,833	0,464	4,759	0,747	4,539	0,505	4,081	0,629	3,786
Standard deviation of repeatability, s_r	0,033	0,234	0,021	0,160	0,017	0,147	0,054	0,209	0,022	0,122	0,026	0,216
Repeatability, 2,83 s_r	0,092	0,662	0,060	0,453	0,049	0,417	0,154	0,590	0,063	0,344	0,073	0,612
Coefficient of variation of repeatability, %	3,664	6,597	1,603	3,311	3,755	3,098	7,237	4,594	4,419	2,979	4,092	5,709
Standard deviation of reproducibility, s_R	0,123	0,749	0,068	0,636	0,049	0,598	0,087	0,831	0,039	0,199	0,056	0,771
Reproducibility, 2,83 s_R	0,349	2,120	0,191	2,365	0,139	1,691	0,247	2,353	0,110	0,562	0,158	2,182
Coefficient of variation of reproducibility, %	13,832	21,141	5,105	17,289	10,556	12,559	11,707	18,319	7,729	4,868	8,895	20,366



Table A.3 — Determination of galactose content for instant coffees

Sample	1		2		3		4		5		6	
	Free	Total	Free	Total	Free	Total	Free	Total	Free	Total	Free	Total
Number of laboratories retained after eliminating outliers	11	10	11	10	11	11	11	11	9	9	11	10
Galactose content	Free	Total	Free	Total	Free	Total	Free	Total	Free	Total	Free	Total
Mean content, % (m/m)	0,562	17,779	0,339	18,481	0,191	8,085	0,438	16,259	0,475	18,444	0,362	17,687
Standard deviation of repeatability, s_r	0,017	1,441	0,014	0,430	0,019	0,219	0,026	0,632	0,015	0,306	0,018	0,780
Repeatability, 2,83 s_r	0,047	4,078	0,039	1,216	0,053	0,619	0,074	1,789	0,041	0,867	0,051	2,152
Coefficient of variation of repeatability, %	2,964	8,105	4,070	2,325	3,810	2,704	5,948	3,836	3,085	1,661	5,006	4,300
Standard deviation of reproducibility, s_R	0,030	1,582	0,027	2,181	0,025	0,647	0,036	2,097	0,019	1,377	0,045	1,503
Reproducibility, 2,83 s_R	0,084	4,478	0,077	6,171	0,070	1,831	0,103	5,935	0,055	3,898	0,127	4,255
Coefficient of variation of reproducibility, %	5,261	8,900	8,048	11,800	13,032	8,003	8,317	12,891	4,075	7,468	12,417	8,500

Table A.4 — Determination of glucose content for instant coffees

Sample	1		2		3		4		5		6	
	Free	Total	Free	Total	Free	Total	Free	Total	Free	Total	Free	Total
Number of laboratories retained after eliminating outliers	11	11	9	10	10	11	10	11	10	11	11	11
Glucose content	Free	Total	Free	Total	Free	Total	Free	Total	Free	Total	Free	Total
Mean content, % (m/m)	0,105	0,684	0,042	0,326	2,041	16,649	1,857	4,385	0,186	1,951	0,186	1,025
Standard deviation of repeatability, s_r	0,010	0,059	0,004	0,062	0,050	0,990	0,068	0,166	0,019	0,111	0,018	0,081
Repeatability, $2,83 s_r$	0,029	0,168	0,012	0,174	0,143	2,803	0,192	0,469	0,053	0,313	0,052	0,230
Coefficient of variation of repeatability, %	9,911	8,661	10,151	7,449	2,470	5,949	4,092	3,780	10,006	5,674	9,853	7,937
Standard deviation of reproducibility, s_R	0,022	0,114	0,009	0,178	0,128	4,053	0,101	1,038	0,040	0,245	0,044	0,140
Reproducibility, $2,83 s_R$	0,062	0,323	0,024	0,504	0,361	11,471	0,286	2,938	0,113	0,693	0,125	0,397
Coefficient of variation of reproducibility, %	21,057	16,667	20,419	21,551	6,249	24,345	6,091	23,672	21,370	12,541	23,808	13,702



Table A.5 — Determination of mannose content for instant coffees

Sample	1		2		3		4		5		6	
	Free	Total	Free	Total	Free	Total	Free	Total	Free	Total	Free	Total
Number of laboratories retained after eliminating outliers	11	10	10	11	11	11	11	11	10	11	11	10
Mannose content	Free	Total	Free	Total	Free	Total	Free	Total	Free	Total	Free	Total
Mean content, % (m/m)	0,583	17,913	0,155	14,365	0,470	2,601	0,329	5,598	0,277	7,653	0,991	19,067
Standard deviation of repeatability, s_r	0,028	1,038	0,013	0,380	0,020	0,052	0,023	0,171	0,012	0,212	0,038	0,426
Repeatability, 2,83 s_r	0,080	2,938	0,036	1,075	0,056	0,147	0,065	0,483	0,033	0,600	0,107	1,206
Coefficient of variation of repeatability, %	4,875	5,796	8,155	2,645	4,176	2,003	6,960	3,047	4,166	2,772	3,801	2,234
Standard deviation of reproducibility, s_R	0,142	2,029	0,055	2,180	0,082	0,352	0,061	0,859	0,111	0,808	0,168	4,137
Reproducibility, 2,83 s_R	0,402	5,741	0,153	6,168	0,233	0,997	0,174	2,431	0,313	2,288	0,474	11,708
Coefficient of variation of reproducibility, %	24,373	11,324	36,065	15,173	17,514	13,546	18,647	15,346	39,969	10,563	16,918	21,698



Table A.6 — Determination of fructose content for instant coffees

Sample	1		2		3		4		5		6	
	Free	Total	Free	Total	Free	Total	Free	Total	Free	Total	Free	Total
Number of laboratories retained after eliminating outliers	9	6	8	—	9	9	10	7	10	7	8	6
Fructose content	Free	Total	Free	Total	Free	Total	Free	Total	Free	Total	Free	Total
Mean content, % (m/m)	0,171	0,189	0,054	—	3,622	2,010	3,124	1,368	0,282	0,244	0,480	0,363
Standard deviation of repeatability, s_r	0,029	0,046	0,011	—	0,105	0,114	0,091	0,076	0,026	0,049	0,024	0,027
Repeatability, 2,83 s_r	0,082	0,130	0,032	—	0,297	0,323	0,258	0,214	0,072	0,140	0,067	0,075
Coefficient of variation of repeatability, %	16,882	24,416	21,035	—	2,900	5,683	0,219	5,527	9,045	20,261	5,167	7,310
Standard deviation of reproducibility, s_R	0,054	0,129	0,018	—	0,667	1,440	0,569	0,958	0,128	0,144	0,072	0,247
Reproducibility, 2,83 s_R	0,152	0,366	0,052	—	1,886	4,076	1,611	2,713	0,361	0,406	0,203	0,700
Coefficient of variation of reproducibility, %	31,370	68,482	33,846	—	18,403	71,660	13,217	70,068	45,292	58,874	15,549	68,236

Table A.7 — Determination of xylose content for instant coffees

Sample	1		2		3		4		5		6	
	Free	Total	Free	Total	Free	Total	Free	Total	Free	Total	Free	Total
Number of laboratories retained after eliminating outliers	—	8	—	9	—	11	—	11	7	11	—	9
Xylose content	Free	Total	Free	Total	Free	Total	Free	Total	Free	Total	Free	Total
Mean content, % (m/m)	—	0,097	—	0,146	—	1,956	—	0,736	0,029	1,826	—	0,133
Standard deviation of repeatability, s_r	—	0,022	—	0,014	—	0,085	—	0,027	0,007	0,134	—	0,019
Repeatability, 2,83 s_r	—	0,063	—	0,040	—	0,239	—	0,077	0,020	0,380	—	0,053
Coefficient of variation of repeatability, %	—	22,929	—	9,750	—	4,555	—	3,717	24,731	7,356	—	14,175
Standard deviation of reproducibility, s_R	—	0,037	—	0,030	—	0,424	—	0,204	0,008	0,410	—	0,032
Reproducibility, 2,83 s_R	—	0,103	—	0,084	—	1,200	—	0,578	0,023	1,161	—	0,090
Coefficient of variation of reproducibility, %	—	37,730	—	20,226	—	22,844	—	27,759	28,215	22,468	—	24,006

Table A.8 — Determination of sucrose content for instant coffees

Sample	1	2	3	4	5	6
Number of laboratories retained after eliminating outliers		10	10	10	11	9
Mean sucrose content, % (m/m)	—	0,149	1,319	0,746	0,181	0,158
Standard deviation of repeatability, s_r	—	0,007	0,023	0,051	0,027	0,005
Repeatability, $2,83 s_r$	—	0,020	0,066	0,144	0,077	0,015
Coefficient of variation of repeatability, %	—	4,813	1,773	6,807	15,081	3,369
Standard deviation of reproducibility, s_R	—	0,057	0,132	0,087	0,075	0,052
Reproducibility, $2,83 s_R$	—	0,162	0,372	0,246	0,213	0,147
Coefficient of variation of reproducibility, %	—	38,395	9,972	11,632	41,584	33,074
NOTE — Some information on fructose and ribose contents was available, but is not reported here.						



Annex B

(informative)

Bibliography

[1] *AFCASOLE Statement*, July 1995; available from AFCASOLE, 18 rue de la Pépinière, 75008 Paris.

NOTE 1 This statement is a statistical interpretation of an available database which establishes limits for relevant carbohydrates within acceptable confidence levels. ISO 11292 provides an analytical tool to establish carbohydrate levels in soluble coffee with the intention of identifying potential adulterants.

[2] *UK Code of Practice for the Soluble Coffee Industry in the UK*, June 1995; available either from the British Soluble Coffee Packers and Importers Association, Ltd., Suite 13, Castle House, Castlereagh Street, London W18 5YR, or the British Soluble Coffee Manufacturer's Association, 8 Catherine Street, London WC2 B5JJ.

NOTE 2 This is a working interpretation of the AFCASOLE statement which reflects limits for carbohydrates with a level of significance that would provide a basis for legal enforcement by the authorities.

BS 5752 :
Part 15 : 1997
ISO 11292 : 1995

BSI — British Standards Institution

BSI is the independent national body responsible for preparing British Standards. It presents the UK view on standards in Europe and at the international level. It is incorporated by Royal Charter.

Contract requirements

A British Standard does not purport to include all the necessary provisions of a contract. Users of British Standards are responsible for their correct application.

Revisions

British Standards are updated by amendment or revision. Users of British Standards should make sure that they possess the latest amendments or editions.

It is the constant aim of BSI to improve the quality of our products and services. We would be grateful if anyone finding an inaccuracy or ambiguity while using this British Standard would inform the Secretary of the responsible technical committee, the identity of which can be found on the inside front cover. Tel: 0181 996 9000; Fax: 0181 996 7400.

BSI offers members an individual updating service called PLUS which ensures that subscribers automatically receive the latest editions of standards.

Buying standards

Orders for all BSI, international and foreign standards publications should be addressed to Customer Services, Sales Department at Chiswick: Tel: 0181 996 7000; Fax: 0181 996 7001.

In response to orders for international standards, it is BSI policy to supply the BSI implementation of those that have been published as British Standards, unless otherwise requested.

Information on standards

BSI provides a wide range of information on national, European and international standards through its Library, the Standardline Database, the BSI Information Technology Service (BITS) and its Technical Help to Exporters Service. Contact the Information Department at Chiswick: Tel: 0181 996 7111; Fax: 0181 996 7048.

Subscribing members of BSI are kept up to date with standards developments and receive substantial discounts on the purchase price of standards. For details of these and other benefits contact Customer Services, Membership at Chiswick: Tel: 0181 996 7002; Fax: 0181 996 7001.

Copyright

Copyright subsists in all BSI publications. BSI also holds the copyright, in the UK, of the publications of the international standardization bodies. Except as permitted under the Copyright, Designs and Patents Act 1988 no extract may be reproduced, stored in a retrieval system or transmitted in any form or by any means – electronic, photocopying, recording or otherwise – without prior written permission from BSI.

This does not preclude the free use, in the course of implementing the standard, of necessary details such as symbols, and size, type or grade designations. If these details are to be used for any other purpose than implementation then the prior written permission of BSI must be obtained.

If permission is granted, the terms may include royalty payments or a licensing agreement. Details and advice can be obtained from the Copyright Manager, BSI, 389 Chiswick High Road, London W4 4AL.

BSI
389 Chiswick High Road
London
W4 4AL