

Methods of

Sampling and test for sodium carbonate for industrial use —

Part 7: Determination of copper content

NOTE It is recommended that this Part be read in conjunction with the information in the “*General introduction*” published separately as BS 6070-0.

UDC 661.833.622:546.33'264:543.42.062:546.56

Confirmed
January 2011

A British Standard does not purport to include all the necessary provisions of a contract. Users of British Standards are responsible for their correct application.

Compliance with a British Standard does not of itself confer immunity from legal obligations.

Summary of pages

This document comprises a front cover, an inside front cover, pages i and ii, pages 1 and 2, an inside back cover and a back cover.

This standard has been updated (see copyright date) and may have had amendments incorporated. This will be indicated in the amendment table on the inside front cover.

This British Standard, having been prepared under the direction of the Chemicals Standards Committee, was published under the authority of the Executive Board and comes into effect on 27 February 1981

© BSI 12-1999

The following BSI reference relates to the work on this standard:
Committee reference CIC/22

ISBN 0 580 12006 6

Amendments issued since publication

Amd. No.	Date of issue	Comments

Contents

	Page	
1	Scope	1
2	Reference	1
3	Principle	1
4	Reagents	1
5	Apparatus	1
6	Procedure	1
7	Expression of results	2
Table 1 — Volumes of standard copper solution		1
Publication referred to		Inside back cover

1 Scope

This Part of BS 6070 describes a method for the determination of the copper content of sodium carbonate for industrial use. The method is applicable to products having copper contents, expressed as Cu, in the range 0.5 mg/kg to 10 mg/kg.

There is no International Standard corresponding to this Part of BS 6070.

2 Reference

The title of the publication referred to in this standard is given on the inside back cover.

3 Principle

The copper present is reduced with ascorbic acid and a violet coloured complex is formed by addition of 2,2'-biquinolyl. This complex is extracted with amyl alcohol and its colour measured spectrophotometrically.

4 Reagents

The following reagents are required and shall be of recognized analytical quality. Water complying with the requirements of BS 3978 shall be used throughout.

4.1 Sodium sulphate, anhydrous.

4.2 Hydrochloric acid solution, ρ approximately 1.18 g/ml, approximately 36 % m/m solution or approximately 11 N.

4.3 Amyl alcohol

4.4 (+)-Tartaric acid, 500 g/l solution.

4.5 Sodium hydroxide, 200 g/l solution.

4.6 L-Ascorbic acid, 100 g/l solution, freshly prepared.

4.7 2,2'-Biquinolyl, 0.5 g/l solution. Dissolve 0.25 g of 2,2'-biquinolyl in the amyl alcohol (4.3) and dilute with more of the amyl alcohol to 500 ml.

4.8 Bromine water, saturated solution.

4.9 Copper, standard solution corresponding to 0.1 g of copper per litre. Dissolve 0.3928 g of copper (II) sulphate pentahydrate ($\text{CuSO}_4 \cdot 5\text{H}_2\text{O}$) in water in a 1 000 ml one-mark volumetric flask, add 25 ml of approximately 6 N sulphuric acid solution, dilute to the mark and mix.

1 ml of this standard solution contains 0.100 mg of Cu.

4.10 Copper, standard solution corresponding to 0.01 g of copper per litre. Dilute 10.0 ml of the standard copper solution (4.9) to the mark in a 100 ml one-mark volumetric flask and mix.

1 ml of this standard solution contains 10 μg of Cu.

4.11 Narrow range indicator papers, covering the range pH 5.5 to pH 7.0.

4.12 Methyl orange indicator, 0.5 g/l aqueous solution.

5 Apparatus

Ordinary laboratory apparatus and the following are required.

5.1 Spectrophotometer, or

5.2 Photoelectric absorptiometer, fitted with filters providing maximum transmission at a wavelength of about 545 nm.

5.3 Optical cells, 4 cm optical path length.

6 Procedure

6.1 Test portion. Weigh, to the nearest 0.1 g, about 10 g of the sample.

6.2 Blank test. Carry out a blank test at the same time as the determination, following the same procedure and using the same quantities of reagents as specified in 6.3.2 but omitting the test portion.

6.3 Preparation of the calibration graph

6.3.1 Preparation of standard matching solutions

Into each of a series of six 500 ml stoppered separating funnels introduce the volumes of the standard copper solution (4.10) shown in Table 1.

Table 1 — Volumes of standard copper solution

Standard copper solution (4.10)	Corresponding mass of copper (Cu)
ml	g
0 ^a	0
2.0	20
4.0	40
6.0	60
8.0	80
10.0	100
^a Compensation solution	

6.3.2 Colour development. Treat the contents of each funnel (6.3.1) as follows.

Dilute with water to approximately 400 ml, then add 2 ml of the tartaric acid solution (4.4).

Adjust the pH of the solution to about 6.0 by addition of the sodium hydroxide solution (4.5) using the narrow range indicator paper (4.11) externally. Add 2 ml of the ascorbic acid solution (4.6), shake so as to mix thoroughly and allow to stand for 5 min. Add 10 ml of the 2,2'-biquinolyl solution (4.7) and shake well for about 2 min. Extract the copper complex with two 20 ml portions of the amyl alcohol (4.3), and transfer the extracts to a 100 ml beaker. Add about 2 g of the anhydrous sodium sulphate (4.1) to the combined extracts and stir thoroughly to remove traces of water. Filter the dried extract into a 50 ml one-mark volumetric flask, wash the residual sodium sulphate twice with 2 ml portions of the amyl alcohol (4.3). Transfer the washings to the flask, dilute to the mark with the amyl alcohol (4.3) and mix.

6.3.3 Photometric measurements. Carry out the photometric measurements either with the spectrophotometer (5.1) at the wavelength of maximum absorption (about 545 nm), or with the photoelectric absorptiometer (5.2), fitted with suitable filters, after having adjusted the instrument to zero absorbance against the amyl alcohol (4.3).

6.3.4 Plotting the calibration graph. Deduct the absorbance of the compensation solution (6.3.1) from those of the standard matching solutions (6.3.1). Plot a graph having, for example, the copper content, expressed in micrograms per 50 ml of standard matching solution, as abscissae and the corresponding values of absorbance as ordinates.

6.4 Determination

6.4.1 Preparation of the test solution. Transfer the test portion (6.1) quantitatively to a 400 ml beaker. Add about 100 ml of water and one drop of the methyl orange indicator solution (4.12). Neutralize the solution with the hydrochloric acid solution (4.2), then add 5 ml in excess. Add 10 ml of the bromine water (4.8) and boil the solution until free from bromine and allow to cool. Transfer the contents of the beaker quantitatively to a 500 ml separating funnel fitted with a stopper and add 1 ml of the hydrochloric acid solution (4.2).

6.4.2 Colour development. Treat the test solution in the separating funnel (6.4.1) as specified in 6.3.2.

6.4.3 Photometric measurement. Carry out the photometric measurement on the test solution (6.4.2) and on the blank test solution (6.2) following the procedure described in 6.3.3, after having adjusted the instrument to zero absorbance against the amyl alcohol (4.3).

NOTE If the absorbance exceeds the maximum of the calibration graph, repeat the determination using a smaller amount of the test portion (6.1) and modifying the calculation accordingly.

7 Expression of results

By means of the calibration graph (6.3.4), determine the quantity of copper (Cu) corresponding to the value of the photometric measurement.

The copper content, expressed as milligrams of copper (Cu) per kilogram, is given by the following formula

$$\frac{m_1 - m_2}{1000} \times \frac{1000}{m_o} = \frac{(m_1 - m_2)}{m_o}$$

where

m_0 is the mass of the test portion (in g);

m_1 is the mass of copper in the test solution (in μg);

m_2 is the mass of copper in the blank test solution (in μg).

Publication referred to

BS 3978, *Water for laboratory use.*

BSI — British Standards Institution

BSI is the independent national body responsible for preparing British Standards. It presents the UK view on standards in Europe and at the international level. It is incorporated by Royal Charter.

Revisions

British Standards are updated by amendment or revision. Users of British Standards should make sure that they possess the latest amendments or editions.

It is the constant aim of BSI to improve the quality of our products and services. We would be grateful if anyone finding an inaccuracy or ambiguity while using this British Standard would inform the Secretary of the technical committee responsible, the identity of which can be found on the inside front cover. Tel: 020 8996 9000. Fax: 020 8996 7400.

BSI offers members an individual updating service called PLUS which ensures that subscribers automatically receive the latest editions of standards.

Buying standards

Orders for all BSI, international and foreign standards publications should be addressed to Customer Services. Tel: 020 8996 9001. Fax: 020 8996 7001.

In response to orders for international standards, it is BSI policy to supply the BSI implementation of those that have been published as British Standards, unless otherwise requested.

Information on standards

BSI provides a wide range of information on national, European and international standards through its Library and its Technical Help to Exporters Service. Various BSI electronic information services are also available which give details on all its products and services. Contact the Information Centre. Tel: 020 8996 7111. Fax: 020 8996 7048.

Subscribing members of BSI are kept up to date with standards developments and receive substantial discounts on the purchase price of standards. For details of these and other benefits contact Membership Administration. Tel: 020 8996 7002. Fax: 020 8996 7001.

Copyright

Copyright subsists in all BSI publications. BSI also holds the copyright, in the UK, of the publications of the international standardization bodies. Except as permitted under the Copyright, Designs and Patents Act 1988 no extract may be reproduced, stored in a retrieval system or transmitted in any form or by any means – electronic, photocopying, recording or otherwise – without prior written permission from BSI.

This does not preclude the free use, in the course of implementing the standard, of necessary details such as symbols, and size, type or grade designations. If these details are to be used for any other purpose than implementation then the prior written permission of BSI must be obtained.

If permission is granted, the terms may include royalty payments or a licensing agreement. Details and advice can be obtained from the Copyright Manager. Tel: 020 8996 7070.