

Methods of

Sampling and test for sodium hydroxide for industrial use —

Part 9: Determination of calcium and magnesium contents

[ISO title: Sodium hydroxide for industrial use — Determination of calcium and magnesium contents — Flame atomic absorption method]

NOTE It is recommended that this Part be read in conjunction with the information in the “General introduction” published separately as BS 6075-0.

UDC 661.322.1:546.33 – 36:543.42.062:[546.41 + 546.46]

This British Standard, having been prepared under the direction of the Chemicals Standards Committee, was published under the authority of the Executive Board and comes into effect on 27 February 1981

© BSI 10-1999

The following BSI references relate to the work on this standard:
Committee reference CIC/22
Draft for comment 75/50812 DC

ISBN 0 580 12059 7

Amendments issued since publication

Amd. No.	Date of issue	Comments
3851	December 1981	Indicated by a sideline in the margin

Contents

	Page
National foreword	ii
<hr/>	
1 Scope and field of application	1
2 Reference	1
3 Principle	1
4 Reagents	1
5 Apparatus	2
6 Procedure	2
7 Expression of results	3
8 Test report	3
<hr/>	
Annex ISO publication relating to sodium hydroxide for industrial use	4
<hr/>	
Publications referred to	Inside back cover
<hr/>	

National foreword

This Part of BS 6075 is identical with ISO 3697 “*Sodium hydroxide for industrial use — Determination of calcium and magnesium contents — Flame atomic absorption method*”, published in 1976 by the International Organization for Standardization (ISO).

Terminology and conventions. The text of the International Standard has been approved as suitable for publication as a British Standard without deviation. Some terminology and certain conventions are not identical with those used in British Standards; attention is especially drawn to the following.

The comma has been used throughout as a decimal marker. In British Standards it is current practice to use a full point on the baseline as the decimal marker.

Wherever the words “International Standard” appear, referring to this standard, they should be read as “British Standard”.

Cross-reference

International Standard	Corresponding British Standard
	BS 6075 <i>Methods of sampling and test for sodium hydroxide for industrial use</i>
ISO 3195:1975	Part 5:1981 <i>Sampling and preparation of main test solution</i> (Identical)

The standards listed in the Annex are intended for information only. Corresponding British Standards are listed in BS 6075-0.

Additional information

Hydrochloric acid, ρ approximately 1.18 g/ml, about 36 % (*m/m*) solution, which is the corresponding reagent normally obtainable in the UK, is suitable for use in place of the solution specified in 4.1.

WARNING. Acetylene is highly flammable and, in some circumstances, can be explosive. Instrument manufacturers’ recommendations concerning the use of gaseous mixtures containing acetylene (see 5.1) should therefore be closely followed. Attention is also drawn to the need for compliance with the requirements of the Compressed Acetylene Order, 1947 (SR and O 1947: No. 805), relating to permissible pressures for the use of acetylene.

A British Standard does not purport to include all the necessary provisions of a contract. Users of British Standards are responsible for their correct application.

Compliance with a British Standard does not of itself confer immunity from legal obligations.

Summary of pages

This document comprises a front cover, an inside front cover, pages i and ii, pages 1 to 4, an inside back cover and a back cover.

This standard has been updated (see copyright date) and may have had amendments incorporated. This will be indicated in the amendment table on the inside front cover.

1 Scope and field of application

This International Standard specifies a flame atomic absorption method for the determination of the calcium and magnesium contents of sodium hydroxide for industrial use.

The method is applicable to products having Ca and Mg contents greater than

- 2,5 mg/kg and 0,5 mg/kg respectively, if an acetylene-air flame is used;
- 0,6 mg/kg and 1,0 mg/kg respectively, if an acetylene-dinitrogen monoxide flame is used.

2 Reference

ISO 3195, *Sodium hydroxide for industrial use — Sampling — Test sample — Preparation of the main solution for carrying out certain determinations.*

3 Principle

Acidification of a test portion with hydrochloric acid.

Aspiration of the solution into an acetylene-dinitrogen monoxide or acetylene-air flame, after addition of lanthanum ions in the latter case in order to suppress certain interferences.

Measurement of the absorption of the 422,7 nm and 285,2 nm lines emitted by calcium and magnesium hollow-cathode lamps.

4 Reagents

During the analysis, use only reagents of recognized analytical grade and only water doubly distilled in borosilicate glass apparatus with ground joints, or water of equivalent purity.

4.1 Hydrochloric acid, ρ approximately 1,19 g/ml, about 38 % (m/m) solution, or approximately 12 N.

4.2 Lanthanum chloride solution, corresponding to 5 g of lanthanum per litre.

NOTE This solution is not required when an acetylene-dinitrogen monoxide flame is employed.

Prepare either of the following solutions:

4.2.1 Dissolve 5,9 g of lanthanum oxide (La_2O_3) in 15 ml of water and 15 ml of the hydrochloric acid solution (4.1). Dilute to the mark with water in a 1 000 ml one-mark volumetric flask and mix.

4.2.2 Dissolve 13,4 g of lanthanum chloride heptahydrate ($\text{LaCl}_3 \cdot 7\text{H}_2\text{O}$) in water, dilute to the mark with water in a 1 000 ml one-mark volumetric flask and mix.

Either solution 4.2.1 or 4.2.2 shall pass the following test: 20,0 ml of the solution, diluted to the mark in a 100 ml one-mark volumetric flask, shall not give absorbances of the calcium and magnesium lines greater than those obtained with a standard aqueous solution containing 3 μg of Ca and 1 μg of Mg per 100 ml, when tested in accordance with 6.2.2 and 6.2.3.

In addition, the quantities of Ca and Mg corresponding to the measured absorbances shall be recorded so that account may be taken of these values in checking the purity of the sodium hydroxide (4.3).

4.3 Sodium hydroxide

This product, which is used in the preparation of the sodium chloride solution (4.4), should preferably not have Ca and Mg contents greater than 2,5 mg/kg and 0,5 mg/kg respectively. Check these contents as follows:

Note, from the calibration graphs (6.2.2 and 6.2.3) the absorbances obtained with the standard matching solution No. 0 (6.2.1). The Ca and Mg contents corresponding to these absorbances shall not exceed 2,5 μg and 0,5 μg , respectively, discounting quantities of Ca and Mg found in the check test on the lanthanum chloride solution (4.2) in the case where this solution is used.

4.4 Sodium chloride, 58,5 g/l acid solution.

Place 20,0 g of the sodium hydroxide (4.3) in a 600 ml polyethylene beaker. Dissolve, while cooling, in 100 ml of water, then acidify by adding carefully, while stirring 80 ml of the hydrochloric acid solution (4.1). Transfer quantitatively to 500 ml conical flask, boil for 5 min, allow to cool, transfer quantitatively to a 500 ml one-mark volumetric flask, dilute to the mark and mix.

NOTE If very pure sodium chloride is available, this solution (4.4) can be prepared as follows:

Place 29 g of sodium chloride in a 500 ml conical flask. Add 250 ml of water and 40 ml of the hydrochloric acid solution (4.1). Boil for 5 min, allow to cool, transfer quantitatively to a 500 ml one-mark volumetric flask, dilute to the mark and mix.

4.5 Calcium, standard solution corresponding to 0,100 g of calcium per litre.¹⁾

Weigh, to the nearest 0,000 1 g, 0,249 7 g of calcium carbonate previously dried at about 250 °C for 2 h and cooled in a desiccator. Place in a 250 ml beaker and dissolve in a mixture of 10 ml of the hydrochloric acid solution (4.1) and 15 ml of water. Transfer quantitatively to a 1 000 ml one-mark volumetric flask, dilute to the mark and mix.

1 ml of this standard solution contains 0,100 mg of Ca.

¹⁾ The spectrophotometric standard reference solutions available commercially can also be used.

4.6 Calcium, standard solution corresponding to 10 mg of calcium per litre.

Transfer 20,0 ml of the standard calcium solution (4.5) to a 200 ml one-mark volumetric flask, dilute to the mark and mix.

1 ml of this standard solution contains 10 μg of Ca. Prepare this solution just before use.

4.7 Magnesium, standard solution corresponding to 0,100 g of magnesium per litre.¹⁾

Weigh, to the nearest 0,000 1 g, 0,100 0 g of magnesium metal of high purity (minimum 99,95 %). Place in a 250 ml beaker and dissolve in a mixture of 10 ml of the hydrochloric acid solution (4.1) and 15 ml of water. Transfer quantitatively to a 1 000 ml one-mark volumetric flask, dilute to the mark and mix.

1 ml of this standard solution contains 0,100 mg of Mg.

4.8 Magnesium, standard solution corresponding to 10 mg of magnesium per litre.

Transfer 20,0 ml of the standard magnesium solution (4.7) to a 200 ml one-mark volumetric flask, dilute to the mark and mix.

1 ml of this standard solution contains 10 μg of Mg. Prepare this solution just before use.

5 Apparatus

Ordinary laboratory apparatus, and

5.1 Atomic absorption spectrophotometer, fitted with a burner fed either with acetylene and dinitrogen monoxide or with acetylene and air.

5.2 Hollow-cathode calcium lamp

5.3 Hollow-cathode magnesium lamp

NOTE All glassware and reagent bottles shall be either of borosilicate glass or of a quality not yielding calcium or magnesium.

6 Procedure

6.1 Test portion

Take 25,0 ml of the solution A, prepared in accordance with ISO 3195 and containing 40 g of the test sample per 1 000 ml, stored in a polyethylene bottle.

6.2 Preparation of the calibration graphs

6.2.1 Preparation of the standard matching solutions

Into each of a series of five 100 ml one-mark volumetric flasks, place 25 ml of the acid sodium chloride solution (4.4) and, if using the acetylene-air flame, 20 ml of the lanthanum chloride solution (4.2).

Then add the volumes of the standard calcium solution (4.6) and standard magnesium solution (4.8) indicated in the following table:

Standard matching solution No.	Standard solutions		Corresponding masses	
	calcium (4.6)	magnesium (4.8)	Ca	Mg
	ml	ml	μg	μg
0 ^a	0	0	0	0
1	1,0	0,2	10	2
2	5,0	1,0	50	10
3	10,0	2,0	100	20
4	20,0	4,0	200	40

^a Blank test on reagents for calibration graphs.

Dilute to the mark and mix.

6.2.2 Calibration graphs for calcium

6.2.2.1 ADJUSTMENT OF THE APPARATUS

Install the hollow-cathode calcium lamp (5.2) in the apparatus (5.1) and leave it switched on for the time necessary to achieve stability. Adjust the lamp current, the attenuation and the slit, to suit the characteristics of the apparatus. Adjust the wavelength in the region of 422,7 nm in order to obtain the maximum absorbance. Adjust the acetylene and air or dinitrogen monoxide pressures according to the characteristics of the aspirator-burner.

6.2.2.2 SPECTROPHOTOMETER MEASUREMENTS

Aspirate the series of standard matching solutions (6.2.1) into the flame and measure the absorbance for each. Take care to keep the rate of aspiration constant throughout the preparation of the calibration graph.

NOTE Pass water through the burner after each measurement.

6.2.2.3 PLOTTING THE CALIBRATION GRAPH

Plot a graph having, for example, the numbers of micrograms of Ca contained in 100 ml of the standard matching solutions as abscissae and the corresponding values of the measured absorbances, reduced by the measured value for the standard matching solution No. 0, as ordinates.

6.2.3 Calibration graph for magnesium

Repeat the operations specified in 6.2.2.1, 6.2.2.2 and 6.2.2.3, after having fitted the hollow-cathode magnesium lamp (5.3) to the apparatus and having adjusted the wavelength to about 285,2 nm.

NOTE The calibration graphs obtained should be linear as far as possible.

6.3 Determination

6.3.1 Preparation of the test solution

Transfer the test portion (6.1) to a 100 ml conical flask, add 20 ml of water and, slowly while stirring, 4,0 ml of the hydrochloric acid solution (4.1). Boil gently for 5 min, cool and transfer quantitatively to a 100 ml one-mark volumetric flask. If using the acetylene-air flame, add 20,0 ml of the lanthanum chloride solution (4.2). Dilute to the mark and mix.

6.3.2 Blank test

Carry out a blank test at the same time as the determination, following the same procedure and using the same quantities of reagents as used in the determination, but omitting the test portion.

6.3.3 Spectrophotometric measurements

Following the procedures specified in 6.2.2 and 6.2.3, determine, in order, the absorbances for the following solutions:

- the five standard matching solutions (6.2.1);
- the test solution (6.3.1);
- the blank test solution (6.3.2);
- the standard matching solution (6.2.1) for which the absorbance is closest to that of the test solution.

6.3.4 Check test

Verify the absence of interferences by repeating the operations specified in 6.3.1 with the addition of known quantities of the standard calcium and magnesium solutions (4.6 and 4.8). If interferences are detected, repeat the calibration using the standard additions method, starting with test solutions prepared as specified in 6.3.1. In this case, ensure that the operations are carried out in the linear part of the calibration graph.

7 Expression of results

7.1 The calcium content, expressed as milligrams of Ca per kilogram, is given by the formula

$$m_1 \times \frac{1}{1\,000} \times \frac{A_1 - A_0}{A_2 - A_3} \times \frac{1\,000}{25} \times \frac{1\,000}{m_0} = \frac{40 m_1 (A_1 - A_0)}{m_0 \times (A_2 - A_3)}$$

where

A_0 is the absorbance of the blank test solution (6.3.2);

A_1 is the absorbance of the test solution (6.3.1);

A_2 is the mean of the measurements on the standard matching solution closest in absorbance to that of the test solution, carried out before and after the latter;

A_3 is the absorbance of the standard matching solution No. 0 (6.2.1);

m_0 is the mass, in grams, of the test sample used in the preparation of the main solution A;

m_1 is the mass, in micrograms, of Ca contained in the standard matching solution closest in absorbance to that of the test solution.

7.2 Calculate the magnesium content according to 7.1, using the relevant absorbances for magnesium.

8 Test report

The test report shall include the following particulars:

- a) the reference of the method used;
- b) the results and the method of expression used;
- c) any unusual features noted during the determination;
- d) any operation not included in this International Standard or in the International Standard to which reference is made, or regarded as optional.

Annex ISO publications relating to sodium hydroxide for industrial use

ISO 979, *Method of assay.*

ISO 980, *Determination of carbonates content — Gas-volumetric method.*

ISO 981, *Determination of chloride content — Mercurimetric method.*

ISO 982, *Determination of sulphate content — Barium sulphate gravimetric method.*

ISO 983, *Determination of iron content — 1,10-Phenanthroline photometric method.*

ISO 984, *Determination of silica content — Reduced silicomolybdic complex photometric method.*

ISO 985, *Determination of silica content — Gravimetric method by precipitation of quinoline molybdosilicate.*

ISO 986, *Determination of calcium — EDTA (disodium salt) complexometric method.*

ISO 3195, *Sampling — Test sample — Preparation of the main solution for carrying out certain determinations.*

ISO 3196, *Determination of carbon dioxide content — Titrimetric method.*

ISO 3197, *Determination of chloride content — Photometric method.*

ISO 3198, *Determination of sulphur compounds — Method by reduction and titrimetry.*

ISO 3697, *Determination of calcium and magnesium contents — Flame atomic absorption method.*

Publications referred to

See national foreword.

BSI — British Standards Institution

BSI is the independent national body responsible for preparing British Standards. It presents the UK view on standards in Europe and at the international level. It is incorporated by Royal Charter.

Revisions

British Standards are updated by amendment or revision. Users of British Standards should make sure that they possess the latest amendments or editions.

It is the constant aim of BSI to improve the quality of our products and services. We would be grateful if anyone finding an inaccuracy or ambiguity while using this British Standard would inform the Secretary of the technical committee responsible, the identity of which can be found on the inside front cover.
Tel: 020 8996 9000. Fax: 020 8996 7400.

BSI offers members an individual updating service called PLUS which ensures that subscribers automatically receive the latest editions of standards.

Buying standards

Orders for all BSI, international and foreign standards publications should be addressed to Customer Services. Tel: 020 8996 9001. Fax: 020 8996 7001.

In response to orders for international standards, it is BSI policy to supply the BSI implementation of those that have been published as British Standards, unless otherwise requested.

Information on standards

BSI provides a wide range of information on national, European and international standards through its Library and its Technical Help to Exporters Service. Various BSI electronic information services are also available which give details on all its products and services. Contact the Information Centre.
Tel: 020 8996 7111. Fax: 020 8996 7048.

Subscribing members of BSI are kept up to date with standards developments and receive substantial discounts on the purchase price of standards. For details of these and other benefits contact Membership Administration.
Tel: 020 8996 7002. Fax: 020 8996 7001.

Copyright

Copyright subsists in all BSI publications. BSI also holds the copyright, in the UK, of the publications of the international standardization bodies. Except as permitted under the Copyright, Designs and Patents Act 1988 no extract may be reproduced, stored in a retrieval system or transmitted in any form or by any means – electronic, photocopying, recording or otherwise – without prior written permission from BSI.

This does not preclude the free use, in the course of implementing the standard, of necessary details such as symbols, and size, type or grade designations. If these details are to be used for any other purpose than implementation then the prior written permission of BSI must be obtained.

If permission is granted, the terms may include royalty payments or a licensing agreement. Details and advice can be obtained from the Copyright Manager.
Tel: 020 8996 7070.