



BSI Standards Publication

Chemical analysis of ferrous materials — Determination of selenium in steels — Electrothermal atomic absorption spectrometric method

National foreword

This Published Document is the UK implementation of CEN/TR 10362:2014.

The UK participation in its preparation was entrusted to Technical Committee ISE/102, Methods of Chemical Analysis for Iron and Steel.

A list of organizations represented on this committee can be obtained on request to its secretary.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

© The British Standards Institution 2014. Published by BSI Standards Limited 2014

ISBN 978 0 580 84816 2

ICS 77.040.30; 77.080.20

Compliance with a British Standard cannot confer immunity from legal obligations.

This Published Document was published under the authority of the Standards Policy and Strategy Committee on 30 November 2014.

Amendments issued since publication

Date	Text affected
------	---------------

ICS 77.040.30; 77.080.20

English Version

**Chemical analysis of ferrous materials - Determination of
selenium in steels - Electrothermal atomic absorption
spectrometric method**

Analyse chimique des produits ferreux - Détermination du
sélénium dans les aciers - Méthode par spectrométrie
d'absorption atomique électrothermique

Chemische Analyse von Eisenwerkstoffen - Bestimmung
von Selen in Stahl - Spektrometrisches Verfahren mit
elektrothermischer Atomabsorption

This Technical Report was approved by CEN on 10 May 2014. It has been drawn up by the Technical Committee ECISS/TC 102.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

Contents

Page

Foreword.....	3
1 Scope	4
2 Normative references	4
3 Principle.....	4
4 Reagents.....	4
5 Apparatus	5
6 Sampling.....	5
7 Procedure	6
8 Measurements.....	7
9 Expression of results	7
10 Precision.....	8
11 Test report	8
Annex A (informative) Typical operating parameters	9
Annex B (informative) Instruments and the instrumental conditions used by the participants to the precision test	11
Annex C (informative) Composition of the samples used for the precision test.....	12
Annex D (informative) Detailed results obtained from the precision test	14
Bibliography.....	19

Foreword

This document (CEN/TR 10362:2014) has been prepared by Technical Committee ECISS/TC 102 "Methods of chemical analysis for iron and steel", the secretariat of which is held by SIS.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

1 Scope

This Technical Report specifies an electrothermal atomic absorption spectrometric method for the determination of selenium in steels.

The method is applicable to selenium contents between 0,000 4 % (m/m) and 0,02 % (m/m).

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN ISO 648, *Laboratory glassware - Single-volume pipettes (ISO 648)*

EN ISO 1042, *Laboratory glassware - One-mark volumetric flasks (ISO 1042)*

EN ISO 3696, *Water for analytical laboratory use - Specification and test methods (ISO 3696)*

3 Principle

Dissolution of a test portion in hydrochloric and nitric acids and dilution of the solution to a known volume.

Introduction of a known volume of the solution into the electrothermal atomizer of an atomic absorption spectrometer.

Measurement of the absorption of the 196,0 nm spectral line energy emitted by a selenium hollow-cathode lamp, using Zeeman effect background correction.

Calibration by the standard addition technique.

4 Reagents

During the analysis, use only reagents of recognised analytical grade and only grade 3 water, as specified in EN ISO 3696.

4.1 Nitric acid, HNO₃ ($\rho_{20} = 1,40$ g/ml)

4.2 Hydrochloric acid, HCl ($\rho_{20} = 1,19$ g/ml)

4.3 Matrix modifiers

The matrix modifiers described in 4.3.1 and 4.3.2 are recommended. Each laboratory has to investigate on its own equipment which of them is the most suitable, regarding sensitivity and recovery.

4.3.1 Palladium-nickel modifier

Prepare a palladium solution (1 mg/ml Pd) by dissolving 167 mg of PdCl₂ in 100 ml of hot water and 1 ml of nitric acid (4.1).

Prepare a nickel solution (1 mg/ml) by dissolving 1 g of nickel (Ni > 99,999 %) in 20 ml of water, 20 ml of nitric acid (4.1) and 5 ml of hydrochloric acid (4.1). Heat until the metal is dissolved. After cooling, transfer the solution into a 1 l one-mark volumetric flask, dilute to the mark with water and mix well.

Into a 50 ml volumetric flask, mix 35 ml of the 1 mg/ml palladium solution with 15 ml of the 1 mg/ml nickel solution. This solution contains 700 µg/ml Pd and 300 µg/ml Ni.

4.3.2 Palladium-magnesium modifier

Prepare a PdCl_2 solution by dissolving 500 mg of PdCl_2 in 100 ml of hot water and 1 ml of nitric acid (4.1).

Prepare a $\text{Mg}(\text{NO}_3)_2$ solution by dissolving 350 mg of $\text{Mg}(\text{NO}_3)_2 \cdot 6\text{H}_2\text{O}$ in 100 ml of water and 1 ml of nitric acid (4.1).

Mix equal volumes of the PdCl_2 solution and the $\text{Mg}(\text{NO}_3)_2$ solution.

4.4 Selenium standard solution, 1 g/l

Weigh, to the nearest 0,001 g, 1 g of high purity selenium [min 99,9 % (mass fraction)], transfer into a 100 ml beaker and cover with a watch glass.

Dissolve it in 35 ml of nitric acid (4.1). Heat to complete dissolution at a temperature just below the boiling point (approximately 150 °C) during at least 30 minutes. After cooling, transfer the solution quantitatively into a 1 000 ml one-mark volumetric flask, dilute to the mark with water and mix well.

1 ml of this solution contains 1 mg of selenium.

4.5 Selenium standard solution, 0,01 g/l

Transfer 10,0 ml of selenium standard solution (4.4) into a 1 000 ml one-mark volumetric flask. Add 120 ml of hydrochloric acid (4.2) and 40 ml of nitric acid (4.1). Dilute to the mark with water and mix well.

1 ml of this solution contains 0,01 mg of selenium.

4.6 Pure iron, containing less than 0,000 1 % (mass fraction) of selenium

5 Apparatus

All volumetric glassware shall be Class A and calibrated, in accordance with ISO 648 or ISO 1042 as appropriate.

Before use, all glassware shall be cleaned by boiling with hydrochloric acid to remove any chemical contamination.

5.1 Auto sampler equipped with micropipettes of capacity 10 μl to 50 μl

5.2 Atomic absorption spectrometer and electrothermal atomizer

This shall be equipped with a selenium hollow-cathode lamp or an electrodeless discharge lamp and supplied with pure argon.

The instrument shall be fitted with Zeeman effect background correction.

An electrothermal atomizer equipped with a L'vov platform, mounted in a pyrolytically coated graphite tube, supplied with argon as purge gas, is recommended.

The characteristic mass for selenium shall be less than 30 pg of selenium.

6 Sampling

Sampling shall be carried out in accordance with EN ISO 14284 or with an appropriate national standard for steels.

7 Procedure

7.1 Test portion

According to the expected selenium content (x) of the test sample, weigh to the nearest 0,1 mg, a test portion as indicated in Table 1.

Table 1 — Mass of the test portion

Expected selenium content (x) of the test sample ($\mu\text{g/g}$)	Mass of the test portion (mg)
$x < 15$	2 000
$15 \leq x < 35$	1 000
$35 \leq x < 75$	400
$75 \leq x < 150$	200
$150 \leq x \leq 200$	100

7.2 Blank test

Carry out a blank test simultaneously with the determination, following the same procedure and using the same quantities of all reagents as used for the determination, but substituting pure iron (4.6) for the test portion.

7.3 Preparation of the test solution

Transfer the test portion (7.1) into a 250 ml beaker tall form. Cover the beaker with a watch glass and add 10 ml of water, 10 ml of nitric acid (4.1) and 30 ml of hydrochloric acid (4.2).

Heat at a temperature just below the boiling point (approximately 150 °C) during at least two hours. Allow to cool. If necessary, filter through a medium texture filter paper and collect the filtrate in a 250 ml one-mark volumetric flask. Wash the filter paper several times with hot water and collect the washings in the same 250 ml volumetric flask.

Allow to cool, dilute to the mark with water and mix.

7.4 Preparation of the standard addition solutions for the test solution

Into a series of 50 ml one-mark volumetric flasks introduce 25 ml aliquots of the test solution (7.3) and add the volumes of selenium standard solution (4.5) shown in Table 2. Dilute to the mark with water and mix.

These solutions are labelled S1, S2, S3 and S4 respectively.

Table 2 — Standard addition solutions for the test solution

Label of the solution	Selenium standard solution volume μl	Concentration of selenium in the test addition solutions $\mu\text{g/ml}$
S1	0	0
S2	200	0,04
S3	400	0,08
S4	600	0,12

7.5 Preparation of the standard addition solutions for the blank test

Into a series of 50 ml one-mark volumetric flasks introduce 25 ml aliquots of the blank solution (7.2) and add the volumes of selenium standard solution (4.5) shown in Table 3. Dilute to the mark with water and mix.

These solutions are labelled B1, B2, B3 and B4 respectively.

Table 3 — Standard addition solutions for the blank test

Label of the solution	Selenium standard solution volume μl	Concentration of selenium in the test addition solutions $\mu\text{g/ml}$
B1	0	0
B2	200	0,04
B3	400	0,08
B4	600	0,12

8 Measurements

Set the required instrument parameters and align the electrothermal atomiser according to the manufacturer's instructions (see NOTE 1 and 2).

Adjust the wavelength in the region of 196,0 nm to minimum absorbance.

NOTE 1 The operating parameters for electrothermal atomic absorption spectrometry vary considerably from an instrument to another and much more than for flame atomic absorption spectrometry. Typical operating parameters are given in Annex A. Annex B details the instruments and the instrumental conditions used by the laboratories having participated to the interlaboratory test programme (see Clause 10).

Fill the auto sampler with the blank standard addition solutions (7.5), the test standard addition solutions (7.4) and the matrix-modifier (4.3.1 or 4.3.2).

Inject sequentially 10 μl of each of the solutions 7.5 and 7.4 and add to each solution 10 μl of the matrix modifier (4.3.1 or 4.3.2). Measure the absorbance of each solution at the adjusted wavelength. Correct the absorbances for background simultaneously.

NOTE 2 In general the absorbance is measured by peak height or peak area. The determination by peak area is recommended.

9 Expression of results

9.1 Plotting of the standard addition graphs

Plot the absorbances corresponding to the blank standard addition solutions against the concentration of selenium added, expressed in $\mu\text{g/ml}$.

Plot the absorbances corresponding to the test standard addition solutions against the concentration of selenium added, expressed in $\mu\text{g/ml}$.

Use least squares regression method to calculate the lines for the blank standard addition solutions and for the test standard addition solutions.

9.2 Method of calculation

Determine the concentrations of selenium, expressed in µg/ml, in the test solution and in the blank test C_S and C_B as the two intercepts on the concentration axis (see 9.1).

The difference ($C_S - C_B$) gives the net concentration of selenium in the test solution (solution S1).

Calculate the selenium mass fraction, in per cent (%), as follows:

$$W_{Se} = \frac{(C_S - C_B) \times 2 \times 250 \times 100}{m \times 1000}$$

where:

- W_{Se} is the selenium mass fraction in per cent (%);
- C_S is the selenium concentration in the test solution, in micrograms per millilitre (µg/ml);
- C_B is the selenium concentration in the blank test solution, in micrograms per millilitre (µg/ml);
- m is the mass of the test portion, in milligrams (mg).

10 Precision

Four laboratories in four European countries participated in an inter laboratory test programme under the auspices of ECISS/TC 102/WG 3, involving three determinations of selenium at nine levels.

Each laboratory carried out two determinations under repeatability conditions as defined in ISO 5725-1, i.e. one operator, same apparatus, identical operating conditions, same standard addition solutions and a minimum period of time. The third determination was carried out on a different day using the same apparatus with different standard addition solutions.

Details on the samples used are given in Annex C and the results obtained are reported in Annex D.

The sets of data available were too few for a statistical evaluation based on ISO 5725-2 and ISO 5725-3 and CEN/TR 10345. Only Mandel's h and k statistics were used (see Figures D.1 and D.2) rather for illustrative purposes.

Annex D also provides some considerations on the data presented.

11 Test report

The test report shall contain the following information:

- a) identification of the test sample;
- b) method used;
- c) results;
- d) any unusual characteristics noted during the determination;
- e) any operation not included in this Technical Report or in the document to which reference is made or regarded as optional;
- f) date of the test and/or date of preparation or signature of the test report;
- g) signature of the responsible person.

Annex A (informative)

Typical operating parameters

A.1 Typical conditions for the determination of selenium by electrothermal atomic absorption spectrometry

Injection volumes	10 µl of the test standard addition solutions and of the blank standard addition solutions and 10 µl of the matrix modifier solution;
Matrix modifier	Pd-Ni or Pd-Mg;
Instrumental settings	See Table A.1

Table A.1 — Instrumental settings

Wavelength nm	Slit width nm	Ashing temperature °C	Atomisation temperature °C
196,0	0,7	500	2000

Background Correction: Zeeman effect;

Typical temperature profile: see Table A.2;

Table A.2 — Typical temperature profile

Step	Temperature °C	Ramp s	Hold s	Gas flow ml/min	Read
1	100	10	10	300	
2	140	10	20	300	
3	500	10	20	300	
4	2000	0	4	0	YES
5	2800	10	2	300	

Measurement mode: Peak area;

Integration time: 4 s.

A.2 Performance criteria

Detection limit	0,02 µg/ml Se
Characteristic mass	< 30 pg Se
Maximum concentration	0,4 µg/ml Se

A.3 Criterion for graph linearity

A.3.1 General

In the case of standard addition method, the graph linearity should be checked as described in A.3.2 and A.3.3.

A.3.2 Linearity of the blank standard addition graph

Having established the blank standard additions graph, the slope of the graph covering the top concentration range, B3 to B4 (expressed as a change in absorbance) should preferably not be less than 0,90 times of the value of the slope of the bottom concentration range, B1 to B2 (expressed as a change in absorbance).

A.3.3 Linearity of the test standard addition graph

Having established the test standard additions graph, the slope of the graph covering the top concentration range, S3 to S4 (expressed as a change in absorbance) should preferably not be less than 0,9 times of the value of the slope of the bottom concentration range, S1 to S2 (expressed as a change in absorbance).

Annex B (informative)

Instruments and the instrumental conditions used by the participants to the precision test

Table B.1 shows the instrument models and the instrumental conditions used by the participants to the precision test.

Table B.1 — Instruments and the instrumental conditions used by the participants to the precision test

Laboratory	1				2				3				4			
Instrument	Analytik Jena AAS Zeenit 650s				Unicam Solaar 989 QZ				Unicam Solaar 989 QZ				Thermo Solaar M6			
Matrix modifier	Pd-Mg(NO ₃) ₂				Pd-Ni				Mg(NO ₃) ₂ -Pd				Pd-Mg			
Wavelength (nm)	196,0				196,0				196,0				196,0			
Slit width (nm)	1,2				0,5				0,5				0,5			
Ashing temperature (°C)	700				500				500 (samples 1 and 2) 600 (samples 3 to 9)				500			
Atomisation Temperature (°C)	2 000				2 100				2 000				2 000			
Temperature profiles	Temp (°C)	Ramp (°C/s)	Hold (s)	Gas (l/min)	Temp (°C)	Ramp (°C/s)	Hold (s)	Gas (l/min)	Temp (°C)	Ramp (°C/s)	Hold (s)	Gas (l/min)	Temp (°C)	Ramp (°C/s)	Hold (s)	Gas (l/min)
	90	5	15		100	10	10	0,3	100	5	10		100	10	10	0,3
	100	2	20		140	5	20	0,3	140	5	20		140	5	20	0,3
	700	200	25		500	35	20	0,3	500	150	20		500	35	20	0,3
	700	0	5		2 100		5	0	2 000	0	4		2 000		5	0
	2 000	1 400	4		2 900		8	0,3	2 800	0	2		2 700		8	0,3
	2 400	500	4													
Other details	Graphite tubes with platform								Coated platform				cuvette without			

Annex C (informative)

Composition of the samples used for the precision test

The composition of the samples used for the precision test are listed in Table C.1.

Table C.1 — Composition of the samples used for the precision test

Sample label	Se (µg/g)	C %	Si %	Mn %	P %	S %	Cr %	Mo %	Ni %	Al %	As %	B %	Co %	Cu %	N %	Nb %	Pb %
C1153a	7	0,225	1,00	0,544	0,030	0,019	16,70	0,24	8,76	<i>0,004</i>	<i>0,007</i>	<i>0,001</i>	0,127	0,226	<i>0,11</i>	<i>0,48</i>	0,006
GBW01620	16										0,0011						0,0004
NIST 1765	35	0,006	<i>0,004</i>	0,144	0,0052	0,0038	0,051	0,005	0,154	<i>0,006</i>	0,0010	0,0009	0,0012	0,0013	0,0010	0,0004	0,0003
NIST 1766	35	0,015	0,010	0,067	0,002	0,0024	0,024	0,0035	0,021	0,012	0,0035	0,0001	0,0020	0,015	0,0033	0,005	0,003
NIST 1261a	40	0,39	0,228	0,67	0,016	0,015	0,69	0,19	2,00	0,02	0,017	0,0005	0,032	0,042	<i>0,0037</i>	0,022	
GBW01622	43										0,0072						0,0002
MBH 12x352	60	0,275	0,414	0,200	0,126	0,070	0,313	0,261	0,376	0,257	0,036		0,020	0,140	<i>0,006</i>	0,104	0,006
BS 12B	120	0,95	0,365	0,081	0,010	0,010	1,58	0,30	0,127	0,016	0,002		0,023	0,116	0,0141	0,156	0,004
MBH 12x353	170	0,194	0,142	0,817	0,0205	0,025	0,491	0,113	0,190	0,0070	0,056		0,057	0,302	0,016	0,096	<i>0,018</i>

NOTE 1 Values given in italics are non-certified.

NOTE 2 Samples GBW01620 and GBW01622 are “defined” as an “iron and nickel base high temperature alloy”.

Table C.1 (cont.)

Sample label	Sn %	Ti %	V %	W %	Ag %	Bi %	Ca %	Cd %	Ce %	In %	Mg %	Sb %	Ta %	Te %	Zn %	Zr %
C1153a	<i>0,013</i>	<i>0,002</i>	0,176		<i>0,001</i>							<i>0,003</i>	<i>0,03</i>	<i>0,01</i>	<i>0,005</i>	
GBW01620	0,0053	0,0022			0,0005		0,0032	0,0005		0,0003	0,0016	0,0095		0,0011	0,0032	
NIST 1765	0,002	0,0055	0,004		0,0002							0,0010	<i>0,004</i>	<i>0,003</i>		<i>0,0002</i>
NIST 1766	0,0010	0,0005	0,009	<i>0,001</i>	0,0005				<i>0,002</i>			0,0005	<i>0,006</i>	<i>0,003</i>		<i>0,0004</i>
NIST 1261a	0,010	0,020	0,011	0,017	0,0004	0,0004			0,0014			0,0042	0,020	0,0006		0,009
GBW01622	0,104						0,0032	0,0002			0,0053	0,0007		0,0083	0,0020	
MBH 12x352	0,110	0,29	0,029	0,202		0,008						0,023			0,005	
BS 12B	0,024	0,055	0,105	0,014			0,0008		0,020			0,014	0,024	0,02	0,0003	0,005
MBH 12x353	0,075	0,057	0,0481	0,108		0,021						0,156			0,048	

NOTE 1 Values given in italics are non-certified.

NOTE 2 Samples GBW01620 and GBW01622 are "defined" as an "iron and nickel base high temperature alloy".

Annex D (informative)

Detailed results obtained from the precision test

D.1 Results

The results from the precision test are shown in Table D.1. Mandel's h and k statistics are represented in Figures D.1 and D.2.

Table D.1 — Results obtained from the precision test ($\mu\text{g/g}$)

		C 1153a		GBW0 1620		NIST 1765	
		Day 1	Day 2	Day 1	Day 2	Day 1	Day 2
LAB 1	Individual values	1,8	3,4	17,5	17,2	39,3	40,5
		5,6		19,3		23,3	
	Laboratory mean	3,6		18,0		34,4	
LAB 2	Individual values	2,2	2,7	13,0	10,0	35,0	27,0
		2,7		15,0		36,0	
	Laboratory mean	2,5		12,7		32,7	
LAB 3	Individual values	4,4	4,8	15,0	17,0	48,0	38,0
		3,2		15,0		39,0	
	Laboratory mean	4,1		15,7		41,7	
LAB 4	Individual values	3,1	2,8	11,0	18,0	33,0	36,0
		2,5		13,0		29,0	
	Laboratory mean	2,8		14,0		32,7	
Mean value ($\mu\text{g/g}$)		3,3		15,1		35,3	
Standard deviation ($\mu\text{g/g}$)		0,7		2,3		4,3	
Relative standard deviation (RSD) (%)		22,5		15,2		12,1	
Certified value ($\mu\text{g/g}$) and uncertainty (1σ)		7		16 \pm 2		35	

Table D.1 (cont.)

		NIST 1766		NIST 1261a		GBW0 1622	
		Day 1	Day 2	Day 1	Day 2	Day 1	Day 2
LAB 1	Individual values	42,7	37,2	36,1	47,1	54,1	50,7
		40,2		38,2		52,4	
	Laboratory mean	40,0		40,5		52,4	
LAB 2	Individual values	26,0	23,0	30,0	27,0	37,0	34,0
		26,0		34,0		39,0	
	Laboratory mean	25,0		30,3		36,7	
LAB 3	Individual values	34,0	28,0	39,0	41,0	49,0	45,0
		31,0		40,0		49,0	
	Laboratory mean	31,0		40,0		47,7	
LAB 4	Individual values	32,0	37,0	36,0	42,0	51,0	51,0
		39,0		40,0		49,0	
	Laboratory mean	36,0		39,3		50,3	
Mean value ($\mu\text{g/g}$)		33,0		37,5		46,8	
Standard deviation ($\mu\text{g/g}$)		6,5		4,8		7,0	
Relative standard deviation (RSD) (%)		19,7		12,9		15,0	
Certified value ($\mu\text{g/g}$) and uncertainty (1σ)		35		38 ± 10		47	

Table D.1 (cont.)

		MBH 12x352		BS 12B		MBH 12x353	
		Day 1	Day 2	Day 1	Day 2	Day 1	Day 2
LAB 1	Individual values	60,1	70,5	126,3	105,0	197,0	203,0
		58,4		118,7		199,0	
	Laboratory mean	63,0		116,7		199,7	
LAB 2	Individual values	41,0	39,0	106,0	123,0	163,0	172,0
		45,0		120,0		165,0	
	Laboratory mean	41,7		116,3		166,7	
LAB 3	Individual values	50,0	49,0	102,0	130,0	145,0	148,0
		48,0		123,0		148,0	
	Laboratory mean	49,0		118,3		147,0	
LAB 4	Individual values	52,0	62,0	133,0	125,0	166,0	182,0
		57,0		128,0		182,0	
	Laboratory mean	57,0		128,7		176,7	
Mean value ($\mu\text{g/g}$)		52,7		120,0		172,5	
Standard deviation ($\mu\text{g/g}$)		9,3		5,8		21,9	
Relative standard deviation (RSD) (%)		17,7		4,9		12,7	
Certified value ($\mu\text{g/g}$) and uncertainty (1σ)		60 \pm 5		120 \pm 20		170 \pm 15	




Figure D.1 — Selenium — Mandel's k test for intra laboratory consistency (99 % confidence level)




Figure D.2 — Selenium — Mandel's h test for between laboratory consistency (99 % confidence level)

D.2 Comments on the suitability of the data

On the basis of the data shown in Table D.1 and on the Mandel's statistics from Figures D.1 and D.2, no significant discrepancies within the whole results are to be noted.

Results from Lab 2 show a systematic low tendency but this remains within the “acceptable limit” for most of the samples.

Except for sample C 1153a, there is a good agreement between the mean value from the “precision test” and the corresponding “certified or referee values”. This “confirms” on the trueness of the method described in the present Technical Report.

Bibliography

- [1] EN ISO 14284, *Steel and iron - Sampling and preparation of samples for the determination of chemical composition (ISO 14284)*

British Standards Institution (BSI)

BSI is the national body responsible for preparing British Standards and other standards-related publications, information and services.

BSI is incorporated by Royal Charter. British Standards and other standardization products are published by BSI Standards Limited.

About us

We bring together business, industry, government, consumers, innovators and others to shape their combined experience and expertise into standards-based solutions.

The knowledge embodied in our standards has been carefully assembled in a dependable format and refined through our open consultation process. Organizations of all sizes and across all sectors choose standards to help them achieve their goals.

Information on standards

We can provide you with the knowledge that your organization needs to succeed. Find out more about British Standards by visiting our website at bsigroup.com/standards or contacting our Customer Services team or Knowledge Centre.

Buying standards

You can buy and download PDF versions of BSI publications, including British and adopted European and international standards, through our website at bsigroup.com/shop, where hard copies can also be purchased.

If you need international and foreign standards from other Standards Development Organizations, hard copies can be ordered from our Customer Services team.

Subscriptions

Our range of subscription services are designed to make using standards easier for you. For further information on our subscription products go to bsigroup.com/subscriptions.

With **British Standards Online (BSOL)** you'll have instant access to over 55,000 British and adopted European and international standards from your desktop. It's available 24/7 and is refreshed daily so you'll always be up to date.

You can keep in touch with standards developments and receive substantial discounts on the purchase price of standards, both in single copy and subscription format, by becoming a **BSI Subscribing Member**.

PLUS is an updating service exclusive to BSI Subscribing Members. You will automatically receive the latest hard copy of your standards when they're revised or replaced.

To find out more about becoming a BSI Subscribing Member and the benefits of membership, please visit bsigroup.com/shop.

With a **Multi-User Network Licence (MUNL)** you are able to host standards publications on your intranet. Licences can cover as few or as many users as you wish. With updates supplied as soon as they're available, you can be sure your documentation is current. For further information, email bsmusales@bsigroup.com.

BSI Group Headquarters

389 Chiswick High Road London W4 4AL UK

Revisions

Our British Standards and other publications are updated by amendment or revision.

We continually improve the quality of our products and services to benefit your business. If you find an inaccuracy or ambiguity within a British Standard or other BSI publication please inform the Knowledge Centre.

Copyright

All the data, software and documentation set out in all British Standards and other BSI publications are the property of and copyrighted by BSI, or some person or entity that owns copyright in the information used (such as the international standardization bodies) and has formally licensed such information to BSI for commercial publication and use. Except as permitted under the Copyright, Designs and Patents Act 1988 no extract may be reproduced, stored in a retrieval system or transmitted in any form or by any means – electronic, photocopying, recording or otherwise – without prior written permission from BSI. Details and advice can be obtained from the Copyright & Licensing Department.

Useful Contacts:

Customer Services

Tel: +44 845 086 9001

Email (orders): orders@bsigroup.com

Email (enquiries): cservices@bsigroup.com

Subscriptions

Tel: +44 845 086 9001

Email: subscriptions@bsigroup.com

Knowledge Centre

Tel: +44 20 8996 7004

Email: knowledgecentre@bsigroup.com

Copyright & Licensing

Tel: +44 20 8996 7070

Email: copyright@bsigroup.com



...making excellence a habit.™