

PD IEC/PAS 63077:2016



BSI Standards Publication

# Good refurbishment practices for medical imaging equipment

### **National foreword**

This Published Document is the UK implementation of IEC/PAS 63077:2016.

The UK participation in its preparation was entrusted by Technical Committee CH/62, Electrical Equipment in Medical Practice, to Subcommittee CH/62/2, Diagnostic imaging equipment.

A list of organizations represented on this committee can be obtained on request to its secretary.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

© The British Standards Institution 2016.

Published by BSI Standards Limited 2016

ISBN 978 0 580 95743 7

ICS 11.040.55

**Compliance with a British Standard cannot confer immunity from legal obligations.**

This Published Document was published under the authority of the Standards Policy and Strategy Committee on 30 November 2016.

### **Amendments/corrigenda issued since publication**

<b>Date</b>	<b>Text affected</b>
-------------	----------------------

---



# PUBLICLY AVAILABLE SPECIFICATION

## PRE-STANDARD

---

**Good refurbishment practices for medical imaging equipment**

INTERNATIONAL  
ELECTROTECHNICAL  
COMMISSION

---

ICS 11.040.55

ISBN 978-2-8322-3761-8

**Warning! Make sure that you obtained this publication from an authorized distributor.**

## CONTENTS

FOREWORD.....	3
INTRODUCTION.....	5
1 Scope.....	6
2 Normative references .....	6
3 Terms and definitions .....	6
4 General requirements for refurbishment of used medical devices .....	8
4.1 Quality management system .....	8
4.2 Resource management .....	8
4.3 Corrective and preventive action.....	9
4.4 Customer complaints .....	9
4.5 Production and service provision .....	9
4.6 Control of non-conforming product.....	9
4.7 Post-market surveillance process.....	9
4.8 Document control.....	9
4.9 Purchasing.....	9
4.10 Control of design and design changes .....	10
4.11 Risk management process .....	10
5 Specific requirements for good refurbishment practice.....	10
5.1 General.....	10
5.2 Selection of medical imaging equipment for refurbishment.....	10
5.3 Evaluating market access requirements .....	10
5.4 Preparation for refurbishment, disassembly, packing, and shipment.....	10
5.5 Planning .....	11
5.6 Installation of safety updates (hardware/software) .....	11
5.7 Performance and safety test .....	11
5.8 Packing, shipment, and installation of refurbished medical imaging equipment.....	11
5.9 Record of refurbishment.....	11
5.10 Refurbishment label.....	11
Bibliography.....	12

## INTERNATIONAL ELECTROTECHNICAL COMMISSION

---

**GOOD REFURBISHMENT PRACTICES  
FOR MEDICAL IMAGING EQUIPMENT**
**FOREWORD**

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

A PAS is a technical specification not fulfilling the requirements for a standard, but made available to the public.

IEC PAS 63077 has been processed by subcommittee 62B: Diagnostic imaging equipment, of IEC technical committee 62: Electrical equipment in medical practice.

The text of this PAS is based on the following document:

This PAS was approved for publication by the P-members of the committee concerned as indicated in the following document

<b>Draft PAS</b>	<b>Report on voting</b>
62B/1022/PAS	62B/1030/RVC

Following publication of this PAS, which is a pre-standard publication, the technical committee or subcommittee concerned may transform it into an International Standard.

This PAS shall remain valid for an initial maximum period of 3 years starting from the publication date. The validity may be extended for a single period up to a maximum of 3 years, at the end of which it shall be published as another type of normative document, or shall be withdrawn.

## INTRODUCTION

Keeping up with the latest innovations in medical technology often involves replacing equipment in medical practice before it reaches the end of its expected service life. This is because innovation cycles for medical technology are much shorter than the functional lifecycle of capital equipment. As a result, a sustainable resource management model is required: early replacement of installed medical imaging equipment by newer technology is more economically feasible if the residual value of the existing medical imaging equipment is utilized.

Conserving assets is a fundamental principle of ecological thinking in a recycling economy. Several medical imaging equipment companies have already set up quality management systems to refurbish used medical imaging equipment and have delivered this refurbished equipment across the healthcare sector for many years. Refurbishment addresses the high demand for affordable and reliable products. Customers of this used equipment are not only small hospitals with limited budgets but also leading medical institutes. The EU and the US represent by far the two largest markets for refurbished medical equipment. Refurbishment of used medical imaging equipment is a well-established element of the healthcare economy.

If used medical imaging equipment is not accurately maintained according to requirements defined by the manufacturer, it may result in additional risk for patients and operators. Consequently, some countries have imposed bans on the importation of used medical imaging equipment to protect public health. These bans fail to distinguish between quality-assured refurbished medical imaging equipment and second-hand medical imaging equipment of undefined quality, with the effect that patients may be denied access to the safe and economical medical imaging equipment they need.

Safety and effectiveness are the most important aspects to consider with medical imaging equipment, including used equipment. To ensure safety and effectiveness, used medical imaging equipment has to be refurbished in a highly specialized and quality-assured way.

# GOOD REFURBISHMENT PRACTICES FOR MEDICAL IMAGING EQUIPMENT

## 1 Scope

This document describes and defines the process of refurbishment of used medical imaging equipment and applies to the restoring of used medical imaging equipment to a condition of safety and effectiveness comparable to that of new equipment. This restoration includes actions such as repair, rework, software/hardware updates, and the replacement of worn parts with original parts. This document enumerates the actions that must be performed and the manner consistent with product specifications and service procedures required to ensure that the refurbishment of medical imaging equipment is done without changing the finished medical imaging equipment's performance, safety specifications, or intended use according to its original or applicable valid registration.

## 2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 13485:2016, *Medical devices – Quality management systems – Requirements for regulatory purposes*

ISO 14971:2007, *Medical devices – Application of risk management to medical devices*

## 3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at <http://www.electropedia.org/>
- ISO Online browsing platform: available at <http://www.iso.org/obp>

### 3.1

#### **expected service life**

time period specified by the manufacturer during which the medical electrical equipment or medical electrical system is expected to remain safe for use (i.e. maintain basic safety and essential performance)

Note 1 to entry: Maintenance can be necessary during the expected service life.

[SOURCE: IEC 60601-1:2005/AMD1:2012, 3.28]

### 3.2

#### **intended use**

#### **intended purpose**

use for which a product, process, or service is intended according to the specifications, instructions and information provided by the manufacturer



Note 1 to entry: Intended use should not be confused with normal use. While both include the concept of use as intended by the manufacturer, intended use focuses on the medical purpose while normal use incorporates not only the medical purpose, but maintenance, transport, etc. as well.

[SOURCE: IEC 60601-1:2005/AMD1:2012, 3.44]

### 3.3

#### **manufacturer**

natural or legal person with responsibility for the design, manufacture, packaging, labelling, assembling, or adapting medical imaging equipment, regardless of whether these operations are performed by that person or on that person's behalf by a third party

Note 1 to entry: ISO 13485:2016 defines "labelling" as "written, printed or graphic matter as:

- a) affixed to a medical device or any of its containers or wrappers, or
- b) accompanying a medical device,

related to identification, technical description, and use of the medical device, but excluding shipping documents. In this document, that material is described as markings and accompanying document.

Note 2 to entry: Adapting includes making substantial modifications to medical imaging equipment already in use.

Note 3 to entry: In some jurisdictions, the responsible refurbisher can be considered a manufacturer when involved in the activities described.

Note 4 to entry: Adapted from ISO 14971:2007, 2.8.

[SOURCE: IEC 60601-1:2005/AMD1:2012, 3.55, modified – The definition and Note 2 have been reworded.]

### 3.4

#### **medical imaging equipment**

medical electrical equipment that provides images for clinical applications

Note 1 to entry: See IEC 60601-1:2005/AMD1:2012, 3.63 for a definition of medical electrical equipment.

### 3.5

#### **normal use**

operation, including routine inspection and adjustments by any operator, and stand-by, according to the instructions for use

Note 1 to entry: Normal use should not be confused with intended use. While both include the concept of use as intended by the manufacturer, intended use focuses on the medical purpose while normal use incorporates not only the medical purpose, but maintenance, transport, etc. as well.

[SOURCE: IEC 60601-1:2005/AMD1:2012, 3.71]

### 3.6

#### **operator**

person handling the medical imaging equipment

[SOURCE: IEC 60601-1:2005/AMD1:2012, 3.73, modified – The definition has been reworded.]

### 3.7

#### **refurbisher**

natural or legal person who conducts refurbishment of medical imaging equipment

### 3.8

#### **patient**

living being (person or animal) undergoing a medical, surgical or dental procedure

Note 1 to entry: A patient can be an operator.

[SOURCE: IEC 60601-1:2005/AMD1:2012, 3.76]

### **3.9**

#### **process**

set of inter-related or interacting activities which transforms inputs into outputs

[SOURCE: IEC 60601-1:2005/AMD1:2012, 3.89]

### **3.10**

#### **refurbishment**

process or combination of processes applied during the expected service life to restore used medical imaging equipment to a condition of safety and effectiveness comparable to when new

Note 1 to entry: Refurbishment can include activities such as repair, rework, replacement of worn parts, and update of software/hardware but shall not include activities that result in regulatory submissions.

### **3.11**

#### **repair**

means for restoring to a safe, functional, normal condition

[SOURCE: IEC 62353:2014, 3.39]

### **3.12**

#### **rework**

action taken on a nonconforming product so that it will fulfill the specified Device Master Record requirements before it is released for distribution

[SOURCE: 21 CFR 820, 3 (x)]

### **3.13**

#### **risk**

combination of the probability of occurrence of harm and the severity of that harm

[SOURCE: IEC 60601-1:2005/AMD1:2012, 3.102]

### **3.14**

#### **used medical imaging equipment**

medical imaging equipment that has been put into service

## **4 General requirements for refurbishment of used medical devices**

### **4.1 Quality management system**

Refurbishment of used medical imaging equipment shall be conducted under a quality management system (QMS) of the refurbisher in compliance with ISO 13485:2016. In addition to ISO 13485:2016, the provisions in paragraphs 4.2 to 4.11 shall be applied.

### **4.2 Resource management**

The refurbisher shall determine, and provide adequate resources, including trained and qualified personal, maintained and calibrated equipment, and instructions, procedures, files, records, or documents to perform the refurbishment, as well as an environment for refurbishment that is in complete compliance with the applicable environmental, occupational health and safety, and pest control requirements.

#### **4.3 Corrective and preventive action**

The refurbisher shall implement a comprehensive corrective action and preventive action (CAPA) process, addressing the specific aspects of the refurbishment of used medical imaging equipment.

In addition, in the event that the refurbisher identifies, through its CAPA system, safety-related issues that are the responsibility of the original manufacturer and not related to the refurbishment, it shall inform the original manufacturer accordingly.

#### **4.4 Customer complaints**

The refurbisher shall have in place a system for managing complaints.

In addition, the refurbisher shall communicate to the original manufacturer all customer complaints that are not related to the refurbishment of the equipment.

#### **4.5 Production and service provision**

The refurbisher shall have documented procedures for refurbishment and service including but not limited to process validation, disinfection processes, identification, traceability and packaging. In addition the organization shall make provisions to have the knowledge and the ability for installing and servicing medical imaging equipment, or to ensure that servicing can be made available in those markets where the refurbisher makes refurbished medical imaging equipment available on the market.

#### **4.6 Control of non-conforming product**

The refurbisher shall ensure that a product that does not conform to product requirements is identified during refurbishment and controlled to prevent its unintended use or delivery. When a non-conforming product is corrected during refurbishment, it shall be subject to re-verification to demonstrate conformity to the requirements of the original manufacturer.

#### **4.7 Post-market surveillance process**

The refurbisher shall collect feedback from customers and establish documented procedures to notify regulatory authorities of adverse events. The process shall also determine if the adverse event is related to the refurbishment of the used medical imaging equipment or needs to be reported to the original manufacturer.

The refurbisher shall also establish his or her own post-market surveillance process to monitor whether the additional risks resulting from refurbishment have been adequately mitigated.

The refurbisher shall enable monitoring of its installed base of refurbished medical imaging equipment to allow for update management for safety and effectiveness.

#### **4.8 Document control**

The refurbisher shall control all work instructions and procedures used to refurbish medical imaging equipment.

#### **4.9 Purchasing**

The refurbisher shall establish dedicated supplier management capabilities when components or services are purchased.

#### **4.10 Control of design and design changes**

The refurbisher shall review, verify, and validate potential design changes to ensure that the safety and effectiveness requirements of the equipment are not changed from its original or applicable valid registration. All changes, including parts, shall be evaluated to determine if the refurbisher needs registration, as he or she may become the legal manufacturer.

#### **4.11 Risk management process**

The refurbisher shall also establish a risk management process that includes any risk introduced by the refurbishment of used medical imaging equipment. This includes changes that would affect parts.

### **5 Specific requirements for good refurbishment practice**

#### **5.1 General**

The refurbisher shall establish a specific process for the refurbishment of used medical imaging equipment that, in addition to the general requirements described in Clause 4 includes the following specific requirements.

#### **5.2 Selection of medical imaging equipment for refurbishment**

The refurbisher shall determine the criteria that used medical imaging equipment need to meet in order to qualify for refurbishment, based on an assessment of the risk (in accordance with ISO 14971:2007) associated with refurbishment, for any type of medical imaging equipment it wishes to process.

This determination shall consider the following items:

- a) intended use and normal use of the equipment;
- b) expected service life;
- c) applicable standards;
- d) service/maintenance history for the equipment;
- e) existing procedures for the refurbishment of medical imaging equipment, such as service, repair, production, and maintenance.

Used medical imaging equipment that is at the end of expected service life or that cannot be restored to at least the original safety and performance levels, including all mandatory safety updates, shall not be refurbished.

#### **5.3 Evaluating market access requirements**

To ensure regulatory compliance, the refurbisher shall have a process in place to evaluate market access requirements, such as valid registrations and licenses or restrictions, and to provide instructions for use in the appropriate languages, safety information, warnings, and labels.

Refurbished medical imaging equipment for which the registrations or licenses of the original or refurbished medical imaging equipment have been discontinued or where there is no original license may require the refurbisher to obtain a valid registration prior to commercialization.

#### **5.4 Preparation for refurbishment, disassembly, packing, and shipment**

The refurbisher shall have procedures in place to ensure that the medical imaging equipment has been suitably cleaned and disinfected to avoid harming any person involved in the disassembly, packing, and shipment. The medical imaging equipment shall be adequately

disassembled (if necessary) and packed to prevent damage during shipment. Appropriate procedures shall be in place to avoid violation of privacy rules concerning patient data possibly stored on the relevant equipment.

## **5.5 Planning**

A refurbishment plan shall be developed and followed to restore the medical imaging equipment to a condition of safety and effectiveness comparable to when new.

## **5.6 Installation of safety updates (hardware/software)**

The refurbisher shall install all safety updates released by the manufacturer for the relevant medical imaging equipment since it was placed on the market.

## **5.7 Performance and safety test**

Tests specified for the original medical imaging equipment shall be conducted to verify that original performance and safety specifications are met, including all mandatory safety updates.

## **5.8 Packing, shipment, and installation of refurbished medical imaging equipment**

Packing and shipment shall be adequate to prevent damage during transit and load/unload operations. Installation, inspection, and any required testing shall be performed according to documented procedures of the manufacturer.

## **5.9 Record of refurbishment**

The record shall reflect for the relevant medical imaging equipment that all operations and processes described in the refurbishment plan have been accomplished. In addition, the record of refurbishment is specifically required to contain, or refer to the location of, the following information:

- a) date of refurbishment;
- b) any medical imaging equipment identification and control numbers used;
- c) the primary identification label and labelling used for each refurbished unit;
- d) the acceptance records that demonstrate that the equipment has been refurbished in accordance with the refurbishment plan;
- e) list of replaced parts and their identification information.

The record shall authenticate any refurbished medical imaging equipment through means that allow inspection by authorities and verification by customers as requested.

## **5.10 Refurbishment label**

The refurbisher shall label all refurbished medical imaging equipment in proximity to the original label identifying that the medical imaging equipment has been refurbished by the refurbisher and the date of refurbishment.

## Bibliography

IEC 60601-1:2005, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*  
IEC 60601-1:2005/AMD1:2012

IEC 62353:2014, *Medical electrical equipment – Recurrent test and test after repair of medical electrical equipment*

US FDA 21 CFR Part 820 Quality System Regulation, *Title 21 – Food and drugs – Chapter I – Food and drug administration department of health and human services – Subchapter H – Medical devices – Part 820: Quality system regulation*

---



# British Standards Institution (BSI)

BSI is the national body responsible for preparing British Standards and other standards-related publications, information and services.

BSI is incorporated by Royal Charter. British Standards and other standardization products are published by BSI Standards Limited.

## About us

We bring together business, industry, government, consumers, innovators and others to shape their combined experience and expertise into standards-based solutions.

The knowledge embodied in our standards has been carefully assembled in a dependable format and refined through our open consultation process. Organizations of all sizes and across all sectors choose standards to help them achieve their goals.

## Information on standards

We can provide you with the knowledge that your organization needs to succeed. Find out more about British Standards by visiting our website at [bsigroup.com/standards](http://bsigroup.com/standards) or contacting our Customer Services team or Knowledge Centre.

## Buying standards

You can buy and download PDF versions of BSI publications, including British and adopted European and international standards, through our website at [bsigroup.com/shop](http://bsigroup.com/shop), where hard copies can also be purchased.

If you need international and foreign standards from other Standards Development Organizations, hard copies can be ordered from our Customer Services team.

## Copyright in BSI publications

All the content in BSI publications, including British Standards, is the property of and copyrighted by BSI or some person or entity that owns copyright in the information used (such as the international standardization bodies) and has formally licensed such information to BSI for commercial publication and use.

Save for the provisions below, you may not transfer, share or disseminate any portion of the standard to any other person. You may not adapt, distribute, commercially exploit, or publicly display the standard or any portion thereof in any manner whatsoever without BSI's prior written consent.

## Storing and using standards

Standards purchased in soft copy format:

- A British Standard purchased in soft copy format is licensed to a sole named user for personal or internal company use only.
- The standard may be stored on more than 1 device provided that it is accessible by the sole named user only and that only 1 copy is accessed at any one time.
- A single paper copy may be printed for personal or internal company use only.

Standards purchased in hard copy format:

- A British Standard purchased in hard copy format is for personal or internal company use only.
- It may not be further reproduced – in any format – to create an additional copy. This includes scanning of the document.

If you need more than 1 copy of the document, or if you wish to share the document on an internal network, you can save money by choosing a subscription product (see 'Subscriptions').

## Reproducing extracts

For permission to reproduce content from BSI publications contact the BSI Copyright & Licensing team.

## Subscriptions

Our range of subscription services are designed to make using standards easier for you. For further information on our subscription products go to [bsigroup.com/subscriptions](http://bsigroup.com/subscriptions).

With **British Standards Online (BSOL)** you'll have instant access to over 55,000 British and adopted European and international standards from your desktop. It's available 24/7 and is refreshed daily so you'll always be up to date.

You can keep in touch with standards developments and receive substantial discounts on the purchase price of standards, both in single copy and subscription format, by becoming a **BSI Subscribing Member**.

**PLUS** is an updating service exclusive to BSI Subscribing Members. You will automatically receive the latest hard copy of your standards when they're revised or replaced.

To find out more about becoming a BSI Subscribing Member and the benefits of membership, please visit [bsigroup.com/shop](http://bsigroup.com/shop).

With a **Multi-User Network Licence (MUNL)** you are able to host standards publications on your intranet. Licences can cover as few or as many users as you wish. With updates supplied as soon as they're available, you can be sure your documentation is current. For further information, email [subscriptions@bsigroup.com](mailto:subscriptions@bsigroup.com).

## Revisions

Our British Standards and other publications are updated by amendment or revision.

We continually improve the quality of our products and services to benefit your business. If you find an inaccuracy or ambiguity within a British Standard or other BSI publication please inform the Knowledge Centre.

## Useful Contacts

### Customer Services

**Tel:** +44 345 086 9001

**Email (orders):** [orders@bsigroup.com](mailto:orders@bsigroup.com)

**Email (enquiries):** [cservices@bsigroup.com](mailto:cservices@bsigroup.com)

### Subscriptions

**Tel:** +44 345 086 9001

**Email:** [subscriptions@bsigroup.com](mailto:subscriptions@bsigroup.com)

### Knowledge Centre

**Tel:** +44 20 8996 7004

**Email:** [knowledgecentre@bsigroup.com](mailto:knowledgecentre@bsigroup.com)

### Copyright & Licensing

**Tel:** +44 20 8996 7070

**Email:** [copyright@bsigroup.com](mailto:copyright@bsigroup.com)

### BSI Group Headquarters

389 Chiswick High Road London W4 4AL UK