
**Acoustics — Reference zero for the
calibration of audiometric equipment —**

Part 6:

**Reference threshold of hearing for test
signals of short duration**

*Acoustique — Zéro de référence pour l'étalonnage d'équipements
audiométriques —*

*Partie 6: Niveaux liminaires d'audition de référence pour signaux d'essai
de courte durée*



Reference number
ISO 389-6:2007(E)

© ISO 2007

PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.



COPYRIGHT PROTECTED DOCUMENT

© ISO 2007

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 389-6 was prepared by Technical Committee ISO/TC 43, *Acoustics*.

ISO 389 consists of the following parts, under the general title *Acoustics — Reference zero for the calibration of audiometric equipment*:

- *Part 1: Reference equivalent threshold sound pressure levels for pure tones and supra-aural earphones*
- *Part 2: Reference equivalent threshold sound pressure levels for pure tones and insert earphones*
- *Part 3: Reference equivalent threshold force levels for pure tones and bone vibrators*
- *Part 4: Reference levels for narrow-band masking noise*
- *Part 5: Reference equivalent threshold sound pressure levels for pure tones in the frequency range 8 kHz to 16 kHz*
- *Part 6: Reference threshold of hearing for test signals of short duration*
- *Part 7: Reference threshold of hearing under free-field and diffuse-field listening conditions*
- *Part 8: Reference equivalent threshold sound pressure levels for pure tones and circumaural earphones*
- *Part 9: Preferred test conditions for the determination of reference hearing threshold levels*

Introduction

IEC 60645-3 specifies audiometric test signals of short duration. These short-duration signals consist of clicks and tone-bursts, and they are used in different diagnostic tests, e.g. in Evoked Response Audiometry such as auditory electrophysiology (electrocochleography, auditory brainstem responses, etc.), and transient evoked otoacoustic emissions. This part of ISO 389 specifies reference threshold values for the short-duration signals.

The reference hearing threshold data for specific transducers described herein will promote agreement and uniformity in the expression of results of hearing threshold measurements.

Acoustics — Reference zero for the calibration of audiometric equipment —

Part 6: Reference threshold of hearing for test signals of short duration

1 Scope

This part of ISO 389 specifies reference hearing threshold levels for tests signals of short duration applicable to the calibration of audiometric equipment where such signals are used.

NOTE 1 Some notes on different parameters and their effects on threshold levels are given in Annex A.

NOTE 2 Preferred test conditions for determining hearing thresholds for standardization are specified in ISO 389-9.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 389-1, *Acoustics — Reference zero for the calibration of audiometric equipment — Part 1: Reference equivalent threshold sound pressure levels for pure tones and supra-aural earphones*

IEC 60318-1, *Electroacoustics — Simulators of human head and ear — Part 1: Ear simulator for the calibration of supra-aural and circumaural earphones*

IEC 60318-4, *Electroacoustics — Simulators of human head and ear — Part 4: Occluded-ear simulator for the measurement of earphones coupled to the ear by inserts*¹⁾

IEC 60318-6, *Electroacoustics — Simulators of human head and ear — Part 6: Mechanical coupler for the measurements on bone vibrators*²⁾

IEC 61094-1, *Measurement microphones — Part 1: Specifications for laboratory standard microphones*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 389-1 and the following apply.

3.1

short-duration signal

signal having a duration of less than 200 ms

1) Revision of IEC 60711:1981.

2) Revision of IEC 60373:1990.

3.2
click
transient acoustic or vibratory signal whose frequency spectrum covers a broad frequency range, and which is produced by applying a single rectangular pulse to a transducer

3.3
tone-burst
sinusoidal signal multiplied by a time window of less than 200 ms

3.4
condensation signal
short-duration signal, the initial pressure wave of which causes an over-pressure relative to ambient pressure, or the initial force wave of which causes an over-force relative to the static force at the plane of the output port of the transducer

3.5
rarefaction signal
short-duration signal, the initial pressure wave of which causes an under-pressure relative to ambient pressure, or the initial force wave of which causes an under-force relative to the static force at the plane of the output port of the transducer

3.6
alternating polarity signal
series of short-duration signals, alternating between rarefaction and condensation signals

3.7
reference pulse
electric rectangular pulse (single monophasic rectangular wave) of (100 ± 10) μs duration with rise and fall times less than 25 μs as specified in IEC 60645-3:2007, 5.2

NOTE The signal output of the transducer can vary considerably depending upon the type used.

3.8
reference tone-burst
electric signal consisting of five periods of the fundamental sine wave and having a linear rise and fall of 1,6 periods and a duration of three periods as specified in IEC 60645-3:2007, 5.3

NOTE 1,6 periods of linear rise and fall corresponds to two periods from zero to 100 % amplitude. The signal envelope remains at the 100 % amplitude for one period. The reference tone-burst can also be described according to the '2-1-2' concept where '2' indicates the number of periods from zero to 100 % amplitude and back to zero, and '1' is the 100 % amplitude period.

3.9
reference equivalent threshold sound pressure level
RETSPL
at a specified frequency, the median value of the equivalent threshold sound pressure levels of a sufficiently large number of ears of otologically normal persons, of both sexes, aged between 18 years and 25 years inclusive, expressing the threshold of hearing in a specified ear simulator for a specified type of earphone

3.10
reference equivalent threshold vibratory force level, monaural listening
RETVFL
at a specified frequency, the median value of the equivalent threshold vibratory force levels of a sufficiently large number of ears of otologically normal persons, of both sexes, aged between 18 years and 25 years inclusive, expressing the threshold of hearing in a specified mechanical coupler for a specified configuration of bone vibrator

3.11**reference threshold sound pressure level****RTSPL**

at a specified frequency, the median value of the threshold sound pressure levels of a sufficiently large number of otologically normal persons, of both sexes, aged between 18 years and 25 years inclusive, expressing the threshold of hearing under free-field conditions

3.12**peak-to-peak equivalent signal level**

root-mean-square (r.m.s.) value of a long-duration sinusoidal signal which, when compared under the same test conditions with a short-duration output signal from the transducer under test, has the same peak-to-peak value (i.e. difference between the extreme positive and the extreme negative values) as the short-duration signal

NOTE 1 For clicks (3.2), the long-duration sinusoidal signal should have a frequency of 1 000 Hz, and for tone-bursts (3.3) its frequency should equal the fundamental frequency of the tone-burst.

NOTE 2 The term peak-to-peak equivalent signal level is often called peak equivalent signal level.

NOTE 3 The peak level of a short-duration signal is between 3 dB and 9 dB greater than the peak-to-peak equivalent level, i.e. 3 dB when the signal is quite symmetrical around the zero baseline and 9 dB when it is completely on one side of the zero level.

NOTE 4 The recommended abbreviations for peak-to-peak equivalent (pe) sound pressure level and vibratory force level are peSPL and peVFL.

3.13**peak-to-peak equivalent reference equivalent threshold sound pressure level or vibratory force level**
peRETSPL/peRETVFL

peak-to-peak equivalent RETSPL or RETVFL value

NOTE The equivalence in this clause when the full terms (peRETSPL or peRETVFL) are used is between two different signals. When short-duration signals are standardized as sound pressure levels or vibratory force levels for hearing threshold purposes, a further equivalence occurs due to the method of measuring hearing threshold levels in an ear simulator or on a mechanical coupler, respectively, i.e. in terms of equivalent threshold sound pressure or vibratory force levels.

3.14**peak-to-peak equivalent reference threshold sound pressure level****peRTSPL**

peak-to-peak equivalent RTSPL value

4 Specifications

The reference hearing threshold levels depend on the type of test signal (click, tone-burst), the type of transducer (earphone, bone vibrator, loudspeaker) and the equipment used for calibration (ear simulator, free field). The recommended standard values for the click signal based on the reference pulse are given in Tables 1 to 3. They are derived from determinations of the threshold of hearing of otologically normal persons (conditions given in References [4] to [6]).

Similarly, the recommended standard values for the reference tone-burst signals are given in Table 4. They are derived from determinations of the threshold of hearing of otologically normal persons (conditions given in References [6] to [8]).

Table 1 — Recommended peRET SPLs for groups of clicks based on the electric reference pulse with a repetition rate of 20 Hz and a duration of 1 s for the click train sequence — Earphones (monaural hearing)

Type of earphone	Ear simulator used according to	peRET SPL (re 20 µPa) dB ^a
Sennheiser HDA 200	IEC 60318-1	28,0
Maico BERAphone	IEC 60318-1	33,0 ^b
Telephonics TDH-39	IEC 60318-1	31,0
Beyer DT 48	IEC 60318-1	32,0 ^b
Sennheiser HDA 280	IEC 60318-1	31,5 ^b
Etymotic Research ER-2	IEC 60318-4	43,5 ^b
Etymotic Research ER-3A	IEC 60318-4	35,5 ^b
^a Values are rounded to the nearest half decibel.		
^b Experimental data were reported from one laboratory only.		
NOTE The listed median values are based on alternating polarity clicks with 100 µs electric pulse length.		

Table 2 — Recommended peRT SPLs for groups of clicks based on the electric reference pulse with a repetition rate of 20 Hz and a duration of 1 s for the click train sequence — Loudspeakers (binaural hearing)

Type of loudspeaker	Microphone used	peRT SPL (re 20 µPa) dB ^a
KEF RDM1	WS2F according to IEC 61094-1	20,0 ^b
^a Values are rounded to the nearest half decibel.		
^b Experimental data were reported from one laboratory only.		
NOTE 1 The listed median values are based on alternating polarity clicks with 100 µm electric pulse length.		
NOTE 2 Loudspeaker values are only applicable under free-field conditions (ISO 8253-2).		

Table 3 — Recommended peRETVFLs for groups of clicks based on the electric reference pulse with a repetition rate of 20 Hz and a duration of 1 s for the click train sequence — Bone vibrators (without any masking of the non-test ear)

Type of bone vibrator	Mechanical coupler used according to	peRETVFL (re 1 µN) dB ^a
Radioear B-71	IEC 60318-6	51,5 ^b
^a Values are rounded to the nearest half decibel.		
^b Experimental data were reported from one laboratory only.		
NOTE The listed median values are based on alternating polarity clicks with 100 µs electric pulse length.		

Table 4 — Recommended peRETSPLs for groups of reference tone-bursts with a repetition rate of 20 Hz and a duration of 1 s for the tone-burst train sequence — Earphones (monaural hearing, median values)

Earphone (ear simulator)	peRETSPL (re 20 µPa)							
	dB ^a							
	Frequency							
	250 Hz	500 Hz	1 000 Hz	2 000 Hz	3 000 Hz	4 000 Hz	6 000 Hz	8 000 Hz
Sennheiser HDA 200 (ear simulator: IEC 60318-1)	28,0 ^b	21,5	19,5	20,0	22,0*	29,0	38,0 ^b	41,0
Sennheiser HDA 280 (ear simulator: IEC 60318-1)	33,0 ^b	23,5 ^b	21,5 ^b	25,0 ^b	—	29,5 ^b	—	41,0 ^b
Telephonics TDH-39 (ear simulator: IEC 60318-1)	32,0 ^b	23,0	18,5	25,0	25,5 ^b	27,5	36,0 ^b	41,0
Etymotic Research ER-3A (ear simulator: IEC 60318-4)	28,0 ^b	23,5 ^b	21,5 ^b	28,5 ^b	—	32,5 ^b	—	—

^a Values are rounded to the nearest half decibel.

^b Experimental data were reported from one laboratory only.

Annex A (informative)

Information on the correction values for hearing threshold levels with different repetition rates and using head and torso measurements

A.1 Repetition rate, clicks

The values presented in Table A.1 show correction values for hearing threshold levels when repetition rates other than 20 Hz are used. At higher repetition rates the measured threshold levels are lower.

Duration of the click train sequence is 1 s.

Table A.1 — Correction values for the dependence of hearing threshold levels on the repetition rate for trains of clicks based on the reference pulse

Values in decibels

Transducer	Repetition rate					
	10 Hz	20 Hz	40 Hz	60 Hz	80 Hz	100 Hz
Sennheiser HDA 200	1,1	0	– 2,1	– 3,4	– 4,2	– 5,5
Telephonics TDH-39	1,3	0	– 1,7	– 2,9	– 3,2	– 4,9

A.2 Difference between single clicks and 20 Hz repetition rate

The values presented in Table A.2 show correction values for hearing threshold levels when single clicks are used as compared to a repetition rate of 20 Hz. Single clicks have the higher threshold.

Table A.2 — Correction values for hearing thresholds of single clicks compared to click trains (1 s) with a repetition rate of 20 Hz based on the reference pulse

Values in decibels

Transducer	Correction values
Sennheiser HDA 200	8,0
Telephonics TDH-39	9,0

Further information about parameters is found in References [4] to [8].

Annex B (informative)

Head and torso measurements on clicks

Using a specific head and torso simulator (B&K 4128) with soft pinna simulators, maximum differences in reference hearing threshold levels between different transducers are reduced to approximately 4 dB for all earphones and loudspeakers. The preliminary value of $p_{eRET SPL}$ re. 20 μPa with this device is 35,0 dB. See Reference [5]. Future reference equivalent threshold data may be based on this kind of simulator. However, the reference hearing threshold levels for the couplers listed in Table 1 provide less uncertainty.

Annex C (informative)

Notes on the derivation of the reference hearing threshold levels for clicks and tone-bursts

Table C.1 — Investigations of thresholds for clicks based on reference pulse and/or for reference tone-bursts

Investigation	Reference [4]	Reference [5]	Reference [6]	Reference [7]	Reference [8]
Year	1998	2005	2004	2005	2007
Test signals	clicks	clicks	clicks and tone- bursts	tone-bursts	tone-bursts
Repetition rate, Hz	20	20 and 10 to 100	20	20	20
Fundamental frequencies, Hz	—	—	250, 500, 1000, 2000, 3000, 4000, 6000, 8000	500, 1000, 2000, 4000, 8000	250, 500, 1000, 2000, 4000, 8000
Type of transducer	HDA 200	HDA 200, BERAphone, TDH-39,DT 48, HDA 280, ER-2, ER-3A, B-71, RDM 1	HDA 200, TDH-39	HDA 200, TDH-39	HDA 280, ER-3A
Type of ear simulator	IEC 60318-1	IEC 60318-1, IEC 60318-4	IEC 60318-1	IEC 60318-1	IEC 60318-1 IEC 60318-4
Number of subjects	28	25	26	25	25
Number of ears	56	25	52	25	25
Age range	18 years to 25 years	18 years to 25 years	18 years to 25 years	18 years to 25 years	18 years to 25 years
Males/females	15/13	12/13	13/13	13/12	HDA 280:13/12 ER-3A: 15/10
Statistics	median	median	median	median	median

Bibliography

- [1] ISO/CD 389-9, *Acoustics — Reference zero for the calibration of audiometric equipment — Part 9: Preferred test conditions for the determination of reference hearing threshold levels* ³⁾
- [2] ISO 8253-2, *Acoustics — Audiometric test methods — Part 2: Sound field audiometry with pure tone and narrow-band test signals*
- [3] IEC 60645-3:2007, *Electroacoustics — Audiometric equipment — Part 3: Test signals of short duration*
- [4] POULSEN, T. and DAUGAARD, C. Equivalent threshold sound pressure levels for acoustic test signals of short duration. In: *Proceedings of the NATO Advanced Study Institute on Computational Hearing, II Ciocco (Tuscany)*, pp. 245-249, 1998
- [5] RICHTER, U. and FEDTKE, T. Reference zero for the calibration of audiometric equipment using “clicks” as test signals. *International Journal of Audiology*, **44**, pp. 478-87, 2005
- [6] LEGARTH, S. and POULSEN, T. Hearing threshold of signals for evoked response audiometry. Acoustic Technology, Ørsted-DTU, Technical University of Denmark, Lyngby, Denmark, 2004. (*Internal AT Report No. 30.*)
- [7] FEDTKE, T. and RICHTER, U. Bezugsschwellen für audiometrische Kurzzeitsignale — Messungen mit Kurztoenen (Tonbursts). In: *Fortschritte der Akustik, DAGA, München*, pp. 419-20, 2005
- [8] FEDTKE, T. and RICHTER, U. Reference zero for the calibration of air-conduction audiometric equipment using ‘tone bursts’ as test signals. *International Journal of Audiology*, **46**, pp. 1-10, 2007

3) To be published.

ICS 13.140

Price based on 9 pages