
Laboratory glassware — Single-volume pipettes

Verrerie de laboratoire — Pipettes à un volume



Reference number
ISO 648:2008(E)

© ISO 2008

PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

.....



COPYRIGHT PROTECTED DOCUMENT

© ISO 2008

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Contents

Page

Foreword.....	iv
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Basis of adjustment	2
4.1 Unit of volume	2
4.2 Delivery volume	2
4.3 Reference temperature	2
5 Types and classes of accuracy	2
5.1 Classes of accuracy	2
5.2 Types of single-volume pipette	2
6 Maximum permissible errors	3
7 Construction	3
7.1 Material	3
7.2 Shape	3
7.3 Bulb	3
7.4 Dimensions	4
7.5 Top of pipette	5
7.6 Delivery jet	5
7.7 Delivery time	5
7.8 Waiting time	6
8 Graduation line(s)	6
9 Setting of the meniscus	6
10 Marking	6
11 Visibility of graduation lines, figures and inscriptions	7
12 Colour coding	7
Bibliography	8

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 648 was prepared by Technical Committee ISO/TC 48, *Laboratory equipment*, Subcommittee SC 6, *Laboratory and volumetric ware*.

This second edition cancels and replaces the first edition (ISO 648:1977), which has been technically revised to incorporate the following modifications:

- a) single-volume pipettes with two marks have been added;
- b) the waiting time for pipettes Class AS has been reduced to 5 s;
- c) dimensions have been modified;
- d) the nominal volume of 200 ml has been deleted.

Laboratory glassware — Single-volume pipettes

1 Scope

This International Standard specifies metrological and constructional requirements for volumetric pipettes with one mark (total delivery) and for volumetric pipettes with two marks, both of which are adequate for general laboratory purposes.

The details specified are in conformity with the principles of design and construction of volumetric glassware given in ISO 384.

NOTE For graduated pipettes, see ISO 835. For piston-operated pipettes, see ISO 8655-2.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 384:1978, *Laboratory glassware — Principles of design and construction of volumetric glassware*

ISO 719, *Glass — Hydrolytic resistance of glass grains at 98 °C — Method of test and classification*

ISO 1769, *Laboratory glassware — Pipettes — Colour coding*

ISO 3696, *Water for analytical laboratory use — Specification and test methods*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

3.1

delivery volume

volume of liquid discharged from the pipette

NOTE Due to retention of liquid on the inner surface of the volumetric instrument, the volume of liquid delivered is not identical with the volume of liquid contained by the volumetric instrument.

3.2

delivery time

time required for the descent of the liquid meniscus from the graduation line to the point at which the meniscus appears to come to rest in the jet for one-mark pipettes and from the upper graduation line to the lower graduation line for two-mark pipettes

3.3

waiting time

time to be observed after apparent completion of the liquid delivery of the volumetric instrument

NOTE A waiting time applies for Class AS single-volume pipettes (see Clause 5 and 7.8).

4 Basis of adjustment

4.1 Unit of volume

The unit of volume shall be the millilitre (ml), which is equivalent to one cubic centimetre (cm³).

4.2 Delivery volume

Single-volume pipettes shall be adjusted with water of grade 3 quality, in accordance with to ISO 3696, to deliver their nominal volume (Ex). The delivery should be under gravity and unrestricted. For details, see 7.7 and 7.8.

4.3 Reference temperature

The standard reference temperature, i.e. the temperature at which the pipette is intended to deliver its nominal volume, shall be 20 °C.

When the pipette is required for use in a country which has adopted a standard reference temperature of 27 °C (the alternative recommended in ISO 384 for tropical use), this figure shall be substituted for 20 °C.

5 Types and classes of accuracy

5.1 Classes of accuracy

Two classes of accuracy are specified:

- Class A and AS for the higher grade;
- Class B for the lower grade.

5.2 Types of single-volume pipette

The following two types of single-volume pipettes are specified:

- those for which no waiting time is specified (Classes A and B);
- those for which a waiting time of 5 s is specified (Class AS).

6 Maximum permissible errors

Maximum permissible errors in the delivered volume shall not exceed the limits specified in Table 1.

Table 1 — Nominal volumes and maximum permissible errors

Nominal volume ml	Maximum permissible error	
	Class A and AS ± ml	Class B ± ml
0,5	0,005	0,010
1	0,008	0,015
2	0,010	0,02
5	0,015	0,03
10	0,02	0,04
20	0,03	0,06
25	0,03	0,06
50	0,05	0,10
100	0,08	0,15

The relationships between maximum permissible error and nominal volume, as well as between maximum permissible error and meniscus diameter as given in ISO 384:1978, Annex A and Annex B, shall be observed.

For the maximum permissible errors of single-volume pipettes with intermediate nominal volumes between those given in Table 1, the value for the nearest nominal volume shall apply. If the intermediate volume is exactly halfway between those given in Table 1, the value for the next greater nominal volume shall apply.

7 Construction

7.1 Material

Single-volume pipettes shall be manufactured from glass of chemical resistance and thermal properties at least to HGB3 in accordance with ISO 719, shall be as free as possible from visible defects and reasonably free from internal stress.

7.2 Shape

Pipettes of 0,5 ml nominal volume shall consist of a straight tube with a jet at the lower end.

Pipettes of nominal volume 1 ml and 2 ml may be of similar straight tube pattern or may alternatively be of bulb pattern, as described herein for the larger sizes.

All larger pipettes shall consist of a bulb with suction and delivery tubes, the three portions being straight and coaxial.

7.3 Bulb

The main portion of the bulb shall be cylindrical, except that 1 ml and 2 ml nominal-volume pipettes may have an oval bulb if desired. The two ends of the bulb shall merge gradually into the suction and delivery tubes, the junctions being of tapered formation to facilitate drainage.

7.4 Dimensions

Single-volume pipettes shall comply with the dimensional requirements shown in Tables 2, 3 and 4.

Table 2 — Dimensions

Dimensions mm		Nominal volume ml								
		0,5	1	2	5	10	20	25	50	100
Overall length	max.	300	325	350	410	450	520	530	550	600
Internal diameter at the graduation line	max.	2,5	3,5	4,5	4,5	5	5,5	5,5	6	7,5
Length of suction tube — valid only for bulb pattern pipettes	min.	—	150	150	150	160	170	170	170	170
Length of delivery tube — valid only for bulb pattern pipettes	min.	—	110	125	145	160	210	220	230	240
External diameter of delivery tube — valid only for bulb pattern pipettes	±1	—	5	6	6,5	7	7	7	8	9
External diameter of bulb — valid only for bulb pattern pipettes	max.	—	9	9	12	16	22	24	30	38
External diameter of tube — valid only for straight pattern pipettes	max.	5	6	7	—	—	—	—	—	—

Table 3 — Dimensions applying to all nominal volumes of one-mark pipettes

Dimension	Size mm
Distance of graduation line from top of pipette	min. 100
Distance from graduation line to top of bulb ^a	min. 10
Distance from graduation line to tip of jet — valid only for straight pattern pipettes	min. 110
Wall thickness	
— bulb pattern pipettes	min. 0,6
— straight pattern pipettes	min. 0,8
^a Top of bulb relates to the point at which the suction tube begins to expand into the bulb.	

Table 4 — Dimensions applying to all nominal volumes of two-mark pipettes

Dimension	Size mm
Distance of upper graduation line from top of pipette	min. 100
Distance from upper graduation line to top of bulb ^a	min. 10
Distance from upper graduation line to tip of jet — valid only for straight pattern pipettes	min. 110
Distance from lower graduation line to bottom of bulb ^a	min. 20
Distance from lower graduation line to tip of jet	min. 30
Wall thickness	
— bulb pattern pipettes	min. 0,6
— straight pattern pipettes	min. 0,8
^a Top of bulb and bottom of bulb relate to the points at which the suction tube and the delivery tube begin to expand into the bulb.	

7.5 Top of pipette

The top of the single-volume pipette shall lie in a plane at right angles to the axis of the pipette and shall be free from any blemishes which might interfere with the required accurate control by a mechanical device or by the finger in setting the meniscus. The end may be lightly fire-polished or smoothly ground with a slight bevel on the outside.

7.6 Delivery jet

The lower end of the pipette shall terminate either in a delivery jet having a smooth and gradual taper or in a capillary end, both without sudden constriction at the orifice which could give rise to turbulent outflow.

The end of the jet may be finished by grinding or machine tooling and may be fire-polished.

7.7 Delivery time

For testing of the delivery time, the clean pipette shall be held in a vertical position and filled with de-ionized water to a few millimetres above the (upper) graduation line; the falling meniscus shall then be set to this line. Any drop adhering to the jet of the pipette shall be removed by bringing the surface of a glass vessel into contact with the tip of the jet.

Delivery shall then be made into another glass vessel which is slightly inclined so that the tip of the jet is in contact with the inside of the vessel, but without movement of one against the other throughout the delivery period. For one-mark pipettes, delivery shall be made from the graduation line down to the jet of the pipette. For two-mark pipettes, delivery shall be made from the upper graduation line down to the lower graduation line.

IMPORTANT — It is important that a glass receiving vessel be used. Capillary effects influencing the delivery time depend considerably on the material down which the water runs.

The delivery time thus determined, in seconds, shall be within the limits specified in Table 5.

Table 5 — Delivery times in seconds

Nominal volume ml		0,5	1	2	5	10	20	25	50	100
Class A	min.	10	10	10	15	15	25	25	30	40
	max.	20	20	25	30	40	50	50	60	60
Class AS	min.	6	7	7	9	11	12	15	20	25
	max.	10	11	11	13	15	16	20	25	30
Class B	min.	4	5	5	7	8	9	10	13	25
	max.	20	20	25	30	40	50	50	60	60
Maximum permissible difference between observed and marked delivery times	max.	2	2	2	3	3	4	4	5	5

7.8 Waiting time

The waiting time for Class AS pipettes shall be 5 s and shall be marked on the pipette.

8 Graduation line(s)

The graduation line(s) shall be (a) clean, permanent, uniform line(s) completely encircling the tube, although a gap, not exceeding 10 % of the circumference, may be permitted.

The line(s) shall lie in a plane at right angles to the longitudinal axis of the pipette. The line(s) thickness shall not exceed 0,4 mm or one half of the limit of volumetric error, whichever is smaller (see ISO 384).

Dimensions limiting the position(s) of the graduation line(s) on the pipette are included in Tables 3 and 4.

9 Setting of the meniscus

Setting of the meniscus shall be performed according to ISO 384:1978, 5.4.

10 Marking

10.1 The following shall be marked on each single-volume pipette:

- a) a number indicating the nominal volume and the symbol “ml” or the symbol “cm³” to indicate the unit in terms of which the pipette is adjusted;
- b) the inscription “20 °C” to indicate the reference temperature; where the reference temperature is 27 °C, this value shall be substituted for 20 °C;
- c) the letters “Ex” to indicate that the pipette has been adjusted to deliver its indicated nominal volume;
- d) the inscription “A”, “AS” or “B” to indicate the class of accuracy for which the pipette has been adjusted;
- e) the waiting time in the form: “Ex + 5 s”, when appropriate;
- f) the manufacturer's and/or supplier's name or mark.

It is recommended to mark single-volume pipettes complying with this International Standard with the inscription "ISO 648".

10.2 If required for certification purposes, the delivery time, in seconds, shall be marked on Class A and Class AS single-volume pipettes intended for official verification or certification.

10.3 The maximum permissible error according to Table 1 may be marked on all pipettes, for example by the inscription " \pm ... ml".

11 Visibility of graduation lines, figures and inscriptions

All figures, graduation lines and inscriptions shall be permanent, and of such size and form as to be clearly legible under usual conditions of use.

NOTE The permanence of marking can be assessed by the test methods specified in ISO 4794.

12 Colour coding

If colour coding is used on single-volume pipettes, it shall comply with the requirements of ISO 1769.

Bibliography

- [1] ISO 835, *Laboratory glassware — Graduated pipettes*
- [2] ISO 4794, *Laboratory glassware — Methods for assessing the chemical resistance of enamels used for colour coding and colour marking*
- [3] ISO 8655-2, *Piston-operated volumetric apparatus — Part 2: Piston pipettes*

.....

ICS 17.060

Price based on 8 pages