



**INTERNATIONAL STANDARD ISO 3310-1:2000**  
**TECHNICAL CORRIGENDUM 1**

Published 2004-09-01

INTERNATIONAL ORGANIZATION FOR STANDARDIZATION • МЕЖДУНАРОДНАЯ ОРГАНИЗАЦИЯ ПО СТАНДАРТИЗАЦИИ • ORGANISATION INTERNATIONALE DE NORMALISATION

**Test sieves — Technical requirements and testing —**

**Part 1:**  
**Test sieves of metal wire cloth**

**TECHNICAL CORRIGENDUM 1**

*Tamis de contrôle — Exigences techniques et vérifications —*

*Partie 1: Tamis de contrôle en tissus métalliques*

*RECTIFICATIF TECHNIQUE 1*

Technical Corrigendum 1 to ISO 3310-1 was prepared by Technical Committee ISO/TC 24, *Sieves, sieving and other sizing methods*, Subcommittee 1, *Test sieves, test sieving*.

---

*Annex A, page 12, fourth equation*

Replace “ $s = \sqrt{\frac{1}{25} \times 0,049} = 0,045 \text{ mm}$ ” by

$$s = \sqrt{\frac{1}{24} \times 0,049} = 0,045 \text{ mm}$$

*Annex A, page 13, fourth equation*

Replace " $s = \sqrt{\frac{1}{50}} \times 0,183 = 0,061 \text{ mm}$ " by

$$s = \sqrt{\frac{1}{49}} \times 0,183 = 0,061 \text{ mm}$$

*Annex B, page 14, Table B.1*

*Revise the existing note by adding the words "or type of sieving machine" as follows:*

NOTE It is the user's responsibility to ensure that a used sieve or type of sieving machine is re-examined according to the circumstances and with a frequency appropriate to the degree of use that the sieve or type of sieving machine has undergone.