

# INTERNATIONAL STANDARD

**ISO**  
**3823-2**

Second edition  
2003-05-15

**AMENDMENT 1**  
2008-07-01

---

---

## Dentistry — Rotary bur instruments —

### Part 2: Finishing burs

#### AMENDMENT 1

*Art dentaire — Instruments rotatifs de fraisage —*

*Partie 2: Fraises à polir*

*AMENDEMENT 1*

ISO 3823-2:2003/Amd.1:2008(E)



Reference number  
ISO 3823-2:2003/Amd.1:2008(E)

© ISO 2008

**PDF disclaimer**

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.



**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2008

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
Case postale 56 • CH-1211 Geneva 20  
Tel. + 41 22 749 01 11  
Fax + 41 22 749 09 47  
E-mail [copyright@iso.org](mailto:copyright@iso.org)  
Web [www.iso.org](http://www.iso.org)

Published in Switzerland

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

Amendment 1 to ISO 3823-2:2003 was prepared by Technical Committee ISO/TC 106, *Dentistry*, Subcommittee SC 4, *Dental instruments*.

-----

# Dentistry — Rotary bur instruments —

## Part 2: Finishing burs

### AMENDMENT 1

Page 9, 5.3.5.7, Table 7

Change the value for the dimension  $d_3$  from “0,05 mm” to “0,5 mm”.

In the following tables, add the following colour code in the specified lines:

Page 10, 5.3.6.1, Table 8

Table 8: add a column for colour and modify values as follows:

Designation of nominal diameter  (Nominal size)	$d_1$		$d_2$	$l_1$	Number of blades  min.	Colour	$l_2$ $\pm 0,5$			
	nom.	tol.	max.	min.			Shank Type 1 Standard	Shank Type 2 Standard	Shank Type 3 Standard	Shank Type 3 Short
018	1,8	$\pm 0,08$	1,20	1,35	30	white	22,0	44,5	19,0	16,5

Page 13, 5.3.6.4.1, Table 11

Table 11: add a column for colour and modify values as follows:

Designation of nominal diameter  (Nominal size)	$d_1$		$d_2$	$l_1$	Number of blades  min.	Colour	$l_2$ $\pm 0,5$			
	nom.	tol.	max.	min.			Shank Type 1 Standard	Shank Type 2 Standard	Shank Type 3 Standard	Shank Type 3 Short
009	0,9	$\pm 0,05$	0,80	3,3	30	white	22,0	—	19,0	—

Table 12: add a column for colour and modify values as follows:

Designation of nominal diameter (Nominal size)	$d_1$		$d_2$ max.	$l_1$ min.	Number of blades min.	Colour	$l_2$ $\pm 0,5$			
	nom.	tol.					Shank Type 1 Standard	Shank Type 2 Standard	Shank Type 3 Standard	Shank Type 3 Short
012	1,2	$\pm 0,08$	1,2	8,0	20	yellow	—	—	23,0	—
					30	white				

Table 13: add a column for colour and modify values as follows:

Designation of nominal diameter (Nominal size)	$d_1$		$d_2$ max.	$l_1$ min.	Number of blades min.	Colour	$l_2$ $\pm 0,5$			
	nom.	tol.					Shank Type 1 Standard	Shank Type 2 Standard	Shank Type 3 Standard	Shank Type 3 Short
014	1,4	$\pm 0,08$	1,10	2,6	30	white	22,0	—	19,0	—
018	1,8		1,35	3,0	30	white				
023	2,3		1,45	3,3	20	yellow				
		30	white							

Table 16: add a column for colour and modify values as follows:

Designation of nominal diameter (Nominal size)	$d_1$		$d_2$ max.	$l_1$ min.	Number of blades min.	Colour	$l_2$ $\pm 0,5$			
	nom.	tol.					Shank Type 1 Standard	Shank Type 2 Standard	Shank Type 3 Standard	Shank Type 3 Short
007	0,7	$\pm 0,05$	0,68	3,15	20	yellow	22,0	—	19,0	—
					30	white				
009	0,9		0,80	3,15	20	yellow				
			30		white					

Page 22, 5.3.6.9, Table 20


Table 20: add a column for colour and modify values as follows:

Designation of nominal diameter (Nominal size)	$d_1$		$d_2$	$d_3$	$l_1$	Number of blades min.	Colour	$l_2$ $\pm 0,5$			
	nom.	tol.	max.	max.	min.			Shank Type 1 Standard	Shank Type 2 Standard	Shank Type 3 Standard	Shank Type 3 Short
010	1,0	$\pm 0,08$	1,0	0,5	4,5	20	yellow	22,0	—	19,0	—
						30	white				

Page 32, 5.3.6.16

After 5.3.6.16, add the following subclause, figure and table:

**5.3.6.17 Conical, pointed**



$\alpha = 6^\circ$  to  $12^\circ$

NOTE All instruments should have a smooth, rounded, non-cutting tip.

**Figure 35 — Conical, pointed, carbide finishing burs**

Table 34 a) — Conical, pointed, carbide finishing burs: dimensions and number of blades

Designation of nominal diameter (Nominal size)	$d_1$		$d_2$	$l_1$	Number of blades min.	Colour	$l_2$ $\pm 0,5$			
	nom.	tol.	max.	min.			Shank Type 1 Standard	Shank Type 2 Standard	Shank Type 3 Standard	Shank Type 3 Short
008	0,8	$\pm 0,05$	0,70	3,0	8	red	25,0	—	23,0	—
					16	yellow				
					30	white				
010	1,0		0,85	4,2	8	red	—	—	20,0	—
					16	yellow				
					30	white				
014	1,4	$\pm 0,08$	1,10	6,0	8	red	23,0	—	21,0	—
					16	yellow				
					30	white				
014	1,4		1,20	9,0	8	red	27,0	—	24,0	—
					16	yellow				
					30	white				

Page 33, 5.6

After 5.6, add the following subclause:

**5.7 Colour coding of carbide instruments**

Colour coding is provided for carbide instruments. The colour code complements the designation and is commonly used on the instrument itself. The usage of colour coding is optional, at the discretion of the manufacturer. If colour coding is used, the colours shall be those specified in this Amendment [Table 8, Table 11, Table 12, Table 13, Table 16, Table 20, Table 34 a)].

All carbide finishing burs are marked with a colour code at the manufacturer's discretion. Standard finishing burs (fine) shall have a red colour code. Finer colour codes are specified in this Amendment.

Size	Colour
fine	red
extra fine	yellow
ultra fine	white

Page 34, 6.4.2

Replace "Quality 3" by "Grade 3".





---

---

**ICS 11.060.20**

Price based on 4 pages