

INTERNATIONAL STANDARD

ISO 4823

Third edition
2000-12-15

AMENDMENT 1
2007-07-01

Dentistry — Elastomeric impression materials

AMENDMENT 1

Art dentaire — Produits pour empreintes, à base d'élastomères
AMENDEMENT 1



Reference number
ISO 4823:2000/Amd.1:2007(E)

© ISO 2007

PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.



COPYRIGHT PROTECTED DOCUMENT

© ISO 2007

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

Amendment 1 to ISO 4823:2000 was prepared by Technical Committee ISO/TC 106, *Dentistry*, Subcommittee SC 2, *Prosthetic materials*.

Dentistry — Elastomeric impression materials

AMENDMENT 1

This amendment to ISO 4823-2000 exempts the Type 0 (putty) elastomeric impression materials from compliance with the working time requirement specified in 6.3.

Page 3, subclause 6.3

Subclause 6.3 is restructured as shown below to clarify matters pertinent to this exemption.

6.3 Working time (Types 1, 2 and 3 only)

When tested in accordance with 9.3, the working time shall not be less than that stated in the manufacturer's instructions [10.3 f)] and shall be at least 30 s longer than the time required to obtain a homogeneous mixture of the components (see 6.2 and 9.1).

NOTE Currently, Type 0 materials are exempt from compliance with the working time requirement for the following reasons:

- because some of these materials, which are otherwise satisfactory for the purpose intended, have at least one component which, even before being combined in a mixture, exhibit an inherent elasticity behaviour that will not permit the mixed material to provide for objectively generated working time data when tested in accordance with 9.3

and

- because as of now, extensive inter-laboratory testing has not resulted in discovery of reliable ways in which the currently prescribed procedures can be modified to provide for objective working time testing of the materials exhibiting inherent elasticity behaviour; nor has such testing led to the development of an effective new approach for solving the problem.

ICS 11.060.10

Price based on 1 page