



**INTERNATIONAL STANDARD ISO 5832-12:2007**  
**TECHNICAL CORRIGENDUM 1**

Published 2008-09-15

INTERNATIONAL ORGANIZATION FOR STANDARDIZATION • МЕЖДУНАРОДНАЯ ОРГАНИЗАЦИЯ ПО СТАНДАРТИЗАЦИИ • ORGANISATION INTERNATIONALE DE NORMALISATION

**Implants for surgery — Metallic materials —**  
**Part 12:**  
**Wrought cobalt-chromium-molybdenum alloy**

TECHNICAL CORRIGENDUM 1

*Implants chirurgicaux — Matériaux métalliques —*

*Partie 12: Alliage corroyé à base de cobalt, de chrome et de molybdène*

*RECTIFICATIF TECHNIQUE 1*

Technical Corrigendum 1 to ISO 5832-12:2007 was prepared by Technical Committee ISO/TC 150, *Implants for surgery*, Subcommittee SC 1, *Materials*.

---

*Page 2, Table 2, column 2, row 4*

The tensile strength for warm worked material should read 1 172 MPa.