



# Dentistry — Number coding system for rotary instruments — Part 1: General characteristics

## TECHNICAL CORRIGENDUM 1

*Art dentaire — Système de codification numérique pour instruments rotatifs —*

*Partie 1: Caractéristiques générales*

*RECTIFICATIF TECHNIQUE 1*

Technical Corrigendum 1 to ISO 6360-1:2004 was prepared by Technical Committee ISO/TC 106, *Dentistry*, Subcommittee SC 4, *Dental instruments*.

---

*Page 14*

### **5.5 Overall length**

Paragraph 4 shall read:

For hand-held root-canal instruments, the code number for the sixth digit shall be 0.

A new paragraph 5 is added:

For powered (electrical) root-canal instruments of Type 1 and Type 2 of ISO 1797-1:1992 and ISO 1792-2:1992, the codes of Table 5 are used:

2 = miniature      3 = short      4 = standard      5 = long      6 = extra long