

Second edition
2011-07-01

AMENDMENT 1
2016-08-15

**Implants for surgery — Components
for partial and total knee joint
prostheses —**

Part 2:

**Articulating surfaces made of metal,
ceramic and plastics materials**

AMENDMENT 1

*Implants chirurgicaux — Éléments de prothèses partielle et totale de
l'articulation du genou —*

*Partie 2: Surfaces articulaires constituées de matériaux métalliques,
céramiques et plastiques*

AMENDEMENT 1



COPYRIGHT PROTECTED DOCUMENT

© ISO 2016, Published in Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Ch. de Blandonnet 8 • CP 401
CH-1214 Vernier, Geneva, Switzerland
Tel. +41 22 749 01 11
Fax +41 22 749 09 47
copyright@iso.org
www.iso.org

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: www.iso.org/iso/foreword.html.

The committee responsible for this document is ISO/TC 150, *Implants for surgery*, Subcommittee SC 4, *Bone and joint replacements*.

Implants for surgery — Components for partial and total knee joint prostheses —

Part 2: Articulating surfaces made of metal, ceramic and plastics materials

AMENDMENT 1

Page 1, Clause 2

Add the following normative reference and delete it in the Bibliography:

“ISO 7207-1, Implants for surgery — Components for partial and total knee joint prostheses — Part 1: Classification, definitions and designation of dimensions”

Page 1, 3.1

Replace the subclause with the following:

“The principles for the surface finish measurements of knee joint prostheses components are given in ISO 4287. The measurements shall be performed according to the rules and procedures described in ISO 4288.”

Page 1, 3.2.1

Replace the first paragraph with the following:

“When measured in accordance with ISO 4288, all articulating surfaces of a metallic or ceramic femoral component shall be measured across the full articulating surface at locations in an approximately square grid of locations no more than 10 mm apart. The component shall have an R_{max} value $\leq 0,1 \mu\text{m}$, using a cut-off value of 0,25 mm.”

Page 2, 3.2.2

Replace the first paragraph with the following:

“When measured in accordance with ISO 4288, all articulating surfaces of a metallic or ceramic tibial component shall be measured across the full articulating surface at locations in an approximately square grid of locations no more than 10 mm apart. The component shall have an R_{max} value $\leq 0,1 \mu\text{m}$, using a cut-off value of 0,25 mm.”

Page 2, 3.2.3

Replace the first paragraph with the following:

“When measured in accordance with ISO 4288, all articulating surfaces of a tibial and patella component shall be measured across the full articulating surface at locations in an approximately square grid of locations no more than 10 mm apart. The component shall have an R_{max} value $\leq 2 \mu\text{m}$, using a cut-off value of 0,8 mm.

The following details should be reported along with the measured surface roughness, R_a , values:”

