

Second edition
2008-12-15

AMENDMENT 1
2013-07-15

Dentistry —
Evaluation of biocompatibility of
medical devices used in dentistry
AMENDMENT 1: Positive control
material

*Art dentaire — Évaluation de la biocompatibilité des dispositifs
médicaux utilisés en art dentaire*
AMENDEMENT 1: Matériaux de contrôle positif



Reference number
ISO 7405:2008/Amd.1:2013(E)

© ISO 2013



COPYRIGHT PROTECTED DOCUMENT

© ISO 2013

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

© ISO 2013 – All rights reserved

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

Amendment 1 to ISO 7405:2008 was prepared by Technical Committee ISO/TC 106, *Dentistry*.

.....

Dentistry — Evaluation of biocompatibility of medical devices used in dentistry

AMENDMENT 1: Positive control material

Title page

Replace the French title with the following:

Médecine bucco-dentaire — Évaluation de la biocompatibilité des dispositifs médicaux utilisés en médecine bucco-dentaire

Page 7, 6.1.4

Add the following new subclause:

6.1.5 Positive control material

For both *in vitro* and *in vivo* tests, it is advisable to include a standard positive control material, which is handled and processed like the test materials (i.e. being plastic after mixing and then setting) and which is based on freely available chemicals or materials. Such a positive control material for *in vitro* testing of plastic filling materials is described in Annex B, Table B.1. The use of this specific positive control material is optional and other materials with a validated history and other well characterized positive control materials with reproducible data on toxicity can be used instead.

Page 30, Annex B, Table B.1

Add the following note in a footer row at the bottom of the table:

NOTE See Reference [31].

Page 34, Bibliography

Add the following reference:

- [31] SCHMALZ, G., HILLER, K.-A., SEIDENADER, C. and SCHWEIKL, H., *Interlaboratory testing of a new cytotoxic reference dental restorative material*. J Dent Res., **90**, Spec Issue A, 2011 (www.dentalresearch.org), Abstract # 273, <http://iadr.confex.com/iadr/2011sandiego/webprogram/schedule/Paper146565.html>.

ICS 11.060.10; 11.100.99

Price based on 1 page