

First edition
2004-10-01

AMENDMENT 1
2012-02-15

**Mechanical contraceptives — Reusable
natural and silicone rubber contraceptive
diaphragms — Requirements and tests**

AMENDMENT 1

*Contraceptifs mécaniques — Diaphragmes contraceptifs réutilisables
en caoutchouc — Performances et essais*

AMENDEMENT 1



Reference number
ISO 8009:2004/Amd.1:2012(E)

© ISO 2012



COPYRIGHT PROTECTED DOCUMENT

© ISO 2012

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

Amendment 1 to ISO 8009:2004 was prepared by Technical Committee ISO/TC 157, *Non-systemic contraceptives and STI barrier prophylactics*.

.....

Mechanical contraceptives — Reusable natural and silicone rubber contraceptive diaphragms — Requirements and tests

AMENDMENT 1

Page 1, Clause 2

Add the following reference:

ISO 10993-1, *Biological evaluation of medical devices — Part 1: Evaluation and testing within a risk management process*

Page 2, Clause 6

Replace the second, third and fourth paragraphs with the following:

For any new product, or following a significant change to the formulation or manufacturing process, biocompatibility assessments shall be conducted in accordance with ISO 10993-1. Testing for cytotoxicity in accordance with ISO 10993-5, and for irritation and sensitization in accordance with ISO 10993-10, shall be conducted. Spermicides applied at the time of use are exempt from this requirement. Where practicable, manufacturers should take steps to recommend spermicides that minimize irritant effects. Accredited laboratories shall be used for all biocompatibility testing. Regulatory bodies might also specify local requirements and require results to be interpreted by a qualified toxicologist. Any toxicologist's assessment report shall state that the product is safe under normal conditions of use.

Page 26, Bibliography

Delete ISO 10993-1 from the list.

Renumber the subsequent references accordingly.

ICS 11.200.00

Price based on 1 page