



INTERNATIONAL STANDARD ISO 8655-6:2002
TECHNICAL CORRIGENDUM 1

Published 2008-12-15

INTERNATIONAL ORGANIZATION FOR STANDARDIZATION • МЕЖДУНАРОДНАЯ ОРГАНИЗАЦИЯ ПО СТАНДАРТИЗАЦИИ • ORGANISATION INTERNATIONALE DE NORMALISATION

Piston-operated volumetric apparatus —
Part 6:
Gravimetric methods for the determination of measurement error

TECHNICAL CORRIGENDUM 1

Appareils volumétriques à piston —

Partie 6: Méthodes gravimétriques pour la détermination de l'erreur de mesure

RECTIFICATIF TECHNIQUE 1

Technical Corrigendum 1 to ISO 8655-6:2002 was prepared by Technical Committee ISO/TC 48, *Laboratory equipment*, Subcommittee SC 6, *Laboratory and volumetric ware*.

Page iv, Foreword

Add the following part to the list of parts:

— *Part 7: Non-gravimetric methods for the assessment of equipment performance*

Delete “The following part is under preparation:” and the former title of part 7.

Page 1, Scope

In the note, replace the third sentence with the following:

“Alternative test methods such as photometric and titrimetric methods are given in ISO 8655-7.”

ICS 17.060

Ref. No. ISO 8655-6:2002/Cor.1:2008(E)

© ISO 2008 – All rights reserved

Published in Switzerland

Page 1, Clause 2

Replace the reference to OIML R 76-1:1992 with the following:

OIML R 76-1, *Non-automatic weighing instruments — Part 1: Metrological and technical requirements — Tests*

Page 9, 8.3

In the sentence below Equation (2), replace “which is 20 °C” with “which is 20 °C or 27 °C”.

Page 10, Clause 9

In list item b), replace “(20 °C)” with “(20 °C or 27 °C)” to give the following:

- b) basis of adjustment (Ex) or (In) and reference temperature (20 °C or 27 °C);