



**INTERNATIONAL STANDARD ISO 8980-1:2004**  
**TECHNICAL CORRIGENDUM 1**

Published 2006-08-01

INTERNATIONAL ORGANIZATION FOR STANDARDIZATION • МЕЖДУНАРОДНАЯ ОРГАНИЗАЦИЯ ПО СТАНДАРТИЗАЦИИ • ORGANISATION INTERNATIONALE DE NORMALISATION

**Ophthalmic optics — Uncut finished spectacle lenses —**  
**Part 1:**  
**Specifications for single-vision and multifocal lenses**

TECHNICAL CORRIGENDUM 1

*Optique ophtalmique— Verres de lunettes finis non détourés —*  
*Partie 1: Spécifications pour les verres unifocaux et multifocaux*

*RECTIFICATIF TECHNIQUE 1*

Technical Corrigendum 1 to ISO 8980-1:2004 was prepared by Technical Committee ISO/TC 172, *Optics and photonics*, Subcommittee SC 7, *Ophthalmic optics and instruments*.

---

*Page 4, Table 4*

Correct the note, as follows:

NOTE  $S_{\max}$  is the absolute value of the focal power, in dioptries, in the meridian of higher absolute power.