



INTERNATIONAL STANDARD ISO 10139-1:2005
TECHNICAL CORRIGENDUM 1

Published 2006-03-01

INTERNATIONAL ORGANIZATION FOR STANDARDIZATION • МЕЖДУНАРОДНАЯ ОРГАНИЗАЦИЯ ПО СТАНДАРТИЗАЦИИ • ORGANISATION INTERNATIONALE DE NORMALISATION

Dentistry — Soft lining materials for removable dentures —
Part 1
Materials for short-term use

TECHNICAL CORRIGENDUM 1

Art dentaire — Produits souples pour intrados de prothèses dentaires amovibles —

Partie 1: Produits pour usage à court terme

RECTIFICATIF TECHNIQUE 1

Technical Corrigendum 1 to ISO 10139-1:2005 was prepared by Technical Committee ISO/TC 106, *Dentistry*, Subcommittee SC 2, *Prosthetic materials*.

Page 5

7.3.1.2

replace “**Three beakers**, of 100 ml capacity”

by “**Three beakers**, each of sufficient size that 100 ml of water covers the specimen (see 7.3.3)”.

Page 8

7.3.3 (on the page, paragraph 1, line 3)

delete “100 ml” before “beaker”.

Page 9

8.3 d)

after “properly” insert the text “and the time at which 10 % recovery is achieved (see Introduction and 7.2.4)”.