

Second edition
2008-09-01

AMENDMENT 1
2012-06-15

Single-use medical examination gloves —

Part 1:

**Specification for gloves made from
rubber latex or rubber solution**

AMENDMENT 1

Gants en caoutchouc pour examen, non réutilisables —

*Partie 1: Spécifications pour gants fabriqués à partir de latex de
caoutchouc ou d'une solution de caoutchouc*

AMENDEMENT 1



Reference number
ISO 11193-1:2008/Amd.1:2012(E)

© ISO 2012



COPYRIGHT PROTECTED DOCUMENT

© ISO 2012

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

Amendment 1 to ISO 11193-1:2008 was prepared by Technical Committee ISO/TC 45, *Rubber and rubber products*, Subcommittee SC 4, *Products (other than hoses)*.

Single-use medical examination gloves —

Part 1:

Specification for gloves made from rubber latex or rubber solution

AMENDMENT 1

Page 6, Subclause 6.3.3

Add the following paragraph after Table 3:

Testing shall be conducted within two months of the date of receipt by the purchaser of the gloves. For gloves that are older than 3 months from the date of manufacture or for which the date of manufacture is not known, the tensile properties need only conform to the “after accelerated ageing” values in Table 3.

Page 6, Table 3

In the last column of the table (Type 2 glove), change the value of the minimum force at break after accelerated ageing from 7,0 N to 6,0 N.

ICS 11.140; 83.140.99

Price based on 1 page