



INTERNATIONAL STANDARD ISO 11979-2:1999
TECHNICAL CORRIGENDUM 1

Published 2003-11-01

INTERNATIONAL ORGANIZATION FOR STANDARDIZATION • МЕЖДУНАРОДНАЯ ОРГАНИЗАЦИЯ ПО СТАНДАРТИЗАЦИИ • ORGANISATION INTERNATIONALE DE NORMALISATION

Ophthalmic implants — Intraocular lenses —
Part 2:
Optical properties and test methods

TECHNICAL CORRIGENDUM 1

Implants ophtalmiques — Lentilles intraoculaires —

Partie 2: Propriétés optiques et méthodes d'essai

RECTIFICATIF TECHNIQUE 1

Technical Corrigendum 1 to ISO 11979-2:1999 was prepared by Technical Committee ISO/TC 172, *Optics and optical instruments*, Subcommittee SC 7, *Ophthalmic optics and instruments*

Page 3, Subclause 4.4

Replace the first sentence by the following:

For each type of IOL, the spectral transmittance in the range 300 nm to 1 100 nm shall be on record for the IOL with a dioptric power of 20 D or its equivalent.

ICS 11.040.70

Ref. No. ISO 11979-2:1999/Cor.1:2003(E)

© ISO 2003 – All rights reserved

Published in Switzerland