



INTERNATIONAL STANDARD ISO 13897:2003
TECHNICAL CORRIGENDUM 1

Published 2003-12-15

INTERNATIONAL ORGANIZATION FOR STANDARDIZATION • МЕЖДУНАРОДНАЯ ОРГАНИЗАЦИЯ ПО СТАНДАРТИЗАЦИИ • ORGANISATION INTERNATIONALE DE NORMALISATION

Dentistry — Amalgam capsules

TECHNICAL CORRIGENDUM 1

Art dentaire — Capsules pour amalgame

RECTIFICATIF TECHNIQUE 1

Technical Corrigendum 1 to ISO 13897:2003 was prepared by Technical Committee ISO/TC 106, *Dentistry*, Subcommittee SC 6, *Dental equipment*.

Page 4

7.3.5.2 Type 2 capsules

Replace the first sentence with the following new sentence.

“If one of the five tests of any of the capsules fails, test 10 more capsules five times each.”