

INTERNATIONAL
STANDARD

ISO
14490-5

First edition
2005-10-15
AMENDMENT 1
2015-02-01

**Optics and optical instruments — Test
methods for telescopic systems —**

Part 5:
Test methods for transmittance

AMENDMENT 1

*Optique et instruments d'optique — Méthodes d'essai pour systèmes
télescopiques —*

Partie 5: Méthodes d'essai du facteur de transmission

AMENDEMENT 1



Reference number
ISO 14490-5:2005/Amd.1:2015(E)

© ISO 2015



COPYRIGHT PROTECTED DOCUMENT

© ISO 2015

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the WTO principles in the Technical Barriers to Trade (TBT) see the following URL: [Foreword - Supplementary information](#)

The committee responsible for this document is ISO/TC 172, *Optics and photonics*, Subcommittee SC 4, *Telescopic systems*.

Optics and optical instruments — Test methods for telescopic systems —

Part 5: Test methods for transmittance

AMENDMENT 1

Update references to ISO/CIE 10526 and ISO/CIE 10527. This applies to Clause 2, the Bibliography, and all instances ISO/CIE 10526 and ISO/CIE 10527 are referenced.

ISO/CIE 10526 has been cancelled and replaced by ISO 11664-2:

ISO 11664-2:2007, *Colorimetry — Part 2: CIE standard illuminants*

ISO/CIE 10527 has been cancelled and replaced by ISO 11664-1:

ISO 11664-1:2007, *Colorimetry — Part 1: CIE standard colorimetric observers*

In addition, add to the Bibliography:

ISO 11664-3:2012, *Colorimetry — Part 3: CIE tristimulus values*

ISO 11664-4:2008, *Colorimetry — Part 4: CIE 1976 L*a*b* Colour space*

ISO 11664-5:2009, *Colorimetry — Part 5: CIE 1976 L*u*v* Colour space and u', v' uniform chromaticity scale diagram*

Page 3, Subclause 5.5

Add the following paragraphs:

Generally, the smallest possible aperture stop should be used which is compatible with the signal-to-noise requirements of the detector.

Special care should be taken when measuring telescopic systems with variable magnification where at some magnification settings the entrance pupil can be considerably smaller than the free objective lens diameter. In this case, it is recommended to take the entrance pupil as the “maximum available aperture”.

Page 3, Subclause 5.6

Add the following paragraph:

The test specimen should be oriented in a way such that no obstructions occur in the measurement beam (e.g. by reticle structures).

Add the following new Clause B.3: CIELAB values

B.3 CIELAB values

This clause is a description of an alternative method for the determination of a colour hue (tint) of the visible image.

To get a reasonable idea of the colour hue of the test specimen's image, it is also possible to use the CIELAB colour space to determine the values L^* , a^* , and b^* . The value L^* is not necessary for this method.

CIELAB is a colour space with perceptually uniformly spaced colour distribution.

The normalized standardized spectral values X_n , Y_n , and Z_n for the white point of D 65 and the 10° observer are taken from [Table B.3](#).

Table B.3 — Normalized standardized spectral values $X_{10W} = X_n$; $Y_{10W} = Y_n$; $Z_{10W} = Z_n$

Standard illuminant	x_{10}	y_{10}	z_{10}	X_{10W}	Y_{10W}	Z_{10W}
D 65	0,313 8	0,331 0	0,355 2	94,81	100,00	107,34

From the spectral transmittance the values X , Y , and Z , according to ISO 11664-1, are calculated. With these and the standardized spectral values, the following values X^* , Y^* , and Z^* are calculated.

$$X^* = \sqrt[3]{X / X_n} \quad \text{for } X/X_n > (6/29)^3$$

$$X^* = (841 / 108) (X / X_n) + 4 / 29 \quad \text{for } X/X_n \leq (6/29)^3$$

$$Y^* = \sqrt[3]{Y / Y_n} \quad \text{for } Y/Y_n > (6/29)^3$$

$$Y^* = (841 / 108) (Y / Y_n) + 4 / 29 \quad \text{for } Y/Y_n \leq (6/29)^3$$

$$Z^* = \sqrt[3]{Z / Z_n} \quad \text{for } Z/Z_n > (6/29)^3$$

$$Z^* = (841 / 108) (Z / Z_n) + 4 / 29 \quad \text{for } Z/Z_n \leq (6/29)^3$$

With X^* , Y^* , and Z^* , the following values a^* and b^* are calculated.

Coordinate a^* $a^* = 500(X^* - Y^*)$

Coordinate b^* $b^* = 200(Y^* - Z^*)$

Finally the chromaticity (tint value) C^*_{ab} is calculated according to the following equation: (distance from the white point)

$$C^*_{ab} = \sqrt{a^{*2} + b^{*2}}$$

To determine the tint the signs of a^* , b^* , and the amount $|b^*/a^*|$ are used. The names of the colours are taken from the [Table B.4](#).

Table B.4 — Colour naming conventions - tint name

		a^*		$ b^*/a^* $
		-	+	
b^*	+	yellow (Y)		>2,5
		yellow green (YG)	orange (O)	0,4 to 2,5
		green (G)	red (R)	<0,4
	-	green (G)	red (R)	<0,4
		blue green (BG)	purple (P)	0,4 to 2,5
		blue (B)		>2,5

