

---

---

**Microscopes — Designation of  
microscope objectives —**

**Part 1:  
Flatness of field/Plan**

*Microscopes — Désignation des objectifs de microscope —  
Partie 1: Planéité du champ/Plan*





**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2013

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
Case postale 56 • CH-1211 Geneva 20  
Tel. + 41 22 749 01 11  
Fax + 41 22 749 09 47  
E-mail [copyright@iso.org](mailto:copyright@iso.org)  
Web [www.iso.org](http://www.iso.org)

Published in Switzerland



# Contents

|          |   | Page      |
|----------|---|-----------|
|          | <b>Foreword</b> .....   | <b>iv</b> |
| <b>1</b> | <b>Scope</b> .....  | <b>1</b>  |
| <b>2</b> | <b>Normative references</b> .....   | <b>1</b>  |
| <b>3</b> | <b>Terms and definitions</b> .....  | <b>1</b>  |
| <b>4</b> | <b>Requirements</b> .....   | <b>2</b>  |
| 4.1      | Indication .....  | 2         |
| 4.2      | Definition of plan objectives .....   | 2         |
| 4.3      | Determination of plan field number .....  | 2         |
|          | <b>Annex A (informative) Depth of field in object space calculated by Berek's formula</b> ..... | <b>4</b>  |
|          | <b>Bibliography</b> .....   | <b>7</b>  |

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2. [www.iso.org/directives](http://www.iso.org/directives)

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received. [www.iso.org/patents](http://www.iso.org/patents)

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

The committee responsible for this document is ISO/TC 172, *Optics and photonics*, Subcommittee SC 5, *Microscopes and endoscopes*.

This third edition cancels and replaces the second edition (ISO 19012-1:2011) by the addition of “Petzval curvature”.

ISO 19012 consists of the following parts, under the general title *Microscopes — Designation of microscope objectives*:

- *Part 1: Flatness of field/Plan*
- *Part 2: Chromatic correction*

The following parts are under preparation:

- *Part 3: Spectral transmittance*

# Microscopes — Designation of microscope objectives —

## Part 1: Flatness of field/Plan

### 1 Scope

This part of ISO 19012 specifies the use of the marking “Plan” on microscope objectives, and defines the diameter of the sharp region of the primary image of a flat object surface. This part of ISO 19012 applies to visual observation using the combination of objective lens, tube lens and eyepiece, as specified by the manufacturer.

This marking is consistent with ISO 8578.

**NOTE** The flatness of the image field does not imply any degree of correction for other aberrations (ISO 10934-1).

### 2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 10934-1, *Optics and optical instruments — Vocabulary for microscopy — Part 1: Light microscopy*

### 3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 10934-1 and the following apply.

#### 3.1

##### **tangential structured object**

object containing short lines perpendicular to the radii of the object field

#### 3.2

##### **tangential image surface**

surface on which all tangential structured objects are focused and sharply imaged in the primary image space subject to no aberrations other than astigmatism and Petzval curvature

#### 3.3

##### **sagittal structured object**

object containing short lines parallel to the radii of the object field

#### 3.4

##### **sagittal image surface**

surface on which all sagittal structured objects are focused and sharply imaged in the primary image space subject to no aberrations other than astigmatism and Petzval curvature

#### 3.5

##### **astigmatic difference**

dimensional difference along the optical axis in the tangential plane between the tangential and sagittal image surfaces

**3.6**  
**plan field number**  
**PFN**

number which specifies the diameter, in millimetres, of the sharp region of the primary image of a flat object surface

**3.7**  
**objective field number**  
**OFN**

maximum field of view number of the eyepiece for which the objective is designed to be used

**3.8**  
**plan field ratio**  
**PFR**

ratio of the plan field number to the objective field number, defined as  $PFR = PFN/OFN$

**3.9**  
**Petzval curvature**

curvature of the Petzval surface, which denotes the basic field curvature

## 4 Requirements

### 4.1 Indication

Objective lenses named Plan or with Plan as part of the name in the markings shall also indicate the objective field number on the body of the lens. If the words “flat field” are used in the name in the marking, then the lenses shall also be marked “Plan” with an indication of the OFN on the body of the lens. The indication of objective field number does not apply to objective lenses sold before the year 2014.

Objective field numbers shall be expressed as follows:

18, 19, 20, 21, 22, 23, 24, 25, 26, 26.5, 27, 28, 29, 30 and so on

EXAMPLE In the case of the objective field number 25:

**OFN25**

### 4.2 Definition of plan objectives

The plan field ratio of a plan objective lens shall be at least 0,85.

### 4.3 Determination of plan field number

Let  $\tau_t$  and  $\tau_s$  be the respective distances of the tangential and sagittal image surfaces from the image plane, along the optical axis in a tangential plane. Using  $\tau_t$  and  $\tau_s$ , the average image surface distance,  $\Delta$ , is defined as shown in Formula (1):

$$\Delta = (\tau_t + \tau_s)/2 \quad (1)$$

The plan field number shall be specified by the maximum field of view of the primary image which satisfies the following conditions: the absolute values of both  $\Delta$  and astigmatic difference ( $\tau_t - \tau_s$ ) are

less than or equal to the value  $\delta$  calculated by Berek's formula [see Formula (2)], and the magnification of the eyepiece is  $10\times$ .

$$\delta = \left( \frac{\omega}{M_{\text{TOT VIS}}} \cdot \frac{250\,000}{NA} + \frac{\lambda}{2 \cdot NA^2} \right) \cdot M_0^2 \quad (2)$$

where

- $\delta$  is the depth of focus in image space, in micrometres;
- $\omega$  is a physiological constant which describes the resolution of the human eye, taken to be the angle  $5'$  [ $\omega$  is the arc of this angle (0,001 4)];
- $M_{\text{TOT VIS}}$  is the total visual magnification of the microscope;
- $NA$  is the numerical aperture of objective;
- $\lambda$  is the wavelength of the e-line, in micrometres;
- $M_0$  is the magnification in the primary image plane.

The depth of field calculated by Berek's formula is expressed in [Annex A](#).

## Annex A (informative)

### Depth of field in object space calculated by Berek's formula

Berek's formula is as follows:

$$\delta_{\text{ob}} = n \cdot \left( \frac{\omega}{M_{\text{TOT VIS}}} \cdot \frac{250\,000}{NA} + \frac{\lambda}{2 \cdot NA^2} \right)$$

$$\omega = 0,001\,4, \lambda = 0,55 \mu\text{m}$$



|  |         |        |        |        |        |        |        |
|--|---------|--------|--------|--------|--------|--------|--------|
| Magnification of objective lens  | 4       | 4      | 4      | 4      | 5      | 5      | 5      |
| NA of objective lens   | 0,10    | 0,13   | 0,16   | 0,20   | 0,12   | 0,15   | 0,16   |
| Magnification of eyepiece  | 10      | 10     | 10     | 10     | 10     | 10     | 10     |
| Depth of field at specimen:<br>$\delta_{ob}$ ( $\mu\text{m}$ )                     | 114,825 | 83,476 | 65,361 | 50,581 | 77,309 | 58,811 | 54,424 |
| Refractive index: <i>n</i><br>(dry: <i>n</i> = 1, oil immersion: <i>n</i> = 1,515) | 1       | 1      | 1      | 1      | 1      | 1      | 1      |

|  |        |        |        |        |        |       |
|--|--------|--------|--------|--------|--------|-------|
| Magnification of objective lens  | 10     | 10     | 10     | 10     | 10     | 10    |
| NA of objective lens   | 0,22   | 0,25   | 0,30   | 0,32   | 0,40   | 0,45  |
| Magnification of eyepiece  | 10     | 10     | 10     | 10     | 10     | 10    |
| Depth of field at specimen:<br>$\delta_{ob}$ ( $\mu\text{m}$ )                     | 21,555 | 18,372 | 14,703 | 13,606 | 10,458 | 9,127 |
| Refractive index: <i>n</i><br>(dry: <i>n</i> = 1, oil immersion: <i>n</i> = 1,515) | 1      | 1      | 1      | 1      | 1      | 1     |

|  |       |       |       |       |       |       |
|--|-------|-------|-------|-------|-------|-------|
| Magnification of objective lens  | 20    | 20    | 20    | 20    | 20    | 20    |
| NA of objective lens   | 0,40  | 0,45  | 0,50  | 0,60  | 0,70  | 0,75  |
| Magnification of eyepiece  | 10    | 10    | 10    | 10    | 10    | 10    |
| Depth of field at specimen:<br>$\delta_{ob}$ ( $\mu\text{m}$ )                     | 6,083 | 5,238 | 4,593 | 3,676 | 3,058 | 2,819 |
| Refractive index: <i>n</i><br>(dry: <i>n</i> = 1, oil immersion: <i>n</i> = 1,515) | 1     | 1     | 1     | 1     | 1     | 1     |

|  |       |       |       |       |       |       |       |       |       |       |
|--|-------|-------|-------|-------|-------|-------|-------|-------|-------|-------|
| Magnification of objective lens  | 40    | 40    | 40    | 40    | 40    | 40    | 40    | 40    | 40    |       |
| NA of objective lens   | 0,55  | 0,60  | 0,65  | 0,70  | 0,75  | 0,85  | 0,95  | 1,00  | 1,25  | 1,30  |
| Magnification of eyepiece  | 10    | 10    | 10    | 10    | 10    | 10    | 10    | 10    | 10    | 10    |
| Depth of field at specimen:<br>$\delta_{ob}$ ( $\mu\text{m}$ )                     | 2,494 | 2,217 | 1,993 | 1,808 | 1,652 | 1,408 | 1,224 | 1,740 | 1,325 | 1,265 |
| Refractive index: <i>n</i><br>(dry: <i>n</i> = 1, oil immersion: <i>n</i> = 1,515) | 1     | 1     | 1     | 1     | 1     | 1     | 1     | 1,515 | 1,515 | 1,515 |

|  |       |       |       |       |       |       |       |
|--|-------|-------|-------|-------|-------|-------|-------|
| Magnification of objective lens  | 60    | 60    | 60    | 60    | 60    | 60    | 60    |
| NA of objective lens   | 0,70  | 0,85  | 0,90  | 0,95  | 1,25  | 1,30  | 1,40  |
| Magnification of eyepiece  | 10    | 10    | 10    | 10    | 10    | 10    | 10    |
| Depth of field at specimen:<br>$\delta_{ob}$ ( $\mu\text{m}$ )                     | 1,391 | 1,064 | 0,985 | 0,917 | 0,972 | 0,925 | 0,842 |
| Refractive index: <i>n</i><br>(dry: <i>n</i> = 1, oil immersion: <i>n</i> = 1,515) | 1     | 1     | 1     | 1     | 1,515 | 1,515 | 1,515 |

# ISO 19012-1:2013(E)

|   |       |       |       |       |       |       |       |
|---|-------|-------|-------|-------|-------|-------|-------|
| Magnification of objective lens                                       | 63    | 63    | 63    | 63    | 63    | 63    | 63    |
| NA of objective lens  | 0,70  | 0,75  | 0,80  | 0,95  | 1,25  | 1,32  | 1,40  |
| Magnification of eyepiece   | 10    | 10    | 10    | 10    | 10    | 10    | 10    |
| Depth of field at specimen:<br>$\delta_{ob}$ ( $\mu\text{m}$ )        | 1,351 | 1,227 | 1,121 | 0,888 | 0,938 | 0,875 | 0,812 |
| Refractive index: $n$<br>(dry: $n = 1$ , oil immersion: $n = 1,515$ ) | 1     | 1     | 1     | 1     | 1,515 | 1,515 | 1,515 |

|   |       |       |       |       |       |       |       |
|---|-------|-------|-------|-------|-------|-------|-------|
| Magnification of objective lens                                       | 100   | 100   | 100   | 100   | 100   | 100   | 100   |
| NA of objective lens  | 0,90  | 0,95  | 1,25  | 1,30  | 1,35  | 1,40  | 1,40  |
| Magnification of eyepiece   | 10    | 10    | 10    | 10    | 10    | 10    | 10    |
| Depth of field at specimen:<br>$\delta_{ob}$ ( $\mu\text{m}$ )        | 0,726 | 0,671 | 0,689 | 0,653 | 0,620 | 0,590 | 0,590 |
| Refractive index: $n$<br>(dry: $n = 1$ , oil immersion: $n = 1,515$ ) | 1     | 1     | 1,515 | 1,515 | 1,515 | 1,515 | 1,515 |

## Bibliography

- [1] ISO 8578, *Microscopes — Marking of objectives and eyepieces*
- [2] BEREK M.. Grundlagen der Tiefenwahrnehmung im Mikroskop. *Marburger Sitzungsberichte*. 1927, **62** pp. 189–223

