

---

---

**Determination of particle size  
distribution — Single particle light  
interaction methods —**

**Part 2:  
Light scattering liquid-borne particle  
counter**

*Détermination de la distribution granulométrique — Méthodes  
d'interaction lumineuse de particules uniques —*

*Partie 2: Compteur de particules en suspension dans un liquide en  
lumière dispersée*



Reference number  
ISO 21501-2:2007(E)

© ISO 2007

**PDF disclaimer**

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.



**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2007

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
Case postale 56 • CH-1211 Geneva 20  
Tel. + 41 22 749 01 11  
Fax + 41 22 749 09 47  
E-mail [copyright@iso.org](mailto:copyright@iso.org)  
Web [www.iso.org](http://www.iso.org)

Published in Switzerland

# Contents

Page

Foreword.....	iv
Introduction .....	v
<b>1 Scope .....</b>	<b>1</b>
<b>2 Terms and definitions.....</b>	<b>2</b>
<b>3 Requirements .....</b>	<b>2</b>
3.1 Size calibration.....	2
3.2 Verification of size setting .....	2
3.3 Counting efficiency.....	2
3.4 Size resolution .....	2
3.5 False count rate .....	3
3.6 Maximum particle number concentration .....	3
3.7 Sampling flow rate.....	3
3.8 Sampling time .....	3
3.9 Sampling volume .....	3
3.10 Calibration interval .....	3
3.11 Test report .....	3
<b>4 Test method.....</b>	<b>4</b>
4.1 Size calibration.....	4
4.2 Verification of size setting .....	6
4.3 Counting efficiency.....	6
4.4 Size resolution .....	6
4.5 False count rate .....	7
4.6 Maximum particle number concentration .....	7
4.7 Sampling flow rate.....	7
4.8 Sampling time .....	8
4.9 Sampling volume .....	8
4.10 Calibration .....	8
<b>Annex A (informative) Uncertainty of particle size calibration .....</b>	<b>9</b>
<b>Annex B (informative) Counting efficiency.....</b>	<b>11</b>
<b>Annex C (informative) Size resolution.....</b>	<b>12</b>
<b>Annex D (informative) False count rate.....</b>	<b>13</b>
<b>Bibliography .....</b>	<b>15</b>

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 21501-2 was prepared by Technical Committee ISO/TC 24, *Sieves, sieving and other sizing methods*, Subcommittee SC 4, *Sizing by methods other than sieving*.

This first edition of ISO 21501-2, together with ISO 21501-3 and ISO 21501-4, cancels and replaces ISO 13323-1:2000, which has been technically revised.

ISO 21501 consists of the following parts, under the general title *Determination of particle size distribution — Single particle light interaction methods*:

- *Part 2: Light scattering liquid-borne particle counter*
- *Part 3: Light extinction liquid-borne particle counter*
- *Part 4: Light scattering airborne particle counter for clean spaces*

The following part is under preparation:

- *Part 1: Light scattering aerosol spectrometer*

## Introduction

Monitoring particle contamination levels is required in various fields, e.g. in the electronic industry, in the pharmaceutical industry, in the manufacturing of precision machines and in medical operations. Particle counters are useful instruments for monitoring particle contamination in liquid. The purpose of this part of ISO 21501 is to provide a calibration procedure and verification method for particle counters, so as to minimize the inaccuracy in the measurement result by a counter, as well as the differences in the results measured by different instruments.

Copyright International Organization for Standardization  
Provided by IHS under license with ISO  
No reproduction or networking permitted without license from IHS

# Determination of particle size distribution — Single particle light interaction methods —

## Part 2: Light scattering liquid-borne particle counter

### 1 Scope

This part of ISO 21501 describes a calibration and verification method for a light scattering liquid-borne particle counter (LSLPC), which is used to measure the size and particle number concentration of particles suspended in liquid. The light scattering method described in this part of ISO 21501 is based on single particle measurements. The typical size range of particles measured by this method is between 0,1  $\mu\text{m}$  and 10  $\mu\text{m}$  in particle size.

Instruments that conform to this part of ISO 21501 are used for the evaluation of the cleanliness of pure water and chemicals, as well as the measurement of number and size distribution of particles in various liquids. The measured particle size using the LSLPC depends on the refractive index of particles and medium; therefore the measured particle size is equivalent to the calibration particles in pure water.

The following are within the scope of this part of ISO 21501:

- size calibration;
- verification of size setting;
- counting efficiency;
- size resolution;
- false count rate;
- maximum particle number concentration;
- sampling flow rate;
- sampling time;
- sampling volume;
- calibration interval;
- test report.

## 2 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

**2.1 calibration particle**  
mono-disperse spherical particle with a known mean particle size, e.g. polystyrene latex (PSL) particle, that is traceable to an international standard of length, and where the standard uncertainty of the mean particle size is equal to or less than  $\pm 2,5$  %

NOTE The refractive index of calibration particles is close to 1,59 at a wavelength of 589 nm (sodium D line).

**2.2 counting efficiency**  
ratio of the measured result of a light scattering liquid-borne particle counter (LSLPC) to that of a reference instrument using the same sample

**2.3 particle counter**  
instrument that counts the number of particles and measures their size using the light scattering method or the light extinction method

**2.4 pulse height analyser  
PHA**  
instrument that analyses the distribution of pulse heights

**2.5 size resolution**  
measure of the ability of an instrument to distinguish between particles of different sizes

## 3 Requirements

### 3.1 Size calibration

The recommended procedure for the size calibration is described in 4.1.

### 3.2 Verification of size setting

The error in the detectable minimum particle size and other sizes specified by the manufacturer of an LSLPC shall be equal to or less than  $\pm 15$  % when the test is carried out by the method described in 4.2.

### 3.3 Counting efficiency

The counting efficiency shall be  $(50 \pm 30)$  % for calibration particles with a size close to the minimum detectable size, and it shall be  $(100 \pm 30)$  % for calibration particles with the particle size of 1,5 times to 3 times larger than the minimum detectable particle size.

### 3.4 Size resolution

The size resolution shall be equal to or less than 10 % for calibration particles of a size specified by the manufacturer.



### 3.5 False count rate

The false count rate is determined by measuring the particle number concentration in the unit of counts per litre at the minimum reported size range when sampling pure water.

### 3.6 Maximum particle number concentration

The maximum measurable particle number concentration shall be specified by the manufacturer. The coincidence loss at the maximum particle number concentration of an LSLPC shall be equal to or less than 10 %.

**NOTE** When the particle number concentration is higher than the maximum particle number concentration, the number of uncounted particles increases because of an enhanced probability of multiple particles existing in the sensing volume (coincidence error) and/or saturation of the electronic system.

### 3.7 Sampling flow rate

The standard uncertainty of the sampling flow rate shall be specified by the manufacturer. The user shall check that the sampling flow rate is within the range specified by the manufacturer.

### 3.8 Sampling time

The standard uncertainty in the duration of sampling time shall be equal to or less than  $\pm 1$  % of the preset value.

If the LSLPC does not have a sampling time control system, this subclause does not apply.

### 3.9 Sampling volume

The standard uncertainty of sampling volume shall be equal to or less than  $\pm 5$  % of the preset value.

This subclause does not apply when the LSLPC is not equipped with a sampling system.

### 3.10 Calibration interval

It is recommended that the calibration interval of an LSLPC be one year or less.

### 3.11 Test report


The following minimum information shall be recorded:

- a) date of calibration;
- b) calibration particle sizes;
- c) flow rate;
- d) size resolution (with the particle size used);
- e) counting efficiency;
- f) voltage limit or channel of an internal pulse height analyser (PHA).

## 4 Test method

### 4.1 Size calibration

When calibrating an LSLPC with calibration particles of known size, the median voltage (or internal PHA channel), corresponds to the particle size (see Figure 1). The median voltage (or internal PHA channel) should be determined by using a particle counter with variable voltage limit (or internal PHA channel) settings. The median voltage (or internal PHA channel) is the voltage (or internal PHA channel) that equally divides the total number of pulses counted. When a particle counter with variable voltage limit settings is not available, a PHA can be used in place of the counter.



#### Key

X pulse height voltage (or channel)

Y density

1 pulse height distribution with PSL particles


$V_l$  lower voltage limit

$V_m$  median voltage

$V_u$  upper voltage limit

**Figure 1 — Pulse height distribution of PSL particle signals**

When noise signals appear as if there are many small particles in the sample, the median voltage (or internal PHA channel) shall be determined by discarding the pulses due to “false particles” [see Figure 2 a)]. The discarding should only be done when the density at the peak due to real particles is more than twice the density at the valley that separates it from the pulses due to “false particles” [see Figure 2 b)]. In this case,  $V_u$  is the voltage greater than the median voltage,  $V_m$ , where the density is the same as  $V_l$ . The median is calculated using only the population between the voltage limits  $V_l$  and  $V_u$ .




**Key**

- X pulse height voltage (or channel)
- Y density
- 1 pulse height distribution with PSL particles
- 2 noise (false particles, small particles and/or optical, electrical noise)
- $V_l$  lower voltage limit
- $V_m$  median voltage
- $V_u$  upper voltage limit

**Figure 2 — Pulse height distribution of PSL particle signals with noise**

The voltages of channels corresponding to particle size should be determined in accordance with the calibration curve provided by the manufacturer (see Figure 3).



**Key**

- X particle size
- Y median value of calibration particles
- 1 calibration curve
- $V_{m,1}$  median voltage corresponding to particle size  $x_{m,1}$
- $V_{m,2}$  median voltage corresponding to particle size  $x_{m,2}$
- $V_{m,3}$  median voltage corresponding to particle size  $x_{m,3}$

**Figure 3 — Calibration curve**

**NOTE** When the median voltage is determined by using an external PHA, the uncertainty in the voltage of PHA and the voltage uncertainty of the LSLPC are included in setting the voltage limits of the LSLPC (see Annex A).

## 4.2 Verification of size setting

Obtain response voltages (or internal PHA channel) in accordance with the test method given in 4.1, using at least three kinds of calibration particles that span most of the reported size range,  $x_r$ , of the LSLPC. Determine the calibration curve from these response voltages (or internal PHA channel) and the calibration particle sizes. Calculate the corresponding particle size,  $x_s$ , from the voltage (or internal PHA channel) setting of the LSLPC using the calibration curve. Obtain the size setting error,  $\varepsilon$ , by means of Equation (1) below, and examine whether it satisfies the requirement given in 3.2.

$$\varepsilon(\%) = \frac{x_s - x_r}{x_r} \times 100 \% \quad (1)$$

where

$\varepsilon$  is the size setting error, in %;

$x_r$  is the reported size range, in  $\mu\text{m}$ ;

$x_s$  is the calculated particle size, in  $\mu\text{m}$ .

## 4.3 Counting efficiency

To test the counting efficiency of the LSLPC, use calibration particles with two sizes: one that is close to the minimum detectable reported size range, and another that is 1,5 times to 3 times larger than the minimum detectable size.

Measure the particle number concentration of both particles with the LSLPC under test and either a microscopic method or a calibrated LSLPC as a reference instrument.

The counting efficiency is the ratio of the particle number concentration measured by the LSLPC under test and the particle number concentration measured by the reference instrument (see Annex B).

## 4.4 Size resolution

Determine the median voltage (or channel),  $V_m$ , using monodisperse calibration particles, as shown in Figure 4.

The lower voltage limit,  $V_l$ , and upper voltage limit,  $V_u$ , are defined as those corresponding to a density of 61 %. Using the calibration curve, determine the particle sizes corresponding to  $V_l$  and  $V_u$ . Calculate the absolute value of the differences in particle size between PSL particle size and particle size corresponding to  $V_l$  and  $V_u$ . The greater of these is the observed standard deviation,  $\sigma$ . Calculate the percentage of size resolution,  $R$ , of the LSLPC by Equation (2) below (see also Annex C).

$$R(\%) = \frac{\sqrt{\sigma^2 - \sigma_p^2}}{x_p} \times 100 \% \quad (2)$$


where

$R$  is the size resolution, in %;

$\sigma$  is the observed standard deviation of LSLPC, in  $\mu\text{m}$ ;

$\sigma_p$  is the supplier's reported standard deviation of calibration particles, in  $\mu\text{m}$ ;

$x_p$  is the particle size of the calibration particle, in  $\mu\text{m}$ .

**Key**

X pulse height voltage (or channel)  
Y density

- 1 pulse height distribution with PSL particles  
2 lower side resolution  
3 upper side resolution

$V_l$  lower voltage limit  
 $V_m$  median voltage  
 $V_u$  upper voltage limit

**Figure 4 — Size resolution**

#### 4.5 False count rate

The false count rate is the measured particle number concentration (in particles per litre) when the LSLPC is set to the minimum detectable size and particle free liquid flows to the LSLPC. The data should be statistically processed using the Poisson distribution with a 95 % upper confidence limit (see Annex D). The false count rate shall be described in units of particle number concentration (in counts per litre).

#### 4.6 Maximum particle number concentration

The coincidence loss is determined by the flow rate, the time required for particles to pass through the sensing zone and the electrical signal processing time. These values are determined by the design of the LSLPC. Coincidence loss is calculated as in Equation (3) below.

$$L(\%) = [1 - \exp(-q \cdot t \cdot C_{\max})] \times 100 \% \quad (3)$$

where

$L$  is the coincidence loss, in %;

$q$  is the flow rate, in  $\text{cm}^3/\text{s}$ ;

$t$  is the time of passing through the sensing region plus electrical processing time, in s;

$C_{\max}$  is the maximum particle number concentration, in particles per cubic centimetre.

#### 4.7 Sampling flow rate

Obtain a flow rate by the sampling volume (see 4.9) and the sampling time (see 4.8), or use a calibrated flow meter. If the LSLPC does not have a sampling function, this subclause does not apply.

#### 4.8 Sampling time

Sampling time is the time during which the LSLPC measures a sample (from the beginning of counting to the end of counting).

The sampling time tolerance is one minus the ratio of the measured sampling time,  $t$ , to the instrument's specified sampling time,  $t_0$ . This is shown as  $1 - \frac{t}{t_0}$ .

Examine whether the sampling time tolerance satisfies the requirement given in 3.8. Calibrated instruments should be used for sampling time measurement.

This subclause does not apply when the LSLPC is not equipped with a sampling system.

This subclause does not apply when the LSLPC is equipped with a volumetric sampling system.

#### 4.9 Sampling volume

Measure the sampling volume by weighing the pure water with the balance and converting to volume, or measure the volume by means of a calibrated graduated cylinder.

If the LSLPC does not have a sampling function, this subclause does not apply.

#### 4.10 Calibration

Calibration at the calibration interval (see 3.10) should include at least size calibration, size resolution, counting efficiency and sampling volume uncertainty. If the LSLPC does not have a control function, the standard uncertainty of sampling flow rate does not apply.


## Annex A (informative)

### Uncertainty of particle size calibration

#### A.1 Size calibration using external and internal PHA


Figure A.1 shows the particle size calibration using an external PHA and a voltmeter. In this case, there are four sources of uncertainty:

- PSL particles,
- PHA,
- voltmeter, and
- offset voltage at the size setting circuit.



**Figure A.1 — Particle size calibration using external instruments (PHA and voltmeter)**

However, in Figure A.2, the uncertainty of particle size calibration depends only on the PSL particle size uncertainty.



**Figure A.2 — Particle size calibration using an internal PHA**

## A.2 Uncertainty of size calibration

Tables A.1 and A.2 show examples of uncertainty of size calibration. Table A.1 shows an example of combined standard uncertainty for size calibration using an external PHA and voltmeter. Table A.2 shows an example of combined standard uncertainty for size calibration using an internal PHA. The combined standard uncertainty for size calibration using an internal PHA is smaller than when using an external PHA.

**Table A.1 — Relative standard uncertainty of size calibration using an external PHA (for example)**

Items	Standard uncertainty %
PSL particles	2,5
PHA	2,5
Voltmeter	0,1
Offset voltage	0,5
Calibration curve	1,5
Combined standard uncertainty	3,9
Expanded uncertainty (k=2)	7,8

NOTE The standard uncertainty of the calibration curve is the uncertainty in the relationship between particle size and voltage limit or internal PHA channel.

**Table A.2 — Relative standard uncertainty of size calibration using an internal PHA (for example)**

Items	Standard uncertainty %
PSL particles	2,5
Calibration curve	1,5
Combined standard uncertainty	2,9
Expanded uncertainty (k=2)	5,8

NOTE The standard uncertainty of the calibration curve is the uncertainty in the relationship between particle size and voltage limit or internal PHA channel.




## Annex B (informative)

### Counting efficiency

Figure B.1 shows the test system for counting efficiency. The sample contains calibration particles in pure water. The counting efficiency of the reference particle counter at the minimum detectable particle size of the particle counter under test shall be 100 %.

The counting efficiency is obtained by calculating the ratio of the particle number concentration measured by the particle counter under test and the particle number concentration measured by the reference particle counter. The particle number concentration of the sample should be less than 25 % of the maximum particle number concentration of both the reference particle counter and the particle counter under test. The counting efficiency of the reference particle counter shall be established by a method of known uncertainty, such as the microscopic method (see method described in JIS B 9925 [5]).



**Figure B.1 — Example of a counting efficiency test system**

Determine the counting efficiency of  $\eta$  by means of Equation B.1:

$$\eta = \frac{C_1}{C_0} \times 100 \% \quad (\text{B.1})$$

where

$\eta$  is the counting efficiency, in %;

$C_0$  is the particle number concentration measured by reference particle counter, in particles per litre;

$C_1$  is the particle number concentration measured by particle counter under test, in particles per litre.

## Annex C (informative)

### Size resolution

Size resolution is defined as one standard deviation of the measured size distribution of monodisperse calibration particles, expressed as a percentage of the mean size of the monodisperse calibration particles.

If the distribution of calibration particles is assumed to be the Gaussian distribution,

$$f(x) = \frac{1}{\sqrt{2\pi}\sigma} \exp\left\{-\frac{1}{2}\left(\frac{x-\mu}{\sigma}\right)^2\right\} \quad (\text{C.1})$$

where

$f(x)$  is the Gaussian function;

$x$  is the particle size;

$\mu$  is the mean value;

$\sigma$  is the standard deviation.

When  $(x - \mu) = \pm\sigma$ , the ratio of density to the maximum density is  $\exp\left(-\frac{1}{2}\right) \approx 0,61$ . This is the basis for the use of 61 % in the determination of size resolution.

## Annex D (informative)

### False count rate

The probability of appearance of false counts is assumed to be defined by the Poisson distribution. The Poisson distribution is defined by Equation (D.1):

$$P(X : \lambda) = \frac{e^{-\lambda} \lambda^X}{X!} \quad (\text{D.1})$$

where

$X$  is the number of false counts;

$\lambda$  is the mean value of the population;

$P(X : \lambda)$  is the probability of observing value  $X$  from a population having a mean value of  $\lambda$ .

The lower confidence limit,  $\lambda_l$ , is defined by Equation (D.2):

$$\sum_{X=X}^{\infty} P(X; \lambda_l) = \frac{\varepsilon}{2} \quad (\text{D.2})$$

where

$\varepsilon$  is the size setting error.

The upper confidence limit,  $\lambda_u$ , is defined by Equation (D.3):

$$\sum_{X=0}^X P(X; \lambda_u) = \frac{\varepsilon}{2} \quad (\text{D.3})$$

When the confidence limit is 95 %,  $\varepsilon$  is 0,05.

Table D.1 shows the observed count and the calculated upper and lower 95 % confidence limits. When the observed count is zero, it is possible to have up to three counts with a probability of 5 %. For example, if zero counts are observed in one minute at a sampling flow rate of 100 l/min, the false count rate is three counts in the volume sampled in one minute with a 95 % confidence limit, i.e. the false count is 30 counts per litre.

Table D.1 — Observed count and 95 % confidence limit

Observed count	Lower confidence limit $\lambda_l$	Upper confidence limit $\lambda_u$
0	0	3
1	0,05	4,7
2	0,36	6,3
3	0,82	7,8
4	1,37	9,2
5	1,97	10,5
6	2,61	11,8
7	3,28	13,1
8	3,98	14,4
9	4,70	15,7
10	5,43	17,0

.....

## Bibliography

- [1] ISO 9276-1, *Representation of results of particle size analysis — Part 1: Graphical representation*
- [2] ASTM F50-92 (2001)e1, *Standard Practice for Continuous Sizing and Counting of Airborne Particles in Dust-Controlled Areas and Clean Rooms Using Instruments Capable of Detecting Single Sub-Micrometre and Larger Particles*
- [3] ASTM F328-98 (2003), *Standard Practice for Calibration of an Airborne Particle Counter Using Monodisperse Spherical Particles*
- [4] ASTM F649-01, *Standard Practice for Secondary Calibration of Airborne Particle Counter Using Comparison Procedures*
- [5] JIS B 9925:1997, *Light scattering automatic particle counter for liquid*
- [6] *Guide to the expression of uncertainty in measurement (GUM)*, BIPM, IEC, IFCC, ISO, IUPAC, IUPAP, OIML, 1993, corrected and reprinted in 1995

---

---

**ICS 19.120**

Price based on 15 pages

-----